

CALIBRATION VERIFICATION SUMMARY

Contract: SAXT01

Lab Code: CHEM Case No.: Q1938 SAS No.: Q1938 SDG NO.: Q1938

Continuing Calib Date: 05/06/2025 Initial Calibration Date(s): 04/23/2025 04/23/2025

Continuing Calib Time: 16:37 Initial Calibration Time(s): 11:31 12:26

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.12	9.12	9.02	9.22	0.00
Tetrachloro-m-xylene	3.60	3.60	3.50	3.70	0.00
alpha-BHC	4.05	4.05	3.95	4.15	0.00
beta-BHC	4.57	4.56	4.46	4.66	-0.01
delta-BHC	4.81	4.81	4.71	4.91	0.00
gamma-BHC (Lindane)	4.38	4.38	4.28	4.48	0.00
Heptachlor	4.98	4.98	4.88	5.08	0.00
Aldrin	5.32	5.32	5.22	5.42	0.00
Heptachlor epoxide	5.74	5.74	5.64	5.84	0.00
Endosulfan I	6.12	6.12	6.02	6.22	0.00
Dieldrin	6.40	6.40	6.30	6.50	0.00
4,4'-DDE	6.25	6.24	6.14	6.34	0.00
Endrin	6.63	6.62	6.52	6.72	-0.01
Endosulfan II	6.84	6.84	6.74	6.94	0.00
4,4'-DDD	6.75	6.75	6.65	6.85	0.00
Endosulfan sulfate	7.20	7.20	7.10	7.30	0.00
4,4'-DDT	7.07	7.07	6.97	7.17	0.00
Methoxychlor	7.54	7.54	7.44	7.64	0.00
Endrin ketone	7.68	7.68	7.58	7.78	0.00
Endrin aldehyde	6.97	6.96	6.86	7.06	-0.01
alpha-Chlordane	6.08	6.08	5.98	6.18	0.00
gamma-Chlordane	6.00	5.99	5.89	6.09	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: SAXT01

Lab Code: CHEM Case No.: Q1938 SAS No.: Q1938 SDG NO.: Q1938

Continuing Calib Date: 05/06/2025 Initial Calibration Date(s): 04/23/2025 04/23/2025

Continuing Calib Time: 16:37 Initial Calibration Time(s): 11:31 12:26

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.09	8.09	7.99	8.19	0.00
Tetrachloro-m-xylene	2.91	2.91	2.81	3.01	0.00
alpha-BHC	3.42	3.42	3.32	3.52	0.00
beta-BHC	4.05	4.04	3.94	4.14	-0.01
delta-BHC	4.28	4.28	4.18	4.38	0.00
gamma-BHC (Lindane)	3.75	3.75	3.65	3.85	0.00
Heptachlor	4.11	4.10	4.00	4.20	-0.01
Aldrin	4.39	4.39	4.29	4.49	0.00
Heptachlor epoxide	4.90	4.89	4.79	4.99	-0.01
Endosulfan I	5.27	5.27	5.17	5.37	0.00
Dieldrin	5.53	5.53	5.43	5.63	0.00
4,4'-DDE	5.40	5.40	5.30	5.50	0.00
Endrin	5.81	5.81	5.71	5.91	0.00
Endosulfan II	6.10	6.10	6.00	6.20	0.00
4,4'-DDD	5.95	5.95	5.85	6.05	0.00
Endosulfan sulfate	6.50	6.50	6.40	6.60	0.00
4,4'-DDT	6.21	6.20	6.10	6.30	-0.01
Methoxychlor	6.78	6.78	6.68	6.88	0.00
Endrin ketone	7.01	7.00	6.90	7.10	-0.01
Endrin aldehyde	6.28	6.28	6.18	6.38	0.00
alpha-Chlordane	5.21	5.21	5.11	5.31	0.00
gamma-Chlordane	5.15	5.15	5.05	5.25	0.00