

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: SAXT01
 ProjectID: Stan Hope
 Lab Code: CHEM Case No.: Q1938 SAS No.: Q1938 SDG No.: Q1938

Calibration Sequence : FG042425		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	193935875	114080	FG015756.D
850	102468017	120551	FG015757.D
340	41273634	121393	FG015758.D
170	22173981	130435	FG015759.D
85	10166512	119606	FG015760.D
AVG RF : 121213		% RSD : 4.863	AVG RT : 15.0012