

SOIL TPH GC SURROGATE RECOVERY

Lab Name: Chemtech Client: Saxton Falls Sand and Gravel Co. Inc.
Lab Code: CHEM Case No.: Q1938 SAS No.: Q1938 SDG No.: Q1938

EPA SAMPLE NO.	S1 TETRACOSANE-d50	S2	S3	S4	TOT OUT
PIBLK-FF015793.D	83				0
PIBLK-FF015804.D	86				0
PIBLK-FG015783.D	94				0
PIBLK-FG015796.D	91				0
PIBLK-FG015799.D	88				0
PIBLK-FG015803.D	89				0
PB167887BL	67				0
PB167887BS	89				0
SMALL-PILE-AMS	40				0
SMALL-PILE-AMSD	37				0
LOWER-WALL-PILE-A	44				0
LOWER-WALL-PILE-B	49				0
LOWER-WALL-PILE-C	49				0
LOWER-WALL-PILE-D	42				0

QC LIMITS

TETRACOSANE-d50

For Water : 29-130
For Soil : 37-130

Column to be used to flag recovery values
* Values outside of contract required QC limits
D Surrogate Diluted Out