Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: SAXT01

Lab Code: CHEM Case No.: Q1938 SAS No.: Q1938 SDG No.: Q1938

Instrument ID: BNA_M Calibration Date(s): 04/28/2025 04/28/2025

Calibration Time(s): 12:30 17:04

LAB FILE ID: RRF2.5 = BM050024.D RRF005 = BM050025.D RRF010 = BM050026.D
RRF020 = BM050027.D RRF040 = BM050028.D RRF050 = BM050029.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.074	1.079	1.107	1.208	1.211	1.142	5.0
Benzaldehyde		0.940	0.920	0.971	1.031	1.013	0.961	5.1
Phenol-d6		1.290	1.318	1.398	1.543	1.537	1.436	7.1
Phenol		1.349	1.344	1.420	1.539	1.537	1.453	5.7
bis(2-Chloroethyl)ether		1.213	1.105	1.176	1.279	1.271	1.214	4.9
2-Chlorophenol		1.169	1.161	1.198	1.312	1.301	1.235	5.0
2-Methylphenol		0.857	0.893	0.931	1.033	1.026	0.957	7.0
2,2-oxybis(1-Chloropropane)		1.487	1.422	1.445	1.552	1.513	1.475	3.2
Acetophenone		0.483	0.464	0.485	0.531	0.525	0.497	4.9
3+4-Methylphenols		1.137	1.170	1.267	1.393	1.392	1.293	8.1
n-Nitroso-di-n-propylamine	0.802	0.889	0.885	0.921	1.003	0.988	0.921	7.0
Nitrobenzene-d5		0.386	0.374	0.397	0.434	0.433	0.406	5.7
Hexachloroethane		0.552	0.508	0.521	0.556	0.548	0.533	3.8
Nitrobenzene		0.337	0.325	0.345	0.374	0.372	0.351	5.2
Isophorone		0.666	0.648	0.680	0.729	0.729	0.692	4.4
2-Nitrophenol		0.148	0.157	0.172	0.196	0.198	0.179	11.3
2,4-Dimethylphenol		0.263	0.266	0.289	0.324	0.322	0.297	8.4
bis(2-Chloroethoxy)methane		0.407	0.399	0.422	0.455	0.452	0.428	4.9
2,4-Dichlorophenol		0.294	0.298	0.326	0.362	0.359	0.333	8.5
Naphthalene		1.001	0.959	0.993	1.066	1.057	1.013	3.8
4-Chloroaniline		0.361	0.393	0.404	0.441	0.448	0.415	7.5
Hexachlorobutadiene		0.253	0.240	0.249	0.270	0.270	0.258	4.3
Caprolactam		0.090	0.085	0.095	0.104	0.104	0.097	7.7
4-Chloro-3-methylphenol		0.289	0.283	0.302	0.325	0.329	0.308	5.7
2-Methylnaphthalene		0.656	0.639	0.662	0.713	0.714	0.679	4.2
Hexachlorocyclopentadiene			0.303	0.362	0.455	0.468	0.423	17.3
2,4,6-Trichlorophenol		0.372	0.381	0.423	0.476	0.485	0.439	10.8
2-Fluorobiphenyl		1.588	1.555	1.632	1.807	1.822	1.691	6.2
2,4,5-Trichlorophenol		0.410	0.417	0.453	0.517	0.519	0.474	9.8
1,1-Biphenyl		1.427	1.395	1.452	1.573	1.581	1.487	4.8
2-Chloronaphthalene		1.126	1.112	1.155	1.250	1.248	1.177	4.7
2-Nitroaniline		0.238	0.242	0.269	0.302	0.304	0.276	10.0
Dimethylphthalate		1.444	1.377	1.437	1.532	1.543	1.462	4.0
Acenaphthylene		1.794	1.715	1.810	1.972	1.971	1.857	5.1

All other compounds must meet a minimum RRF of 0.010.

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: SAXT01

Lab Code: CHEM Case No.: Q1938 SAS No.: Q1938 SDG No.: Q1938

Instrument ID: BNA_M Calibration Date(s): 04/28/2025 04/28/2025

Calibration Time(s): 12:30 17:04

LAB FILE ID: RRF2.5 = BM050024.D RRF005 = BM050025.D RRF010 = BM050026.D
RRF020 = BM050027.D RRF040 = BM050028.D RRF050 = BM050029.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.266	0.271	0.299	0.327	0.331	0.303	8.5
3-Nitroaniline		0.233	0.247	0.284	0.321	0.325	0.289	12.5
Acenaphthene		1.038	0.998	1.046	1.133	1.145	1.075	4.9
2,4-Dinitrophenol			0.100	0.132	0.178	0.194	0.164	23.6
4-Nitrophenol			0.128	0.180	0.224	0.237	0.205	21.0
Dibenzofuran		1.749	1.681	1.749	1.876	1.888	1.788	4.2
2,4-Dinitrotoluene		0.338	0.366	0.412	0.456	0.468	0.418	11.7
Diethylphthalate		1.372	1.301	1.380	1.433	1.451	1.377	3.9
4-Chlorophenyl-phenylether		0.741	0.716	0.770	0.857	0.871	0.804	7.6
Fluorene		1.387	1.343	1.431	1.558	1.573	1.463	5.8
4-Nitroaniline		0.208	0.234	0.277	0.313	0.321	0.279	15.2
4,6-Dinitro-2-methylphenol			0.088	0.110	0.135	0.142	0.125	17.2
n-Nitrosodiphenylamine		0.580	0.572	0.596	0.655	0.657	0.614	5.6
2,4,6-Tribromophenol		0.253	0.249	0.281	0.318	0.329	0.296	11.6
4-Bromophenyl-phenylether		0.223	0.216	0.229	0.258	0.258	0.242	7.6
Hexachlorobenzene		0.265	0.251	0.264	0.294	0.295	0.279	6.5
Atrazine		0.196	0.198	0.214	0.236	0.242	0.222	8.5
Pentachlorophenol			0.117	0.141	0.167	0.172	0.157	14.5
Phenanthrene		1.075	1.035	1.083	1.187	1.212	1.128	5.8
Anthracene		1.062	1.038	1.096	1.214	1.236	1.142	6.8
Carbazole		0.909	0.900	0.960	1.054	1.075	0.991	7.1
Di-n-butylphthalate		1.106	1.082	1.151	1.246	1.277	1.180	6.2
Fluoranthene		1.178	1.142	1.250	1.410	1.455	1.324	9.9
Pyrene		1.290	1.255	1.332	1.518	1.520	1.404	7.8
Terphenyl-d14		1.176	1.189	1.332	1.524	1.541	1.352	11.6
Butylbenzylphthalate		0.489	0.479	0.515	0.562	0.568	0.526	6.5
3,3-Dichlorobenzidine		0.434	0.432	0.483	0.580	0.597	0.527	14.1
Benzo(a)anthracene		1.280	1.231	1.321	1.478	1.479	1.377	7.2
Chrysene		1.208	1.170	1.214	1.342	1.369	1.275	6.0
Bis(2-ethylhexyl)phthalate		0.726	0.725	0.775	0.843	0.852	0.786	6.6
Di-n-octyl phthalate		1.221	1.197	1.267	1.366	1.404	1.296	5.9
Benzo(b)fluoranthene		1.128	1.149	1.209	1.406	1.415	1.301	10.2
Benzo(k)fluoranthene		1.219	1.150	1.251	1.387	1.449	1.316	8.3
Benzo(a)pyrene		1.094	1.066	1.147	1.295	1.330	1.218	9.2
Indeno(1,2,3-cd)pyrene		1.301	1.293	1.404	1.602	1.646	1.490	10.3

All other compounds must meet a minimum RRF of 0.010.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: SAXT01

Lab Code: CHEM Case No.: Q1938 SAS No.: Q1938 SDG No.: Q1938

Instrument ID: BNA_M Calibration Date(s): 04/28/2025 04/28/2025

Calibration Time(s): 12:30 17:04

LAB FILE ID: RRF2.5 = BM050024.D RRF005 = BM050025.D RRF010 = BM050026.D
RRF020 = BM050027.D RRF040 = BM050028.D RRF050 = BM050029.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo(a,h)anthracene		1.056	1.044	1.142	1.300	1.347	1.215	10.7
Benzo(g,h,i)perylene		1.066	1.030	1.105	1.236	1.268	1.163	8.1
1,2,4,5-Tetrachlorobenzene		0.624	0.622	0.655	0.742	0.745	0.692	8.2
1,4-Dioxane		0.506	0.464	0.478	0.511	0.504	0.490	3.8
2,3,4,6-Tetrachlorophenol		0.366	0.362	0.398	0.435	0.445	0.410	8.4

All other compounds must meet a minimum RRF of 0.010.