

Surrogate Summary

SDG No.: Q1938

Client: Saxton Falls Sand and Gravel Co. Inc.

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1938-02	LOWER-WALL-PILE-A	1,2-Dichloroethane-d4	50	53.2	106	70 (63)	130 (155)
		Dibromofluoromethane	50	50.6	101	70 (70)	130 (134)
		Toluene-d8	50	48.4	97	70 (74)	130 (123)
		4-Bromofluorobenzene	50	33.8	68 *	70 (38)	130 (136)
Q1938-04	LOWER-WALL-PILE-B	1,2-Dichloroethane-d4	50	54.3	109	70 (63)	130 (155)
		Dibromofluoromethane	50	50.8	102	70 (70)	130 (134)
		Toluene-d8	50	48.2	96	70 (74)	130 (123)
		4-Bromofluorobenzene	50	37.3	75	70 (38)	130 (136)
Q1938-06	LOWER-WALL-PILE-C	1,2-Dichloroethane-d4	50	55.8	112	70 (63)	130 (155)
		Dibromofluoromethane	50	51.3	103	70 (70)	130 (134)
		Toluene-d8	50	48.8	98	70 (74)	130 (123)
		4-Bromofluorobenzene	50	37.0	74	70 (38)	130 (136)
Q1938-08	LOWER-WALL-PILE-D	1,2-Dichloroethane-d4	50	53.8	108	70 (63)	130 (155)
		Dibromofluoromethane	50	50.4	101	70 (70)	130 (134)
		Toluene-d8	50	48.7	97	70 (74)	130 (123)
		4-Bromofluorobenzene	50	38.1	76	70 (38)	130 (136)
VY0505SBL01	VY0505SBL01	1,2-Dichloroethane-d4	50	49.8	100	70 (63)	130 (155)
		Dibromofluoromethane	50	50.1	100	70 (70)	130 (134)
		Toluene-d8	50	46.5	93	70 (74)	130 (123)
		4-Bromofluorobenzene	50	55.3	111	70 (38)	130 (136)
VY0505SBS01	VY0505SBS01	1,2-Dichloroethane-d4	50	49.9	100	70 (63)	130 (155)
		Dibromofluoromethane	50	49.4	99	70 (70)	130 (134)
		Toluene-d8	50	48.9	98	70 (74)	130 (123)
		4-Bromofluorobenzene	50	50.1	100	70 (38)	130 (136)
VY0505SBSD01	VY0505SBSD01	1,2-Dichloroethane-d4	50	49.0	98	70 (63)	130 (155)
		Dibromofluoromethane	50	48.3	97	70 (70)	130 (134)
		Toluene-d8	50	48.5	97	70 (74)	130 (123)
		4-Bromofluorobenzene	50	48.4	97	70 (38)	130 (136)

