

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SAXT01
 Lab Code: CHEM Case No.: Q1938 SAS No.: Q1938 SDG NO.: Q1938
 Lab File ID: VY022145.D Date Analyzed: 05/05/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:46
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	221579	13.347				
UPPER LIMIT	443158	13.847				
LOWER LIMIT	110790	12.847				
EPA SAMPLE NO.						
LOWER-WALL-PILE-A	118743	13.35				
LOWER-WALL-PILE-B	135235	13.35				
LOWER-WALL-PILE-C	132998	13.35				
LOWER-WALL-PILE-D	137265	13.35				
VY0505SBL01	191516	13.35				
VY0505SBS01	212383	13.35				
VY0505SBSD01	209502	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.