

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV113	Mercury	3.79	4.0	95	90 - 110	CV	05/05/2025	08:30	LB135655

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV67	Mercury	5.16	5.0	103	90 - 110	CV	05/05/2025	08:34	LB135655
CCV68	Mercury	5.02	5.0	100	90 - 110	CV	05/05/2025	09:07	LB135655
CCV69	Mercury	5.09	5.0	102	90 - 110	CV	05/05/2025	09:37	LB135655
CCV70	Mercury	4.98	5.0	100	90 - 110	CV	05/05/2025	09:56	LB135655

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2360	2500	94	90 - 110	P	05/05/2025	14:18	LB135674
	Antimony	1000	1000	100	90 - 110	P	05/05/2025	14:18	LB135674
	Arsenic	973	1000	97	90 - 110	P	05/05/2025	14:18	LB135674
	Barium	522	520	100	90 - 110	P	05/05/2025	14:18	LB135674
	Beryllium	483	510	95	90 - 110	P	05/05/2025	14:18	LB135674
	Cadmium	501	510	98	90 - 110	P	05/05/2025	14:18	LB135674
	Calcium	9510	10000	95	90 - 110	P	05/05/2025	14:18	LB135674
	Chromium	508	520	98	90 - 110	P	05/05/2025	14:18	LB135674
	Cobalt	489	520	94	90 - 110	P	05/05/2025	14:18	LB135674
	Copper	499	510	98	90 - 110	P	05/05/2025	14:18	LB135674
	Iron	9600	10000	96	90 - 110	P	05/05/2025	14:18	LB135674
	Lead	988	1000	99	90 - 110	P	05/05/2025	14:18	LB135674
	Magnesium	5640	6000	94	90 - 110	P	05/05/2025	14:18	LB135674
	Manganese	485	520	93	90 - 110	P	05/05/2025	14:18	LB135674
	Nickel	490	530	92	90 - 110	P	05/05/2025	14:18	LB135674
	Potassium	9270	9900	94	90 - 110	P	05/05/2025	14:18	LB135674
	Selenium	1020	1000	102	90 - 110	P	05/05/2025	14:18	LB135674
	Silver	228	250	91	90 - 110	P	05/05/2025	14:18	LB135674
	Sodium	9280	10000	93	90 - 110	P	05/05/2025	14:18	LB135674
	Thallium	1020	1000	102	90 - 110	P	05/05/2025	14:18	LB135674
	Vanadium	472	500	94	90 - 110	P	05/05/2025	14:18	LB135674
	Zinc	969	1000	97	90 - 110	P	05/05/2025	14:18	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental SDG No.: Q1939
Contract: GENV01 Lab Code: CHEM Case No.: Q1939 SAS No.: Q1939
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	100	100	100	80 - 120	P	05/05/2025	14:36	LB135674
	Antimony	53.4	50.0	107	80 - 120	P	05/05/2025	14:36	LB135674
	Arsenic	23.4	20.0	117	80 - 120	P	05/05/2025	14:36	LB135674
	Barium	95.2	100	95	80 - 120	P	05/05/2025	14:36	LB135674
	Beryllium	5.87	6.0	98	80 - 120	P	05/05/2025	14:36	LB135674
	Cadmium	5.83	6.0	97	80 - 120	P	05/05/2025	14:36	LB135674
	Calcium	2030	2000	102	80 - 120	P	05/05/2025	14:36	LB135674
	Chromium	10.2	10.0	102	80 - 120	P	05/05/2025	14:36	LB135674
	Cobalt	29.7	30.0	99	80 - 120	P	05/05/2025	14:36	LB135674
	Copper	22.4	20.0	112	80 - 120	P	05/05/2025	14:36	LB135674
	Iron	90.6	100	91	80 - 120	P	05/05/2025	14:36	LB135674
	Lead	10.9	12.0	91	80 - 120	P	05/05/2025	14:36	LB135674
	Magnesium	2150	2000	107	80 - 120	P	05/05/2025	14:36	LB135674
	Manganese	18.4	20.0	92	80 - 120	P	05/05/2025	14:36	LB135674
	Nickel	39.5	40.0	99	80 - 120	P	05/05/2025	14:36	LB135674
	Potassium	1920	2000	96	80 - 120	P	05/05/2025	14:36	LB135674
	Selenium	22.6	20.0	113	80 - 120	P	05/05/2025	14:36	LB135674
	Silver	10.2	10.0	102	80 - 120	P	05/05/2025	14:36	LB135674
	Sodium	1930	2000	97	80 - 120	P	05/05/2025	14:36	LB135674
	Thallium	41.0	40.0	103	80 - 120	P	05/05/2025	14:36	LB135674
	Vanadium	39.7	40.0	99	80 - 120	P	05/05/2025	14:36	LB135674
	Zinc	41.2	40.0	103	80 - 120	P	05/05/2025	14:36	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9770	10000	98	90 - 110	P	05/05/2025	15:20	LB135674
	Antimony	4950	5000	99	90 - 110	P	05/05/2025	15:20	LB135674
	Arsenic	4930	5000	99	90 - 110	P	05/05/2025	15:20	LB135674
	Barium	9850	10000	98	90 - 110	P	05/05/2025	15:20	LB135674
	Beryllium	243	250	97	90 - 110	P	05/05/2025	15:20	LB135674
	Cadmium	2450	2500	98	90 - 110	P	05/05/2025	15:20	LB135674
	Calcium	24400	25000	98	90 - 110	P	05/05/2025	15:20	LB135674
	Chromium	996	1000	100	90 - 110	P	05/05/2025	15:20	LB135674
	Cobalt	2450	2500	98	90 - 110	P	05/05/2025	15:20	LB135674
	Copper	1240	1250	99	90 - 110	P	05/05/2025	15:20	LB135674
	Iron	5100	5000	102	90 - 110	P	05/05/2025	15:20	LB135674
	Lead	4940	5000	99	90 - 110	P	05/05/2025	15:20	LB135674
	Magnesium	24200	25000	97	90 - 110	P	05/05/2025	15:20	LB135674
	Manganese	2430	2500	97	90 - 110	P	05/05/2025	15:20	LB135674
	Nickel	2450	2500	98	90 - 110	P	05/05/2025	15:20	LB135674
	Potassium	24900	25000	100	90 - 110	P	05/05/2025	15:20	LB135674
	Selenium	4960	5000	99	90 - 110	P	05/05/2025	15:20	LB135674
	Silver	1240	1250	99	90 - 110	P	05/05/2025	15:20	LB135674
	Sodium	25000	25000	100	90 - 110	P	05/05/2025	15:20	LB135674
	Thallium	4980	5000	100	90 - 110	P	05/05/2025	15:20	LB135674
	Vanadium	2440	2500	98	90 - 110	P	05/05/2025	15:20	LB135674
	Zinc	2480	2500	99	90 - 110	P	05/05/2025	15:20	LB135674
CCV02	Aluminum	9710	10000	97	90 - 110	P	05/05/2025	16:07	LB135674
	Antimony	4910	5000	98	90 - 110	P	05/05/2025	16:07	LB135674
	Arsenic	4880	5000	98	90 - 110	P	05/05/2025	16:07	LB135674
	Barium	9880	10000	99	90 - 110	P	05/05/2025	16:07	LB135674
	Beryllium	240	250	96	90 - 110	P	05/05/2025	16:07	LB135674
	Cadmium	2410	2500	96	90 - 110	P	05/05/2025	16:07	LB135674
	Calcium	24200	25000	97	90 - 110	P	05/05/2025	16:07	LB135674
	Chromium	974	1000	97	90 - 110	P	05/05/2025	16:07	LB135674
	Cobalt	2430	2500	97	90 - 110	P	05/05/2025	16:07	LB135674
	Copper	1230	1250	98	90 - 110	P	05/05/2025	16:07	LB135674
	Iron	4990	5000	100	90 - 110	P	05/05/2025	16:07	LB135674
	Lead	4870	5000	97	90 - 110	P	05/05/2025	16:07	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24000	25000	96	90 - 110	P	05/05/2025	16:07	LB135674
	Manganese	2430	2500	97	90 - 110	P	05/05/2025	16:07	LB135674
	Nickel	2420	2500	97	90 - 110	P	05/05/2025	16:07	LB135674
	Potassium	24300	25000	97	90 - 110	P	05/05/2025	16:07	LB135674
	Selenium	4920	5000	98	90 - 110	P	05/05/2025	16:07	LB135674
	Silver	1210	1250	97	90 - 110	P	05/05/2025	16:07	LB135674
	Sodium	24400	25000	98	90 - 110	P	05/05/2025	16:07	LB135674
	Thallium	4890	5000	98	90 - 110	P	05/05/2025	16:07	LB135674
	Vanadium	2430	2500	97	90 - 110	P	05/05/2025	16:07	LB135674
	Zinc	2430	2500	97	90 - 110	P	05/05/2025	16:07	LB135674
CCV03	Aluminum	9670	10000	97	90 - 110	P	05/05/2025	16:56	LB135674
	Antimony	5080	5000	102	90 - 110	P	05/05/2025	16:56	LB135674
	Arsenic	5050	5000	101	90 - 110	P	05/05/2025	16:56	LB135674
	Barium	9910	10000	99	90 - 110	P	05/05/2025	16:56	LB135674
	Beryllium	240	250	96	90 - 110	P	05/05/2025	16:56	LB135674
	Cadmium	2410	2500	96	90 - 110	P	05/05/2025	16:56	LB135674
	Calcium	24100	25000	96	90 - 110	P	05/05/2025	16:56	LB135674
	Chromium	975	1000	98	90 - 110	P	05/05/2025	16:56	LB135674
	Cobalt	2430	2500	97	90 - 110	P	05/05/2025	16:56	LB135674
	Copper	1260	1250	100	90 - 110	P	05/05/2025	16:56	LB135674
	Iron	5080	5000	102	90 - 110	P	05/05/2025	16:56	LB135674
	Lead	4880	5000	98	90 - 110	P	05/05/2025	16:56	LB135674
	Magnesium	23800	25000	95	90 - 110	P	05/05/2025	16:56	LB135674
	Manganese	2420	2500	97	90 - 110	P	05/05/2025	16:56	LB135674
	Nickel	2420	2500	97	90 - 110	P	05/05/2025	16:56	LB135674
	Potassium	24800	25000	99	90 - 110	P	05/05/2025	16:56	LB135674
	Selenium	5160	5000	103	90 - 110	P	05/05/2025	16:56	LB135674
	Silver	1230	1250	98	90 - 110	P	05/05/2025	16:56	LB135674
	Sodium	24800	25000	99	90 - 110	P	05/05/2025	16:56	LB135674
	Thallium	4970	5000	99	90 - 110	P	05/05/2025	16:56	LB135674
	Vanadium	2420	2500	97	90 - 110	P	05/05/2025	16:56	LB135674
	Zinc	2460	2500	98	90 - 110	P	05/05/2025	16:56	LB135674
CCV04	Aluminum	9610	10000	96	90 - 110	P	05/05/2025	17:55	LB135674
	Antimony	4970	5000	99	90 - 110	P	05/05/2025	17:55	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4930	5000	99	90 - 110	P	05/05/2025	17:55	LB135674
	Barium	9690	10000	97	90 - 110	P	05/05/2025	17:55	LB135674
	Beryllium	230	250	92	90 - 110	P	05/05/2025	17:55	LB135674
	Cadmium	2420	2500	97	90 - 110	P	05/05/2025	17:55	LB135674
	Calcium	23800	25000	95	90 - 110	P	05/05/2025	17:55	LB135674
	Chromium	977	1000	98	90 - 110	P	05/05/2025	17:55	LB135674
	Cobalt	2420	2500	97	90 - 110	P	05/05/2025	17:55	LB135674
	Copper	1220	1250	98	90 - 110	P	05/05/2025	17:55	LB135674
	Iron	4990	5000	100	90 - 110	P	05/05/2025	17:55	LB135674
	Lead	4870	5000	98	90 - 110	P	05/05/2025	17:55	LB135674
	Magnesium	23600	25000	95	90 - 110	P	05/05/2025	17:55	LB135674
	Manganese	2360	2500	94	90 - 110	P	05/05/2025	17:55	LB135674
	Nickel	2420	2500	97	90 - 110	P	05/05/2025	17:55	LB135674
	Potassium	24200	25000	97	90 - 110	P	05/05/2025	17:55	LB135674
	Selenium	4980	5000	100	90 - 110	P	05/05/2025	17:55	LB135674
	Silver	1200	1250	96	90 - 110	P	05/05/2025	17:55	LB135674
	Sodium	24400	25000	98	90 - 110	P	05/05/2025	17:55	LB135674
	Thallium	4860	5000	97	90 - 110	P	05/05/2025	17:55	LB135674
	Vanadium	2410	2500	96	90 - 110	P	05/05/2025	17:55	LB135674
	Zinc	2420	2500	97	90 - 110	P	05/05/2025	17:55	LB135674
CCV05	Aluminum	9590	10000	96	90 - 110	P	05/05/2025	18:43	LB135674
	Antimony	5050	5000	101	90 - 110	P	05/05/2025	18:43	LB135674
	Arsenic	4990	5000	100	90 - 110	P	05/05/2025	18:43	LB135674
	Barium	9570	10000	96	90 - 110	P	05/05/2025	18:43	LB135674
	Beryllium	235	250	94	90 - 110	P	05/05/2025	18:43	LB135674
	Cadmium	2420	2500	97	90 - 110	P	05/05/2025	18:43	LB135674
	Calcium	23500	25000	94	90 - 110	P	05/05/2025	18:43	LB135674
	Chromium	970	1000	97	90 - 110	P	05/05/2025	18:43	LB135674
	Cobalt	2410	2500	96	90 - 110	P	05/05/2025	18:43	LB135674
	Copper	1230	1250	99	90 - 110	P	05/05/2025	18:43	LB135674
	Iron	4980	5000	100	90 - 110	P	05/05/2025	18:43	LB135674
	Lead	4860	5000	97	90 - 110	P	05/05/2025	18:43	LB135674
	Magnesium	23300	25000	93	90 - 110	P	05/05/2025	18:43	LB135674
	Manganese	2340	2500	93	90 - 110	P	05/05/2025	18:43	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2430	2500	97	90 - 110	P	05/05/2025	18:43	LB135674
	Potassium	24500	25000	98	90 - 110	P	05/05/2025	18:43	LB135674
	Selenium	5070	5000	101	90 - 110	P	05/05/2025	18:43	LB135674
	Silver	1220	1250	98	90 - 110	P	05/05/2025	18:43	LB135674
	Sodium	24800	25000	99	90 - 110	P	05/05/2025	18:43	LB135674
	Thallium	4770	5000	95	90 - 110	P	05/05/2025	18:43	LB135674
	Vanadium	2390	2500	96	90 - 110	P	05/05/2025	18:43	LB135674
	Zinc	2470	2500	99	90 - 110	P	05/05/2025	18:43	LB135674
CCV06	Aluminum	9210	10000	92	90 - 110	P	05/05/2025	19:32	LB135674
	Antimony	4780	5000	96	90 - 110	P	05/05/2025	19:32	LB135674
	Arsenic	4750	5000	95	90 - 110	P	05/05/2025	19:32	LB135674
	Barium	9680	10000	97	90 - 110	P	05/05/2025	19:32	LB135674
	Beryllium	215	250	86	90 - 110	P	05/05/2025	19:32	LB135674
	Cadmium	2340	2500	94	90 - 110	P	05/05/2025	19:32	LB135674
	Calcium	22700	25000	91	90 - 110	P	05/05/2025	19:32	LB135674
	Chromium	940	1000	94	90 - 110	P	05/05/2025	19:32	LB135674
	Cobalt	2330	2500	93	90 - 110	P	05/05/2025	19:32	LB135674
	Copper	1180	1250	94	90 - 110	P	05/05/2025	19:32	LB135674
	Iron	5030	5000	101	90 - 110	P	05/05/2025	19:32	LB135674
	Lead	4680	5000	94	90 - 110	P	05/05/2025	19:32	LB135674
	Magnesium	22100	25000	88	90 - 110	P	05/05/2025	19:32	LB135674
	Manganese	2300	2500	92	90 - 110	P	05/05/2025	19:32	LB135674
	Nickel	2360	2500	94	90 - 110	P	05/05/2025	19:32	LB135674
	Potassium	25600	25000	102	90 - 110	P	05/05/2025	19:32	LB135674
	Selenium	4800	5000	96	90 - 110	P	05/05/2025	19:32	LB135674
	Silver	1190	1250	95	90 - 110	P	05/05/2025	19:32	LB135674
	Sodium	25500	25000	102	90 - 110	P	05/05/2025	19:32	LB135674
	Thallium	4620	5000	92	90 - 110	P	05/05/2025	19:32	LB135674
	Vanadium	2340	2500	94	90 - 110	P	05/05/2025	19:32	LB135674
	Zinc	2380	2500	95	90 - 110	P	05/05/2025	19:32	LB135674
CCV07	Aluminum	9210	10000	92	90 - 110	P	05/05/2025	20:19	LB135674
	Antimony	4700	5000	94	90 - 110	P	05/05/2025	20:19	LB135674
	Arsenic	4700	5000	94	90 - 110	P	05/05/2025	20:19	LB135674
	Barium	9380	10000	94	90 - 110	P	05/05/2025	20:19	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	231	250	92	90 - 110	P	05/05/2025	20:19	LB135674
	Cadmium	2350	2500	94	90 - 110	P	05/05/2025	20:19	LB135674
	Calcium	23300	25000	93	90 - 110	P	05/05/2025	20:19	LB135674
	Chromium	955	1000	96	90 - 110	P	05/05/2025	20:19	LB135674
	Cobalt	2330	2500	93	90 - 110	P	05/05/2025	20:19	LB135674
	Copper	1170	1250	93	90 - 110	P	05/05/2025	20:19	LB135674
	Iron	4820	5000	96	90 - 110	P	05/05/2025	20:19	LB135674
	Lead	4710	5000	94	90 - 110	P	05/05/2025	20:19	LB135674
	Magnesium	23100	25000	92	90 - 110	P	05/05/2025	20:19	LB135674
	Manganese	2310	2500	92	90 - 110	P	05/05/2025	20:19	LB135674
	Nickel	2360	2500	94	90 - 110	P	05/05/2025	20:19	LB135674
	Potassium	23800	25000	95	90 - 110	P	05/05/2025	20:19	LB135674
	Selenium	4700	5000	94	90 - 110	P	05/05/2025	20:19	LB135674
	Silver	1180	1250	95	90 - 110	P	05/05/2025	20:19	LB135674
	Sodium	23800	25000	95	90 - 110	P	05/05/2025	20:19	LB135674
	Thallium	4630	5000	92	90 - 110	P	05/05/2025	20:19	LB135674
	Vanadium	2340	2500	94	90 - 110	P	05/05/2025	20:19	LB135674
	Zinc	2350	2500	94	90 - 110	P	05/05/2025	20:19	LB135674
CCV08	Aluminum	9250	10000	92	90 - 110	P	05/05/2025	21:07	LB135674
	Antimony	4760	5000	95	90 - 110	P	05/05/2025	21:07	LB135674
	Arsenic	4790	5000	96	90 - 110	P	05/05/2025	21:07	LB135674
	Barium	9390	10000	94	90 - 110	P	05/05/2025	21:07	LB135674
	Beryllium	232	250	93	90 - 110	P	05/05/2025	21:07	LB135674
	Cadmium	2380	2500	95	90 - 110	P	05/05/2025	21:07	LB135674
	Calcium	23300	25000	93	90 - 110	P	05/05/2025	21:07	LB135674
	Chromium	961	1000	96	90 - 110	P	05/05/2025	21:07	LB135674
	Cobalt	2360	2500	95	90 - 110	P	05/05/2025	21:07	LB135674
	Copper	1190	1250	95	90 - 110	P	05/05/2025	21:07	LB135674
	Iron	4870	5000	97	90 - 110	P	05/05/2025	21:07	LB135674
	Lead	4770	5000	95	90 - 110	P	05/05/2025	21:07	LB135674
	Magnesium	23300	25000	93	90 - 110	P	05/05/2025	21:07	LB135674
	Manganese	2310	2500	92	90 - 110	P	05/05/2025	21:07	LB135674
	Nickel	2380	2500	95	90 - 110	P	05/05/2025	21:07	LB135674
	Potassium	23400	25000	94	90 - 110	P	05/05/2025	21:07	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	4800	5000	96	90 - 110	P	05/05/2025	21:07	LB135674
	Silver	1180	1250	94	90 - 110	P	05/05/2025	21:07	LB135674
	Sodium	23700	25000	95	90 - 110	P	05/05/2025	21:07	LB135674
	Thallium	4670	5000	94	90 - 110	P	05/05/2025	21:07	LB135674
	Vanadium	2340	2500	94	90 - 110	P	05/05/2025	21:07	LB135674
	Zinc	2350	2500	94	90 - 110	P	05/05/2025	21:07	LB135674
CCV09	Aluminum	9270	10000	93	90 - 110	P	05/05/2025	21:59	LB135674
	Antimony	4680	5000	94	90 - 110	P	05/05/2025	21:59	LB135674
	Arsenic	4740	5000	95	90 - 110	P	05/05/2025	21:59	LB135674
	Barium	9390	10000	94	90 - 110	P	05/05/2025	21:59	LB135674
	Beryllium	242	250	97	90 - 110	P	05/05/2025	21:59	LB135674
	Cadmium	2410	2500	97	90 - 110	P	05/05/2025	21:59	LB135674
	Calcium	23600	25000	94	90 - 110	P	05/05/2025	21:59	LB135674
	Chromium	959	1000	96	90 - 110	P	05/05/2025	21:59	LB135674
	Cobalt	2380	2500	95	90 - 110	P	05/05/2025	21:59	LB135674
	Copper	1180	1250	95	90 - 110	P	05/05/2025	21:59	LB135674
	Iron	4770	5000	95	90 - 110	P	05/05/2025	21:59	LB135674
	Lead	4810	5000	96	90 - 110	P	05/05/2025	21:59	LB135674
	Magnesium	23700	25000	95	90 - 110	P	05/05/2025	21:59	LB135674
	Manganese	2370	2500	95	90 - 110	P	05/05/2025	21:59	LB135674
	Nickel	2410	2500	96	90 - 110	P	05/05/2025	21:59	LB135674
	Potassium	23200	25000	93	90 - 110	P	05/05/2025	21:59	LB135674
	Selenium	4740	5000	95	90 - 110	P	05/05/2025	21:59	LB135674
	Silver	1190	1250	95	90 - 110	P	05/05/2025	21:59	LB135674
	Sodium	23000	25000	92	90 - 110	P	05/05/2025	21:59	LB135674
	Thallium	4640	5000	93	90 - 110	P	05/05/2025	21:59	LB135674
	Vanadium	2360	2500	94	90 - 110	P	05/05/2025	21:59	LB135674
	Zinc	2320	2500	93	90 - 110	P	05/05/2025	21:59	LB135674
CCV10	Aluminum	9230	10000	92	90 - 110	P	05/05/2025	22:33	LB135674
	Antimony	4720	5000	94	90 - 110	P	05/05/2025	22:33	LB135674
	Arsenic	4740	5000	95	90 - 110	P	05/05/2025	22:33	LB135674
	Barium	9090	10000	91	90 - 110	P	05/05/2025	22:33	LB135674
	Beryllium	239	250	96	90 - 110	P	05/05/2025	22:33	LB135674
	Cadmium	2380	2500	95	90 - 110	P	05/05/2025	22:33	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental SDG No.: Q1939
Contract: GENV01 Lab Code: CHEM Case No.: Q1939 SAS No.: Q1939
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Calcium	23300	25000	93	90 - 110	P	05/05/2025	22:33	LB135674
	Chromium	967	1000	97	90 - 110	P	05/05/2025	22:33	LB135674
	Cobalt	2360	2500	94	90 - 110	P	05/05/2025	22:33	LB135674
	Copper	1180	1250	95	90 - 110	P	05/05/2025	22:33	LB135674
	Iron	4620	5000	92	90 - 110	P	05/05/2025	22:33	LB135674
	Lead	4750	5000	95	90 - 110	P	05/05/2025	22:33	LB135674
	Magnesium	23600	25000	94	90 - 110	P	05/05/2025	22:33	LB135674
	Manganese	2300	2500	92	90 - 110	P	05/05/2025	22:33	LB135674
	Nickel	2370	2500	95	90 - 110	P	05/05/2025	22:33	LB135674
	Potassium	22800	25000	91	90 - 110	P	05/05/2025	22:33	LB135674
	Selenium	4730	5000	95	90 - 110	P	05/05/2025	22:33	LB135674
	Silver	1180	1250	94	90 - 110	P	05/05/2025	22:33	LB135674
	Sodium	23300	25000	93	90 - 110	P	05/05/2025	22:33	LB135674
	Thallium	4550	5000	91	90 - 110	P	05/05/2025	22:33	LB135674
	Vanadium	2330	2500	93	90 - 110	P	05/05/2025	22:33	LB135674
	Zinc	2330	2500	93	90 - 110	P	05/05/2025	22:33	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2400	2500	96	90 - 110	P	05/06/2025	14:47	LB135689
	Antimony	972	1000	97	90 - 110	P	05/06/2025	14:47	LB135689
	Arsenic	976	1000	98	90 - 110	P	05/06/2025	14:47	LB135689
	Barium	531	520	102	90 - 110	P	05/06/2025	14:47	LB135689
	Beryllium	552	510	108	90 - 110	P	05/06/2025	14:47	LB135689
	Cadmium	498	510	98	90 - 110	P	05/06/2025	14:47	LB135689
	Calcium	9500	10000	95	90 - 110	P	05/06/2025	14:47	LB135689
	Chromium	498	520	96	90 - 110	P	05/06/2025	14:47	LB135689
	Cobalt	487	520	94	90 - 110	P	05/06/2025	14:47	LB135689
	Copper	500	510	98	90 - 110	P	05/06/2025	14:47	LB135689
	Iron	10500	10000	105	90 - 110	P	05/06/2025	14:47	LB135689
	Lead	975	1000	98	90 - 110	P	05/06/2025	14:47	LB135689
	Magnesium	5640	6000	94	90 - 110	P	05/06/2025	14:47	LB135689
	Manganese	493	520	95	90 - 110	P	05/06/2025	14:47	LB135689
	Nickel	492	530	93	90 - 110	P	05/06/2025	14:47	LB135689
	Potassium	10100	9900	102	90 - 110	P	05/06/2025	14:47	LB135689
	Selenium	1000	1000	100	90 - 110	P	05/06/2025	14:47	LB135689
	Silver	232	250	93	90 - 110	P	05/06/2025	14:47	LB135689
	Sodium	10000	10000	100	90 - 110	P	05/06/2025	14:47	LB135689
	Thallium	1000	1000	100	90 - 110	P	05/06/2025	14:47	LB135689
	Vanadium	478	500	96	90 - 110	P	05/06/2025	14:47	LB135689
	Zinc	975	1000	98	90 - 110	P	05/06/2025	14:47	LB135689

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	105	100	105	80 - 120	P	05/06/2025	15:22	LB135689
	Antimony	50.3	50.0	101	80 - 120	P	05/06/2025	15:22	LB135689
	Arsenic	18.2	20.0	91	80 - 120	P	05/06/2025	15:22	LB135689
	Barium	91.5	100	92	80 - 120	P	05/06/2025	15:22	LB135689
	Beryllium	5.90	6.0	98	80 - 120	P	05/06/2025	15:22	LB135689
	Cadmium	5.98	6.0	100	80 - 120	P	05/06/2025	15:22	LB135689
	Calcium	1930	2000	96	80 - 120	P	05/06/2025	15:22	LB135689
	Chromium	9.31	10.0	93	80 - 120	P	05/06/2025	15:22	LB135689
	Cobalt	28.6	30.0	95	80 - 120	P	05/06/2025	15:22	LB135689
	Copper	21.1	20.0	106	80 - 120	P	05/06/2025	15:22	LB135689
	Iron	98.2	100	98	80 - 120	P	05/06/2025	15:22	LB135689
	Lead	11.1	12.0	92	80 - 120	P	05/06/2025	15:22	LB135689
	Magnesium	2030	2000	102	80 - 120	P	05/06/2025	15:22	LB135689
	Manganese	19.6	20.0	98	80 - 120	P	05/06/2025	15:22	LB135689
	Nickel	38.9	40.0	97	80 - 120	P	05/06/2025	15:22	LB135689
	Potassium	1770	2000	88	80 - 120	P	05/06/2025	15:22	LB135689
	Selenium	19.1	20.0	95	80 - 120	P	05/06/2025	15:22	LB135689
	Silver	9.80	10.0	98	80 - 120	P	05/06/2025	15:22	LB135689
	Sodium	1790	2000	89	80 - 120	P	05/06/2025	15:22	LB135689
	Thallium	39.9	40.0	100	80 - 120	P	05/06/2025	15:22	LB135689
	Vanadium	38.0	40.0	95	80 - 120	P	05/06/2025	15:22	LB135689
	Zinc	40.4	40.0	101	80 - 120	P	05/06/2025	15:22	LB135689

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9630	10000	96	90 - 110	P	05/06/2025	16:29	LB135689
	Antimony	4790	5000	96	90 - 110	P	05/06/2025	16:29	LB135689
	Arsenic	4780	5000	96	90 - 110	P	05/06/2025	16:29	LB135689
	Barium	9340	10000	93	90 - 110	P	05/06/2025	16:29	LB135689
	Beryllium	249	250	100	90 - 110	P	05/06/2025	16:29	LB135689
	Cadmium	2390	2500	95	90 - 110	P	05/06/2025	16:29	LB135689
	Calcium	23900	25000	96	90 - 110	P	05/06/2025	16:29	LB135689
	Chromium	998	1000	100	90 - 110	P	05/06/2025	16:29	LB135689
	Cobalt	2390	2500	96	90 - 110	P	05/06/2025	16:29	LB135689
	Copper	1210	1250	97	90 - 110	P	05/06/2025	16:29	LB135689
	Iron	5040	5000	101	90 - 110	P	05/06/2025	16:29	LB135689
	Lead	5030	5000	101	90 - 110	P	05/06/2025	16:29	LB135689
	Magnesium	23900	25000	95	90 - 110	P	05/06/2025	16:29	LB135689
	Manganese	2400	2500	96	90 - 110	P	05/06/2025	16:29	LB135689
	Nickel	2400	2500	96	90 - 110	P	05/06/2025	16:29	LB135689
	Potassium	23900	25000	95	90 - 110	P	05/06/2025	16:29	LB135689
	Selenium	4800	5000	96	90 - 110	P	05/06/2025	16:29	LB135689
	Silver	1240	1250	99	90 - 110	P	05/06/2025	16:29	LB135689
	Sodium	24100	25000	96	90 - 110	P	05/06/2025	16:29	LB135689
	Thallium	4670	5000	94	90 - 110	P	05/06/2025	16:29	LB135689
	Vanadium	2400	2500	96	90 - 110	P	05/06/2025	16:29	LB135689
	Zinc	2470	2500	99	90 - 110	P	05/06/2025	16:29	LB135689
CCV02	Aluminum	9480	10000	95	90 - 110	P	05/06/2025	17:15	LB135689
	Antimony	4750	5000	95	90 - 110	P	05/06/2025	17:15	LB135689
	Arsenic	4730	5000	95	90 - 110	P	05/06/2025	17:15	LB135689
	Barium	9710	10000	97	90 - 110	P	05/06/2025	17:15	LB135689
	Beryllium	237	250	95	90 - 110	P	05/06/2025	17:15	LB135689
	Cadmium	2390	2500	96	90 - 110	P	05/06/2025	17:15	LB135689
	Calcium	24000	25000	96	90 - 110	P	05/06/2025	17:15	LB135689
	Chromium	962	1000	96	90 - 110	P	05/06/2025	17:15	LB135689
	Cobalt	2390	2500	96	90 - 110	P	05/06/2025	17:15	LB135689
	Copper	1200	1250	96	90 - 110	P	05/06/2025	17:15	LB135689
	Iron	5000	5000	100	90 - 110	P	05/06/2025	17:15	LB135689
	Lead	4890	5000	98	90 - 110	P	05/06/2025	17:15	LB135689

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	23700	25000	95	90 - 110	P	05/06/2025	17:15	LB135689
	Manganese	2410	2500	96	90 - 110	P	05/06/2025	17:15	LB135689
	Nickel	2390	2500	96	90 - 110	P	05/06/2025	17:15	LB135689
	Potassium	24100	25000	96	90 - 110	P	05/06/2025	17:15	LB135689
	Selenium	4750	5000	95	90 - 110	P	05/06/2025	17:15	LB135689
	Silver	1200	1250	96	90 - 110	P	05/06/2025	17:15	LB135689
	Sodium	24100	25000	96	90 - 110	P	05/06/2025	17:15	LB135689
	Thallium	4770	5000	96	90 - 110	P	05/06/2025	17:15	LB135689
	Vanadium	2390	2500	95	90 - 110	P	05/06/2025	17:15	LB135689
	Zinc	2410	2500	96	90 - 110	P	05/06/2025	17:15	LB135689
CCV03	Aluminum	9460	10000	95	90 - 110	P	05/06/2025	18:08	LB135689
	Antimony	4590	5000	92	90 - 110	P	05/06/2025	18:08	LB135689
	Arsenic	4570	5000	92	90 - 110	P	05/06/2025	18:08	LB135689
	Barium	9450	10000	94	90 - 110	P	05/06/2025	18:08	LB135689
	Beryllium	240	250	96	90 - 110	P	05/06/2025	18:08	LB135689
	Cadmium	2340	2500	94	90 - 110	P	05/06/2025	18:08	LB135689
	Calcium	23700	25000	95	90 - 110	P	05/06/2025	18:08	LB135689
	Chromium	954	1000	95	90 - 110	P	05/06/2025	18:08	LB135689
	Cobalt	2340	2500	94	90 - 110	P	05/06/2025	18:08	LB135689
	Copper	1170	1250	94	90 - 110	P	05/06/2025	18:08	LB135689
	Iron	4700	5000	94	90 - 110	P	05/06/2025	18:08	LB135689
	Lead	4690	5000	94	90 - 110	P	05/06/2025	18:08	LB135689
	Magnesium	23600	25000	94	90 - 110	P	05/06/2025	18:08	LB135689
	Manganese	2390	2500	95	90 - 110	P	05/06/2025	18:08	LB135689
	Nickel	2340	2500	94	90 - 110	P	05/06/2025	18:08	LB135689
	Potassium	23200	25000	93	90 - 110	P	05/06/2025	18:08	LB135689
	Selenium	4550	5000	91	90 - 110	P	05/06/2025	18:08	LB135689
	Silver	1180	1250	94	90 - 110	P	05/06/2025	18:08	LB135689
	Sodium	23300	25000	93	90 - 110	P	05/06/2025	18:08	LB135689
	Thallium	4710	5000	94	90 - 110	P	05/06/2025	18:08	LB135689
	Vanadium	2360	2500	94	90 - 110	P	05/06/2025	18:08	LB135689
	Zinc	2370	2500	95	90 - 110	P	05/06/2025	18:08	LB135689
CCV04	Aluminum	9960	10000	100	90 - 110	P	05/06/2025	19:16	LB135689
	Antimony	4840	5000	97	90 - 110	P	05/06/2025	19:16	LB135689

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4840	5000	97	90 - 110	P	05/06/2025	19:16	LB135689
	Barium	9930	10000	99	90 - 110	P	05/06/2025	19:16	LB135689
	Beryllium	264	250	106	90 - 110	P	05/06/2025	19:16	LB135689
	Cadmium	2470	2500	99	90 - 110	P	05/06/2025	19:16	LB135689
	Calcium	25000	25000	100	90 - 110	P	05/06/2025	19:16	LB135689
	Chromium	988	1000	99	90 - 110	P	05/06/2025	19:16	LB135689
	Cobalt	2470	2500	99	90 - 110	P	05/06/2025	19:16	LB135689
	Copper	1240	1250	99	90 - 110	P	05/06/2025	19:16	LB135689
	Iron	4820	5000	96	90 - 110	P	05/06/2025	19:16	LB135689
	Lead	4970	5000	99	90 - 110	P	05/06/2025	19:16	LB135689
	Magnesium	24900	25000	100	90 - 110	P	05/06/2025	19:16	LB135689
	Manganese	2530	2500	101	90 - 110	P	05/06/2025	19:16	LB135689
	Nickel	2470	2500	99	90 - 110	P	05/06/2025	19:16	LB135689
	Potassium	23600	25000	94	90 - 110	P	05/06/2025	19:16	LB135689
	Selenium	4820	5000	96	90 - 110	P	05/06/2025	19:16	LB135689
	Silver	1240	1250	99	90 - 110	P	05/06/2025	19:16	LB135689
	Sodium	23500	25000	94	90 - 110	P	05/06/2025	19:16	LB135689
	Thallium	4760	5000	95	90 - 110	P	05/06/2025	19:16	LB135689
	Vanadium	2470	2500	99	90 - 110	P	05/06/2025	19:16	LB135689
	Zinc	2490	2500	100	90 - 110	P	05/06/2025	19:16	LB135689
CCV05	Aluminum	10000	10000	100	90 - 110	P	05/06/2025	20:01	LB135689
	Antimony	4800	5000	96	90 - 110	P	05/06/2025	20:01	LB135689
	Arsenic	4780	5000	96	90 - 110	P	05/06/2025	20:01	LB135689
	Barium	9940	10000	99	90 - 110	P	05/06/2025	20:01	LB135689
	Beryllium	260	250	104	90 - 110	P	05/06/2025	20:01	LB135689
	Cadmium	2460	2500	99	90 - 110	P	05/06/2025	20:01	LB135689
	Calcium	25000	25000	100	90 - 110	P	05/06/2025	20:01	LB135689
	Chromium	995	1000	100	90 - 110	P	05/06/2025	20:01	LB135689
	Cobalt	2450	2500	98	90 - 110	P	05/06/2025	20:01	LB135689
	Copper	1230	1250	98	90 - 110	P	05/06/2025	20:01	LB135689
	Iron	4860	5000	97	90 - 110	P	05/06/2025	20:01	LB135689
	Lead	4900	5000	98	90 - 110	P	05/06/2025	20:01	LB135689
	Magnesium	25000	25000	100	90 - 110	P	05/06/2025	20:01	LB135689
	Manganese	2520	2500	101	90 - 110	P	05/06/2025	20:01	LB135689

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2450	2500	98	90 - 110	P	05/06/2025	20:01	LB135689
	Potassium	23700	25000	95	90 - 110	P	05/06/2025	20:01	LB135689
	Selenium	4730	5000	95	90 - 110	P	05/06/2025	20:01	LB135689
	Silver	1230	1250	99	90 - 110	P	05/06/2025	20:01	LB135689
	Sodium	23700	25000	95	90 - 110	P	05/06/2025	20:01	LB135689
	Thallium	4800	5000	96	90 - 110	P	05/06/2025	20:01	LB135689
	Vanadium	2470	2500	99	90 - 110	P	05/06/2025	20:01	LB135689
	Zinc	2480	2500	99	90 - 110	P	05/06/2025	20:01	LB135689
CCV06	Aluminum	9830	10000	98	90 - 110	P	05/06/2025	20:31	LB135689
	Antimony	4880	5000	98	90 - 110	P	05/06/2025	20:31	LB135689
	Arsenic	4840	5000	97	90 - 110	P	05/06/2025	20:31	LB135689
	Barium	9850	10000	98	90 - 110	P	05/06/2025	20:31	LB135689
	Beryllium	241	250	96	90 - 110	P	05/06/2025	20:31	LB135689
	Cadmium	2440	2500	98	90 - 110	P	05/06/2025	20:31	LB135689
	Calcium	24400	25000	98	90 - 110	P	05/06/2025	20:31	LB135689
	Chromium	975	1000	98	90 - 110	P	05/06/2025	20:31	LB135689
	Cobalt	2430	2500	97	90 - 110	P	05/06/2025	20:31	LB135689
	Copper	1240	1250	99	90 - 110	P	05/06/2025	20:31	LB135689
	Iron	4920	5000	98	90 - 110	P	05/06/2025	20:31	LB135689
	Lead	4870	5000	97	90 - 110	P	05/06/2025	20:31	LB135689
	Magnesium	24300	25000	97	90 - 110	P	05/06/2025	20:31	LB135689
	Manganese	2440	2500	98	90 - 110	P	05/06/2025	20:31	LB135689
	Nickel	2430	2500	97	90 - 110	P	05/06/2025	20:31	LB135689
	Potassium	24200	25000	97	90 - 110	P	05/06/2025	20:31	LB135689
	Selenium	4820	5000	96	90 - 110	P	05/06/2025	20:31	LB135689
	Silver	1210	1250	97	90 - 110	P	05/06/2025	20:31	LB135689
	Sodium	24400	25000	98	90 - 110	P	05/06/2025	20:31	LB135689
	Thallium	4750	5000	95	90 - 110	P	05/06/2025	20:31	LB135689
	Vanadium	2430	2500	97	90 - 110	P	05/06/2025	20:31	LB135689
	Zinc	2440	2500	98	90 - 110	P	05/06/2025	20:31	LB135689

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2310	2500	92	90 - 110	P	05/07/2025	12:59	LB135699
	Antimony	965	1000	96	90 - 110	P	05/07/2025	12:59	LB135699
	Arsenic	968	1000	97	90 - 110	P	05/07/2025	12:59	LB135699
	Barium	534	520	103	90 - 110	P	05/07/2025	12:59	LB135699
	Beryllium	515	510	101	90 - 110	P	05/07/2025	12:59	LB135699
	Cadmium	501	510	98	90 - 110	P	05/07/2025	12:59	LB135699
	Calcium	9400	10000	94	90 - 110	P	05/07/2025	12:59	LB135699
	Chromium	503	520	97	90 - 110	P	05/07/2025	12:59	LB135699
	Cobalt	488	520	94	90 - 110	P	05/07/2025	12:59	LB135699
	Copper	495	510	97	90 - 110	P	05/07/2025	12:59	LB135699
	Iron	11000	10000	110	90 - 110	P	05/07/2025	12:59	LB135699
	Lead	980	1000	98	90 - 110	P	05/07/2025	12:59	LB135699
	Magnesium	5440	6000	91	90 - 110	P	05/07/2025	12:59	LB135699
	Manganese	488	520	94	90 - 110	P	05/07/2025	12:59	LB135699
	Nickel	492	530	93	90 - 110	P	05/07/2025	12:59	LB135699
	Potassium	10600	9900	107	90 - 110	P	05/07/2025	12:59	LB135699
	Selenium	996	1000	100	90 - 110	P	05/07/2025	12:59	LB135699
	Silver	231	250	93	90 - 110	P	05/07/2025	12:59	LB135699
	Sodium	10100	10000	101	90 - 110	P	05/07/2025	12:59	LB135699
	Thallium	1020	1000	102	90 - 110	P	05/07/2025	12:59	LB135699
	Vanadium	466	500	93	90 - 110	P	05/07/2025	12:59	LB135699
	Zinc	970	1000	97	90 - 110	P	05/07/2025	12:59	LB135699

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental SDG No.: Q1939
Contract: GENV01 Lab Code: CHEM Case No.: Q1939 SAS No.: Q1939
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	101	100	101	80 - 120	P	05/07/2025	13:05	LB135699
	Antimony	50.9	50.0	102	80 - 120	P	05/07/2025	13:05	LB135699
	Arsenic	19.8	20.0	99	80 - 120	P	05/07/2025	13:05	LB135699
	Barium	86.9	100	87	80 - 120	P	05/07/2025	13:05	LB135699
	Beryllium	5.69	6.0	95	80 - 120	P	05/07/2025	13:05	LB135699
	Cadmium	6.10	6.0	102	80 - 120	P	05/07/2025	13:05	LB135699
	Calcium	1950	2000	97	80 - 120	P	05/07/2025	13:05	LB135699
	Chromium	9.59	10.0	96	80 - 120	P	05/07/2025	13:05	LB135699
	Cobalt	28.7	30.0	96	80 - 120	P	05/07/2025	13:05	LB135699
	Copper	21.5	20.0	108	80 - 120	P	05/07/2025	13:05	LB135699
	Iron	97.0	100	97	80 - 120	P	05/07/2025	13:05	LB135699
	Lead	12.8	12.0	106	80 - 120	P	05/07/2025	13:05	LB135699
	Magnesium	2040	2000	102	80 - 120	P	05/07/2025	13:05	LB135699
	Manganese	22.5	20.0	113	80 - 120	P	05/07/2025	13:05	LB135699
	Nickel	38.9	40.0	97	80 - 120	P	05/07/2025	13:05	LB135699
	Potassium	1880	2000	94	80 - 120	P	05/07/2025	13:05	LB135699
	Selenium	18.3	20.0	91	80 - 120	P	05/07/2025	13:05	LB135699
	Silver	10.3	10.0	103	80 - 120	P	05/07/2025	13:05	LB135699
	Sodium	1790	2000	89	80 - 120	P	05/07/2025	13:05	LB135699
	Thallium	43.2	40.0	108	80 - 120	P	05/07/2025	13:05	LB135699
	Vanadium	40.8	40.0	102	80 - 120	P	05/07/2025	13:05	LB135699
	Zinc	41.0	40.0	102	80 - 120	P	05/07/2025	13:05	LB135699

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10100	10000	101	90 - 110	P	05/07/2025	13:39	LB135699
	Antimony	5020	5000	100	90 - 110	P	05/07/2025	13:39	LB135699
	Arsenic	5080	5000	102	90 - 110	P	05/07/2025	13:39	LB135699
	Barium	9770	10000	98	90 - 110	P	05/07/2025	13:39	LB135699
	Beryllium	267	250	107	90 - 110	P	05/07/2025	13:39	LB135699
	Cadmium	2570	2500	103	90 - 110	P	05/07/2025	13:39	LB135699
	Calcium	25600	25000	102	90 - 110	P	05/07/2025	13:39	LB135699
	Chromium	1050	1000	104	90 - 110	P	05/07/2025	13:39	LB135699
	Cobalt	2550	2500	102	90 - 110	P	05/07/2025	13:39	LB135699
	Copper	1280	1250	102	90 - 110	P	05/07/2025	13:39	LB135699
	Iron	4960	5000	99	90 - 110	P	05/07/2025	13:39	LB135699
	Lead	5150	5000	103	90 - 110	P	05/07/2025	13:39	LB135699
	Magnesium	25600	25000	102	90 - 110	P	05/07/2025	13:39	LB135699
	Manganese	2520	2500	101	90 - 110	P	05/07/2025	13:39	LB135699
	Nickel	2570	2500	103	90 - 110	P	05/07/2025	13:39	LB135699
	Potassium	24200	25000	97	90 - 110	P	05/07/2025	13:39	LB135699
	Selenium	5050	5000	101	90 - 110	P	05/07/2025	13:39	LB135699
	Silver	1280	1250	102	90 - 110	P	05/07/2025	13:39	LB135699
	Sodium	23900	25000	96	90 - 110	P	05/07/2025	13:39	LB135699
	Thallium	5090	5000	102	90 - 110	P	05/07/2025	13:39	LB135699
	Vanadium	2530	2500	101	90 - 110	P	05/07/2025	13:39	LB135699
	Zinc	2540	2500	102	90 - 110	P	05/07/2025	13:39	LB135699
CCV02	Aluminum	10100	10000	102	90 - 110	P	05/07/2025	14:35	LB135699
	Antimony	5040	5000	101	90 - 110	P	05/07/2025	14:35	LB135699
	Arsenic	5050	5000	101	90 - 110	P	05/07/2025	14:35	LB135699
	Barium	10000	10000	100	90 - 110	P	05/07/2025	14:35	LB135699
	Beryllium	260	250	104	90 - 110	P	05/07/2025	14:35	LB135699
	Cadmium	2490	2500	100	90 - 110	P	05/07/2025	14:35	LB135699
	Calcium	24900	25000	100	90 - 110	P	05/07/2025	14:35	LB135699
	Chromium	1020	1000	102	90 - 110	P	05/07/2025	14:35	LB135699
	Cobalt	2490	2500	100	90 - 110	P	05/07/2025	14:35	LB135699
	Copper	1270	1250	101	90 - 110	P	05/07/2025	14:35	LB135699
	Iron	4870	5000	97	90 - 110	P	05/07/2025	14:35	LB135699
	Lead	5030	5000	101	90 - 110	P	05/07/2025	14:35	LB135699

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24800	25000	99	90 - 110	P	05/07/2025	14:35	LB135699
	Manganese	2470	2500	99	90 - 110	P	05/07/2025	14:35	LB135699
	Nickel	2500	2500	100	90 - 110	P	05/07/2025	14:35	LB135699
	Potassium	24200	25000	97	90 - 110	P	05/07/2025	14:35	LB135699
	Selenium	5040	5000	101	90 - 110	P	05/07/2025	14:35	LB135699
	Silver	1270	1250	101	90 - 110	P	05/07/2025	14:35	LB135699
	Sodium	24100	25000	96	90 - 110	P	05/07/2025	14:35	LB135699
	Thallium	5120	5000	102	90 - 110	P	05/07/2025	14:35	LB135699
	Vanadium	2490	2500	100	90 - 110	P	05/07/2025	14:35	LB135699
	Zinc	2540	2500	102	90 - 110	P	05/07/2025	14:35	LB135699
CCV03	Aluminum	9790	10000	98	90 - 110	P	05/07/2025	15:23	LB135699
	Antimony	4870	5000	97	90 - 110	P	05/07/2025	15:23	LB135699
	Arsenic	4880	5000	98	90 - 110	P	05/07/2025	15:23	LB135699
	Barium	9500	10000	95	90 - 110	P	05/07/2025	15:23	LB135699
	Beryllium	256	250	102	90 - 110	P	05/07/2025	15:23	LB135699
	Cadmium	2420	2500	97	90 - 110	P	05/07/2025	15:23	LB135699
	Calcium	24500	25000	98	90 - 110	P	05/07/2025	15:23	LB135699
	Chromium	993	1000	99	90 - 110	P	05/07/2025	15:23	LB135699
	Cobalt	2420	2500	97	90 - 110	P	05/07/2025	15:23	LB135699
	Copper	1220	1250	98	90 - 110	P	05/07/2025	15:23	LB135699
	Iron	4820	5000	96	90 - 110	P	05/07/2025	15:23	LB135699
	Lead	4870	5000	97	90 - 110	P	05/07/2025	15:23	LB135699
	Magnesium	24500	25000	98	90 - 110	P	05/07/2025	15:23	LB135699
	Manganese	2430	2500	97	90 - 110	P	05/07/2025	15:23	LB135699
	Nickel	2430	2500	97	90 - 110	P	05/07/2025	15:23	LB135699
	Potassium	23400	25000	94	90 - 110	P	05/07/2025	15:23	LB135699
	Selenium	4900	5000	98	90 - 110	P	05/07/2025	15:23	LB135699
	Silver	1230	1250	98	90 - 110	P	05/07/2025	15:23	LB135699
	Sodium	23100	25000	92	90 - 110	P	05/07/2025	15:23	LB135699
	Thallium	5080	5000	102	90 - 110	P	05/07/2025	15:23	LB135699
	Vanadium	2430	2500	97	90 - 110	P	05/07/2025	15:23	LB135699
	Zinc	2450	2500	98	90 - 110	P	05/07/2025	15:23	LB135699
CCV04	Aluminum	9600	10000	96	90 - 110	P	05/07/2025	16:05	LB135699
	Antimony	4860	5000	97	90 - 110	P	05/07/2025	16:05	LB135699

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental SDG No.: Q1939
Contract: GENV01 Lab Code: CHEM Case No.: Q1939 SAS No.: Q1939
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4850	5000	97	90 - 110	P	05/07/2025	16:05	LB135699
	Barium	9490	10000	95	90 - 110	P	05/07/2025	16:05	LB135699
	Beryllium	236	250	94	90 - 110	P	05/07/2025	16:05	LB135699
	Cadmium	2400	2500	96	90 - 110	P	05/07/2025	16:05	LB135699
	Calcium	24100	25000	96	90 - 110	P	05/07/2025	16:05	LB135699
	Chromium	968	1000	97	90 - 110	P	05/07/2025	16:05	LB135699
	Cobalt	2400	2500	96	90 - 110	P	05/07/2025	16:05	LB135699
	Copper	1220	1250	97	90 - 110	P	05/07/2025	16:05	LB135699
	Iron	4800	5000	96	90 - 110	P	05/07/2025	16:05	LB135699
	Lead	4840	5000	97	90 - 110	P	05/07/2025	16:05	LB135699
	Magnesium	23700	25000	95	90 - 110	P	05/07/2025	16:05	LB135699
	Manganese	2380	2500	95	90 - 110	P	05/07/2025	16:05	LB135699
	Nickel	2410	2500	96	90 - 110	P	05/07/2025	16:05	LB135699
	Potassium	23300	25000	93	90 - 110	P	05/07/2025	16:05	LB135699
	Selenium	4860	5000	97	90 - 110	P	05/07/2025	16:05	LB135699
	Silver	1180	1250	95	90 - 110	P	05/07/2025	16:05	LB135699
	Sodium	23000	25000	92	90 - 110	P	05/07/2025	16:05	LB135699
	Thallium	5120	5000	102	90 - 110	P	05/07/2025	16:05	LB135699
	Vanadium	2410	2500	96	90 - 110	P	05/07/2025	16:05	LB135699
	Zinc	2390	2500	96	90 - 110	P	05/07/2025	16:05	LB135699
CCV05	Aluminum	9820	10000	98	90 - 110	P	05/07/2025	16:47	LB135699
	Antimony	5170	5000	103	90 - 110	P	05/07/2025	16:47	LB135699
	Arsenic	5140	5000	103	90 - 110	P	05/07/2025	16:47	LB135699
	Barium	9560	10000	96	90 - 110	P	05/07/2025	16:47	LB135699
	Beryllium	233	250	93	90 - 110	P	05/07/2025	16:47	LB135699
	Cadmium	2360	2500	95	90 - 110	P	05/07/2025	16:47	LB135699
	Calcium	23500	25000	94	90 - 110	P	05/07/2025	16:47	LB135699
	Chromium	965	1000	96	90 - 110	P	05/07/2025	16:47	LB135699
	Cobalt	2390	2500	95	90 - 110	P	05/07/2025	16:47	LB135699
	Copper	1280	1250	102	90 - 110	P	05/07/2025	16:47	LB135699
	Iron	4960	5000	99	90 - 110	P	05/07/2025	16:47	LB135699
	Lead	4800	5000	96	90 - 110	P	05/07/2025	16:47	LB135699
	Magnesium	22800	25000	91	90 - 110	P	05/07/2025	16:47	LB135699
	Manganese	2330	2500	93	90 - 110	P	05/07/2025	16:47	LB135699

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q1939
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2410	2500	96	90 - 110	P	05/07/2025	16:47	LB135699
	Potassium	25000	25000	100	90 - 110	P	05/07/2025	16:47	LB135699
	Selenium	5200	5000	104	90 - 110	P	05/07/2025	16:47	LB135699
	Silver	1220	1250	98	90 - 110	P	05/07/2025	16:47	LB135699
	Sodium	24900	25000	100	90 - 110	P	05/07/2025	16:47	LB135699
	Thallium	5060	5000	101	90 - 110	P	05/07/2025	16:47	LB135699
	Vanadium	2400	2500	96	90 - 110	P	05/07/2025	16:47	LB135699
	Zinc	2480	2500	99	90 - 110	P	05/07/2025	16:47	LB135699
