



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental SDG No.: Q1939
Contract: GENV01 Lab Code: CHEM Case No.: Q1939 SAS No.: Q1939
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.23	0.2	113	40 - 160	CV	05/05/2025	08:39	LB135655
CRI01	Aluminum	99.6	100	100	40 - 160	P	05/05/2025	14:47	LB135674
	Antimony	52.8	50.0	106	40 - 160	P	05/05/2025	14:47	LB135674
	Arsenic	20.6	20.0	103	40 - 160	P	05/05/2025	14:47	LB135674
	Barium	92.2	100	92	40 - 160	P	05/05/2025	14:47	LB135674
	Beryllium	5.85	6.0	98	40 - 160	P	05/05/2025	14:47	LB135674
	Cadmium	5.79	6.0	96	40 - 160	P	05/05/2025	14:47	LB135674
	Calcium	2000	2000	100	40 - 160	P	05/05/2025	14:47	LB135674
	Chromium	10.3	10.0	103	40 - 160	P	05/05/2025	14:47	LB135674
	Cobalt	29.6	30.0	99	40 - 160	P	05/05/2025	14:47	LB135674
	Copper	22.4	20.0	112	40 - 160	P	05/05/2025	14:47	LB135674
	Iron	93.9	100	94	40 - 160	P	05/05/2025	14:47	LB135674
	Lead	11.2	12.0	93	40 - 160	P	05/05/2025	14:47	LB135674
	Magnesium	2100	2000	105	40 - 160	P	05/05/2025	14:47	LB135674
	Manganese	18.0	20.0	90	40 - 160	P	05/05/2025	14:47	LB135674
	Nickel	39.9	40.0	100	40 - 160	P	05/05/2025	14:47	LB135674
	Potassium	2020	2000	101	40 - 160	P	05/05/2025	14:47	LB135674
	Selenium	22.8	20.0	114	40 - 160	P	05/05/2025	14:47	LB135674
	Silver	10.3	10.0	103	40 - 160	P	05/05/2025	14:47	LB135674
	Sodium	2000	2000	100	40 - 160	P	05/05/2025	14:47	LB135674
	Thallium	42.8	40.0	107	40 - 160	P	05/05/2025	14:47	LB135674
	Vanadium	41.7	40.0	104	40 - 160	P	05/05/2025	14:47	LB135674
	Zinc	42.7	40.0	107	40 - 160	P	05/05/2025	14:47	LB135674
CRI01	Aluminum	102	100	102	40 - 160	P	05/06/2025	15:31	LB135689
	Antimony	50.4	50.0	101	40 - 160	P	05/06/2025	15:31	LB135689
	Arsenic	20.3	20.0	101	40 - 160	P	05/06/2025	15:31	LB135689
	Barium	89.4	100	89	40 - 160	P	05/06/2025	15:31	LB135689
	Beryllium	5.91	6.0	98	40 - 160	P	05/06/2025	15:31	LB135689
	Cadmium	5.91	6.0	98	40 - 160	P	05/06/2025	15:31	LB135689
	Calcium	1950	2000	97	40 - 160	P	05/06/2025	15:31	LB135689
	Chromium	9.54	10.0	95	40 - 160	P	05/06/2025	15:31	LB135689
	Cobalt	28.3	30.0	94	40 - 160	P	05/06/2025	15:31	LB135689
	Copper	21.6	20.0	108	40 - 160	P	05/06/2025	15:31	LB135689
	Iron	95.6	100	96	40 - 160	P	05/06/2025	15:31	LB135689
	Lead	10.0	12.0	83	40 - 160	P	05/06/2025	15:31	LB135689

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental **SDG No.:** Q1939

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1939 **SAS No.:** Q1939

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Magnesium	2050	2000	102	40 - 160	P	05/06/2025	15:31	LB135689
	Manganese	19.8	20.0	99	40 - 160	P	05/06/2025	15:31	LB135689
	Nickel	38.6	40.0	97	40 - 160	P	05/06/2025	15:31	LB135689
	Potassium	1730	2000	87	40 - 160	P	05/06/2025	15:31	LB135689
	Selenium	22.5	20.0	112	40 - 160	P	05/06/2025	15:31	LB135689
	Silver	9.66	10.0	97	40 - 160	P	05/06/2025	15:31	LB135689
	Sodium	1760	2000	88	40 - 160	P	05/06/2025	15:31	LB135689
	Thallium	37.5	40.0	94	40 - 160	P	05/06/2025	15:31	LB135689
	Vanadium	39.4	40.0	98	40 - 160	P	05/06/2025	15:31	LB135689
	Zinc	39.7	40.0	99	40 - 160	P	05/06/2025	15:31	LB135689
CRI01	Aluminum	95.9	100	96	40 - 160	P	05/07/2025	13:16	LB135699
	Antimony	51.7	50.0	103	40 - 160	P	05/07/2025	13:16	LB135699
	Arsenic	19.7	20.0	98	40 - 160	P	05/07/2025	13:16	LB135699
	Barium	88.3	100	88	40 - 160	P	05/07/2025	13:16	LB135699
	Beryllium	5.55	6.0	92	40 - 160	P	05/07/2025	13:16	LB135699
	Cadmium	6.09	6.0	102	40 - 160	P	05/07/2025	13:16	LB135699
	Calcium	1950	2000	98	40 - 160	P	05/07/2025	13:16	LB135699
	Chromium	9.49	10.0	95	40 - 160	P	05/07/2025	13:16	LB135699
	Cobalt	29.1	30.0	97	40 - 160	P	05/07/2025	13:16	LB135699
	Copper	21.5	20.0	108	40 - 160	P	05/07/2025	13:16	LB135699
	Iron	96.4	100	96	40 - 160	P	05/07/2025	13:16	LB135699
	Lead	13.4	12.0	112	40 - 160	P	05/07/2025	13:16	LB135699
	Magnesium	2010	2000	100	40 - 160	P	05/07/2025	13:16	LB135699
	Manganese	22.3	20.0	112	40 - 160	P	05/07/2025	13:16	LB135699
	Nickel	39.2	40.0	98	40 - 160	P	05/07/2025	13:16	LB135699
	Potassium	1850	2000	93	40 - 160	P	05/07/2025	13:16	LB135699
	Selenium	16.5	20.0	82	40 - 160	P	05/07/2025	13:16	LB135699
	Silver	10.2	10.0	102	40 - 160	P	05/07/2025	13:16	LB135699
	Sodium	1790	2000	90	40 - 160	P	05/07/2025	13:16	LB135699
	Thallium	43.0	40.0	108	40 - 160	P	05/07/2025	13:16	LB135699
	Vanadium	40.2	40.0	101	40 - 160	P	05/07/2025	13:16	LB135699
	Zinc	40.3	40.0	101	40 - 160	P	05/07/2025	13:16	LB135699