

Continuing Calibration Report for SequenceID : FC050225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068758.D

Aliphatic C9-C12	5796149.000	60.000	3.343	6.645	96602.483	105282.003	8.244
Aliphatic C12-C16	3655219.000	40.000	6.646	10.046	91380.475	102322.575	10.694
Aliphatic C16-C21	5276212.000	60.000	10.047	13.415	87936.867	98816.539	11.010
Aliphatic C21-C28	6971161.000	80.000	13.416	17.079	87139.513	94185.337	7.481
Aliphatic C28-C40	11514316.000	120.000	17.080	22.085	95952.633	89882.835	-6.753
Aliphatic EPH	33213057.000	360.000	3.343	22.085	92258.492	96276.618	4.174