

Continuing Calibration Report for SequenceID : FC050525AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FC068761.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 6153886.000  | 60.000  | 3.342  | 6.643  | 102564.767 | 105282.003 | 2.581  |
| Aliphatic C12-C16 | 4011540.000  | 40.000  | 6.644  | 10.043 | 100288.500 | 102322.575 | 1.988  |
| Aliphatic C16-C21 | 5821248.000  | 60.000  | 10.044 | 13.410 | 97020.800  | 98816.539  | 1.817  |
| Aliphatic C21-C28 | 7504771.000  | 80.000  | 13.411 | 17.073 | 93809.638  | 94185.337  | 0.399  |
| Aliphatic C28-C40 | 11707657.000 | 120.000 | 17.074 | 22.072 | 97563.808  | 89882.835  | -8.546 |
| Aliphatic EPH     | 35199102.000 | 360.000 | 3.342  | 22.072 | 97775.283  | 96276.618  | -1.557 |