

Continuing Calibration Report for SequenceID : FE050225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE053618.D

Aliphatic C9-C12	8509712.000	60.000	3.114	6.755	141828.533	138562.354	-2.357
Aliphatic C12-C16	5638130.000	40.000	6.756	10.204	140953.250	141806.081	0.601
Aliphatic C16-C21	8230190.000	60.000	10.205	13.579	137169.833	145363.765	5.637
Aliphatic C21-C28	10288413.000	80.000	13.580	17.248	128605.163	142174.626	9.544
Aliphatic C28-C40	14723785.000	120.000	17.249	22.135	122698.208	129049.185	4.921
Aliphatic EPH	47390230.000	360.000	3.114	22.135	131639.528	137688.007	4.393