

Continuing Calibration Report for SequenceID : FE050225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE053633.D

Aliphatic C9-C12	7523288.000	60.000	3.114	6.755	125388.133	138562.354	9.508
Aliphatic C12-C16	5185546.000	40.000	6.756	10.204	129638.650	141806.081	8.580
Aliphatic C16-C21	7773718.000	60.000	10.205	13.579	129561.967	145363.765	10.871
Aliphatic C21-C28	10140274.000	80.000	13.580	17.248	126753.425	142174.626	10.847
Aliphatic C28-C40	15671059.000	120.000	17.249	22.135	130592.158	129049.185	-1.196
Aliphatic EPH	46293885.000	360.000	3.114	22.135	128594.125	137688.007	6.605