

**SOLID EPH\_NF LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY**

**Lab Name:** Chemtech **Client:** G Environmental  
**Lab Code:** CHEM **Cas No:** Q1939 **SAS No :** Q1939 **SDG No:** Q1939  
**Sample No :** PB167833BS **Datafile:** FC068750.D  
**Client ID :** PB167833BS

| COMPOUND          | SPIKE<br>ADDED<br>mg/kg | LCS/LCSD<br>CONCENTRATION<br>mg/kg | % REC | Qual | QC LIMITS |
|-------------------|-------------------------|------------------------------------|-------|------|-----------|
| Aliphatic C28-C40 | 30.0                    | 26.5                               | 88    |      | (40-140)  |
| Aliphatic C9-C28  | 99.9                    | 95.3                               | 94    |      | (40-140)  |