

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN042825

Review By	Rahul	Review On	4/29/2025 8:54:47 AM
Supervise By	Jagrut	Supervise On	4/29/2025 11:57:54 AM
SubDirectory	BN042825	HP Acquire Method	BNA_N, 8270_HP Processing Method BN042825

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6757 SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6735 SP6740,1ul/100ul sample SP6768

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN036922.D	28 Apr 2025 10:56		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN036923.D	28 Apr 2025 11:35	Compound#20 removed from 0.1 ppm	RC/JU	Ok,M
3	SSTDICC0.2	SSTDICC0.2	BN036924.D	28 Apr 2025 12:11		RC/JU	Ok,M
4	SSTDICCC0.4	SSTDICCC0.4	BN036925.D	28 Apr 2025 12:47		RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN036926.D	28 Apr 2025 13:24		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN036927.D	28 Apr 2025 14:00		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN036928.D	28 Apr 2025 14:36		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN036929.D	28 Apr 2025 15:12		RC/JU	Ok
9	SSTDICV0.4	ICVBN042825	BN036930.D	28 Apr 2025 15:51		RC/JU	Ok
10	PB167430BL	PB167430BL	BN036931.D	28 Apr 2025 17:39	Internal Standard Fail, Analyzed for contamination check	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN050825

Review By	Rahul	Review On	5/9/2025 9:44:58 AM
Supervise By	Jagrut	Supervise On	5/9/2025 3:46:05 PM
SubDirectory	BN050825	HP Acquire Method	BNA_N, 8270_HP Processing Method BN042825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731
CCC	SP6735
Internal Standard/PEM	SP6740,1ul/100ul sample
ICV/I.BLK	SP6768
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN036971.D	08 May 2025 13:49		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN036972.D	08 May 2025 14:28		RC/JU	Ok
3	PB167915BL	PB167915BL	BN036973.D	08 May 2025 15:04		RC/JU	Ok
4	Q1939-04	GB2B	BN036974.D	08 May 2025 15:40		RC/JU	Ok
5	Q1939-04MS	GB2BMS	BN036975.D	08 May 2025 16:16		RC/JU	Ok
6	Q1939-04MSD	GB2BMSD	BN036976.D	08 May 2025 16:53		RC/JU	Ok
7	Q1939-02	GB1B	BN036977.D	08 May 2025 17:29		RC/JU	Ok
8	Q1939-06	GB3B	BN036978.D	08 May 2025 18:05	Internal Standard Fail	RC/JU	Ok,M
9	PB167915BS	PB167915BS	BN036979.D	08 May 2025 18:41		RC/JU	Ok
10	SSTDCCC0.4	SSTDCCC0.4EC	BN036980.D	08 May 2025 19:18		RC/JU	Ok

M : Manual Integration