

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q1939 SAS No.: Q1939 SDG No.: Q1939

Instrument ID: MSVOA_N Calibration Date/Time: 05/05/2025 11:32

Lab File ID: VN086478.D Init. Calib. Date(s): 04/15/2025 04/15/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:29 14:29

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.593	0.536		-9.61	20
Chloromethane	0.861	0.805	0.1	-6.5	20
Vinyl Chloride	0.822	0.759		-7.66	20
Bromomethane	0.370	0.461		24.59	20
Chloroethane	0.550	0.523		-4.91	20
Trichlorofluoromethane	0.911	0.869		-4.61	20
1,1,2-Trichlorotrifluoroethane	0.551	0.537		-2.54	20
1,1-Dichloroethene	0.588	0.557		-5.27	20
Acetone	0.290	0.278		-4.14	20
Carbon Disulfide	1.762	1.533		-13	20
Methyl tert-butyl Ether	2.163	2.360		9.11	20
Methyl Acetate	0.870	0.830		-4.6	20
Methylene Chloride	0.670	0.729		8.81	20
trans-1,2-Dichloroethene	0.615	0.632		2.76	20
1,1-Dichloroethane	1.201	1.259	0.1	4.83	20
Cyclohexane	1.175	0.999		-14.98	20
2-Butanone	0.451	0.469		3.99	20
Carbon Tetrachloride	0.441	0.442		0.23	20
cis-1,2-Dichloroethene	0.764	0.829		8.51	20
Bromochloromethane	0.509	0.533		4.72	20
Chloroform	1.180	1.306		10.68	20
1,1,1-Trichloroethane	1.009	1.049		3.96	20
Methylcyclohexane	0.551	0.492		-10.71	20
Benzene	1.485	1.541		3.77	20
1,2-Dichloroethane	0.474	0.533		12.45	20
Trichloroethene	0.352	0.372		5.68	20
1,2-Dichloropropane	0.364	0.382		4.95	20
Bromodichloromethane	0.500	0.559		11.8	20
4-Methyl-2-Pentanone	0.500	0.527		5.4	20
Toluene	0.927	0.979		5.61	20
t-1,3-Dichloropropene	0.558	0.627		12.37	20
cis-1,3-Dichloropropene	0.613	0.648		5.71	20
1,1,2-Trichloroethane	0.333	0.382		14.72	20
2-Hexanone	0.371	0.388		4.58	20
Dibromochloromethane	0.366	0.426		16.39	20
1,2-Dibromoethane	0.336	0.394		17.26	20
Tetrachloroethene	0.385	0.386		0.26	20
Chlorobenzene	1.116	1.173	0.3	5.11	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q1939 SAS No.: Q1939 SDG No.: Q1939

Instrument ID: MSVOA_N Calibration Date/Time: 05/05/2025 11:32

Lab File ID: VN086478.D Init. Calib. Date(s): 04/15/2025 04/15/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:29 14:29

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	2.014	1.999		-0.75	20
m/p-Xylenes	0.754	0.772		2.39	20
o-Xylene	0.746	0.767		2.82	20
Styrene	1.243	1.334		7.32	20
Bromoform	0.277	0.312	0.1	12.64	20
Isopropylbenzene	4.090	3.762		-8.02	20
1,1,2,2-Tetrachloroethane	1.176	1.187	0.3	0.94	20
1,3-Dichlorobenzene	1.704	1.787		4.87	20
1,4-Dichlorobenzene	1.716	1.810		5.48	20
1,2-Dichlorobenzene	1.652	1.755		6.24	20
1,2-Dibromo-3-Chloropropane	0.237	0.233		-1.69	20
1,2,4-Trichlorobenzene	0.791	0.848		7.21	20
1,2,3-Trichlorobenzene	0.749	0.800		6.81	20
1,2-Dichloroethane-d4	0.725	0.817		12.69	20
Dibromofluoromethane	0.232	0.264		13.79	20
Toluene-d8	1.240	1.280		3.23	20
4-Bromofluorobenzene	0.452	0.478		5.75	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.