

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q1939 SAS No.: Q1939 SDG No.: Q1939

Instrument ID: MSVOA_N Calibration Date/Time: 05/06/2025 10:35

Lab File ID: VN086504.D Init. Calib. Date(s): 04/15/2025 04/15/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:29 14:29

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Acetone	250.000	244.34	2.263202	20