

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q1939 SAS No.: Q1939 SDG No.: Q1939

Instrument ID: MSVOA_Y Calibration Date/Time: 05/02/2025 10:56

Lab File ID: VY022119.D Init. Calib. Date(s): 04/22/2025 04/22/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:39 16:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Acetone	250.000	252.24	-0.8960022	20