

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1939 SAS No.: Q1939 SDG NO.: Q1939
 Lab File ID: VN086478.D Date Analyzed: 05/05/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:32
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	154484	8.22	297307	9.10	276417	11.87
UPPER LIMIT	308968	8.718	594614	9.6	552834	12.365
LOWER LIMIT	77242	7.718	148654	8.6	138209	11.365
EPA SAMPLE NO.						
GB3B	176014	8.22	330285	9.10	317139	11.87
VN0505MBL01	156968	8.22	298822	9.10	277824	11.87
VN0505MBS01	174392	8.22	323563	9.10	294006	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.