

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1939 SAS No.: Q1939 SDG NO.: Q1939
 Lab File ID: VY022119.D Date Analyzed: 05/02/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:56
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	263357	13.347				
UPPER LIMIT	526714	13.847				
LOWER LIMIT	131679	12.847				
EPA SAMPLE NO.						
GB1	227037	13.35				
GB1B	220532	13.35				
GB2B	245346	13.35				
GB3	245556	13.35				
VY0502SBL01	242655	13.35				
VY0502SBS01	264963	13.35				
VY0502SBSD01	257067	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.