

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN050525

Review By	John Carlone	Review On	5/6/2025 10:01:09 AM
Supervise By	Maresh Dadoda	Supervise On	5/6/2025 12:40:44 PM
SubDirectory	VN050525	HP Acquire Method	HP Processing Method 82N041525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133808
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133809,VP133810

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086477.D	05 May 2025 10:23		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086478.D	05 May 2025 11:32	pH#Lot#V12668	JC\MD	Ok,M
3	VN0505MBL01	VN0505MBL01	VN086479.D	05 May 2025 12:13		JC\MD	Ok
4	VN0505WBL01	VN0505WBL01	VN086480.D	05 May 2025 12:38		JC\MD	Ok
5	VN0505WBS01	VN0505WBS01	VN086481.D	05 May 2025 13:02		JC\MD	Ok,M
6	VN0505WBSD01	VN0505WBSD01	VN086482.D	05 May 2025 13:36		JC\MD	Ok,M
7	IBLK	IBLK	VN086483.D	05 May 2025 14:00		JC\MD	Ok
8	Q1942-01	STORAGE-BLANK-SO	VN086484.D	05 May 2025 14:24	vial A pH<2	JC\MD	Ok
9	Q1942-02	STORAGE-BLANK-WA	VN086485.D	05 May 2025 14:48	vial A pH<2	JC\MD	Ok
10	Q1942-03	STORAGE-BLANK-WA	VN086486.D	05 May 2025 15:12	vial A pH<2	JC\MD	Ok
11	Q1942-04	STORAGE-BLANK-SAI	VN086487.D	05 May 2025 15:36	vial A pH<2	JC\MD	Ok
12	Q1956-07	FB-05022025	VN086488.D	05 May 2025 16:00	vial A pH<2 FB	JC\MD	Ok
13	Q1929-05RE	WC-A1-03-GRE	VN086489.D	05 May 2025 16:25	vial B pH#5.0 Surrogate Fail	JC\MD	Confirms
14	Q1929-09RE	WC-A1-04-GRE	VN086490.D	05 May 2025 16:49	vial B pH#5.0 Surrogate Fail	JC\MD	Confirms
15	VN0505MBS01	VN0505MBS01	VN086491.D	05 May 2025 17:13		JC\MD	Ok,M
16	VN0505MBSD01	VN0505MBSD01	VN086492.D	05 May 2025 17:37		JC\MD	Ok,M
17	Q1914-06ME	SS-5ME	VN086493.D	05 May 2025 18:01	run straight meoh	JC\MD	Not Ok
18	Q1914-07MEDL	SS-6MEDL	VN086494.D	05 May 2025 18:25		JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN050525

Review By	John Carlone	Review On	5/6/2025 10:01:09 AM
Supervise By	Maresh Dadoda	Supervise On	5/6/2025 12:40:44 PM
SubDirectory	VN050525	HP Acquire Method	HP Processing Method 82N041525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133808		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133809,VP133810		

19	Q1914-08DL	SS-7DL	VN086495.D	05 May 2025 18:49		JC\MD	Ok
20	Q1914-09MEDL	SS-8MEDL	VN086496.D	05 May 2025 19:14		JC\MD	Ok
21	Q1914-12	SS-9	VN086497.D	05 May 2025 19:38	run soil	JC\MD	Not Ok
22	IBLK	IBLK	VN086498.D	05 May 2025 20:02		JC\MD	Ok
23	Q1939-03	GB2	VN086499.D	05 May 2025 20:26	Need Straight Run	JC\MD	Not Ok
24	Q1939-06	GB3B	VN086500.D	05 May 2025 20:50	vial C	JC\MD	Ok
25	IBLK	IBLK	VN086501.D	05 May 2025 21:14		JC\MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VN086502.D	05 May 2025 21:38		JC\MD	Ok,M

M : Manual Integration