

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN050925

Review By	John Carlone	Review On	5/12/2025 9:52:42 AM
Supervise By	Mahesh Dadoda	Supervise On	5/12/2025 1:58:31 PM
SubDirectory	VN050925	HP Acquire Method	HP Processing Method 82N041525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133876
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133877,VP133878

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086552.D	09 May 2025 09:49		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086553.D	09 May 2025 11:00	CCAL failed low for comp. #31	JC\MD	Ok,M
3	VN0509MBL01	VN0509MBL01	VN086554.D	09 May 2025 11:34		JC\MD	Ok
4	VN0509WBL01	VN0509WBL01	VN086555.D	09 May 2025 11:58		JC\MD	Ok
5	VN0509MBS01	VN0509MBS01	VN086556.D	09 May 2025 12:22	Recovery fail	JC\MD	Not Ok
6	VN0509MBS02	VN0509MBS02	VN086557.D	09 May 2025 12:57		JC\MD	Ok,M
7	Q1939-03RE	GB2RE	VN086558.D	09 May 2025 13:31	vial C, CCAL Failed Low	JC\MD	Confirms
8	Q1980-04	3848	VN086559.D	09 May 2025 13:55	Need 40X	JC\MD	Dilution
9	Q1980-05	3837	VN086560.D	09 May 2025 14:19	E flag of previous sample	JC\MD	ReRun
10	IBLK	IBLK	VN086561.D	09 May 2025 14:43		JC\MD	Ok
11	IBLK	IBLK	VN086562.D	09 May 2025 15:07		JC\MD	Ok
12	Q1980-04DL	3848DL	VN086563.D	09 May 2025 15:31	Need 100X	JC\MD	Dilution
13	IBLK	IBLK	VN086564.D	09 May 2025 16:08		JC\MD	Ok
14	IBLK	IBLK	VN086565.D	09 May 2025 16:32		JC\MD	Ok
15	Q1980-05	3837	VN086566.D	09 May 2025 16:56	cloth material	JC\MD	Ok
16	Q1980-04DL2	3848DL2	VN086567.D	09 May 2025 17:20	cloth material	JC\MD	Ok
17	VSTDCCC050	VSTDCCC050EC	VN086568.D	09 May 2025 17:44		JC\MD	Ok,M

M : Manual Integration