

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN050625

Review By	John Carlone	Review On	5/7/2025 9:53:45 AM
Supervise By	Maresh Dadoda	Supervise On	5/7/2025 1:29:39 PM
SubDirectory	VN050625	HP Acquire Method	HP Processing Method 82N041525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133826
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133827,VP133828

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086503.D	06 May 2025 09:13		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086504.D	06 May 2025 10:35	CCC Failed Low For Com.#17,31	JC\MD	Ok,M
3	VN0506MBL01	VN0506MBL01	VN086505.D	06 May 2025 11:11		JC\MD	Ok
4	VN0506WBL01	VN0506WBL01	VN086506.D	06 May 2025 11:35		JC\MD	Ok
5	VN0506WBS01	VN0506WBS01	VN086507.D	06 May 2025 11:59		JC\MD	Ok,M
6	VN0506WBSD01	VN0506WBSD01	VN086508.D	06 May 2025 12:34		JC\MD	Ok,M
7	Q1947-04	MH-P	VN086509.D	06 May 2025 12:58	vial A pH#5.0	JC\MD	Ok
8	Q1947-08	MH-O	VN086510.D	06 May 2025 13:22	vial A pH#5.0	JC\MD	Ok
9	Q1951-04	MH-KK	VN086511.D	06 May 2025 13:46	vial A pH#5.0	JC\MD	Ok,M
10	Q1956-05	COMP1	VN086512.D	06 May 2025 14:11	vial A pH#5.0	JC\MD	Ok,M
11	Q1959-02	SOIL-PILE	VN086513.D	06 May 2025 14:35	vial A pH#5.0	JC\MD	Ok,M
12	Q1960-03	72-11933	VN086514.D	06 May 2025 14:59	vial A pH#5.0	JC\MD	Ok,M
13	VN0506MBS01	VN0506MBS01	VN086515.D	06 May 2025 15:23		JC\MD	Ok,M
14	VN0506MBSD01	VN0506MBSD01	VN086516.D	06 May 2025 15:47		JC\MD	Ok,M
15	Q1939-03	GB2	VN086517.D	06 May 2025 16:11	vial C, CCC Failed Low	JC\MD	ReRun
16	Q1914-06ME	SS-5ME	VN086518.D	06 May 2025 16:35		JC\MD	Ok
17	Q1914-07ME	SS-6ME	VN086519.D	06 May 2025 16:59	Need 20X	JC\MD	Dilution

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN050625

Review By	John Carlone	Review On	5/7/2025 9:53:45 AM
Supervise By	Maresh Dadoda	Supervise On	5/7/2025 1:29:39 PM
SubDirectory	VN050625	HP Acquire Method	HP Processing Method 82N041525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133826		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133827,VP133828		

18	Q1914-08	SS-7	VN086520.D	06 May 2025 17:23	Need 20X	JC\MD	Dilution
19	Q1914-09ME	SS-8ME	VN086521.D	06 May 2025 17:48	Need 20X	JC\MD	Dilution
20	Q1966-03	SILICA-GEL	VN086522.D	06 May 2025 18:12	CCC Failed Low,Run@ Lower Dilution	JC\MD	ReRun
21	IBLK	IBLK	VN086523.D	06 May 2025 18:36		JC\MD	Ok
22	IBLK	IBLK	VN086524.D	06 May 2025 19:00		JC\MD	Ok
23	IBLK	IBLK	VN086525.D	06 May 2025 19:24		JC\MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VN086526.D	06 May 2025 19:48		JC\MD	Ok,M

M : Manual Integration