

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1940 SAS No.: Q1940 SDG No.: Q1940
Instrument ID: MSVOA_N Calibration Date(s): 04/15/2025 04/15/2025
Heated Purge: (Y/N) N Calibration Time(s): 11:29 14:29
GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF001 = VN086277.D		RRF005 = VN086278.D		RRF010 = VN086279.D			
		RRF020 = VN086280.D		RRF050 = VN086281.D		RRF100 = VN086282.D			
COMPOUND	RRF001	RRF005	RRF010	RRF020	RRF050	RRF100	RRF	% RSD	
Chloromethane	1.113	0.904	0.783	0.819	0.754	0.795	0.861	15.5	
Vinyl Chloride	0.897	0.825	0.785	0.846	0.791	0.788	0.822	5.4	
Bromomethane		0.383	0.370	0.395	0.356	0.344	0.370	5.5	
Chloroethane	0.650	0.580	0.519	0.543	0.500	0.509	0.550	10.3	
Tert butyl alcohol		0.145	0.135	0.142	0.124	0.115	0.132	9.4	
1,1-Dichloroethene	0.660	0.622	0.553	0.588	0.551	0.557	0.588	7.6	
Acetone	0.423	0.284	0.252	0.283	0.251	0.247	0.290	23.2	
Carbon Disulfide	2.325	1.864	1.602	1.721	1.558	1.503	1.762	17.3	
Methyl tert-butyl Ether	2.478	2.266	2.013	2.199	2.010	2.014	2.163	8.8	
Methylene Chloride	0.782	0.707	0.616	0.670	0.619	0.629	0.670	9.7	
trans-1,2-Dichloroethene	0.703	0.656	0.557	0.620	0.575	0.580	0.615	9.1	
1,1-Dichloroethane	1.315	1.274	1.122	1.206	1.136	1.152	1.201	6.6	
2-Butanone	0.504	0.473	0.420	0.466	0.425	0.420	0.451	7.7	
Carbon Tetrachloride	0.470	0.471	0.409	0.449	0.423	0.426	0.441	5.9	
cis-1,2-Dichloroethene	0.882	0.810	0.694	0.764	0.714	0.720	0.764	9.3	
Chloroform	1.318	1.233	1.122	1.196	1.105	1.104	1.180	7.3	
1,1,1-Trichloroethane	1.095	1.083	0.953	1.026	0.949	0.949	1.009	6.8	
Benzene	1.648	1.545	1.389	1.508	1.409	1.411	1.485	6.8	
1,2-Dichloroethane	0.526	0.496	0.448	0.482	0.444	0.446	0.474	7.1	
Trichloroethene	0.390	0.357	0.334	0.352	0.338	0.341	0.352	5.8	
1,2-Dichloropropane	0.378	0.383	0.355	0.368	0.349	0.347	0.364	4.2	
Bromodichloromethane	0.531	0.530	0.477	0.510	0.475	0.477	0.500	5.4	
4-Methyl-2-Pentanone	0.522	0.522	0.479	0.521	0.484	0.471	0.500	4.9	
Toluene	1.003	0.963	0.865	0.941	0.892	0.895	0.927	5.6	
t-1,3-Dichloropropene	0.576	0.574	0.528	0.572	0.547	0.554	0.558	3.4	
cis-1,3-Dichloropropene	0.679	0.636	0.580	0.621	0.577	0.587	0.613	6.5	
1,1,2-Trichloroethane	0.367	0.357	0.315	0.332	0.318	0.313	0.333	6.9	
2-Hexanone	0.394	0.387	0.355	0.383	0.355	0.349	0.371	5.3	
Dibromochloromethane	0.377	0.381	0.354	0.375	0.353	0.356	0.366	3.5	
Tetrachloroethene	0.419	0.399	0.358	0.390	0.371	0.370	0.385	5.8	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1940 SAS No.: Q1940 SDG No.: Q1940
 Instrument ID: MSVOA_N Calibration Date(s): 04/15/2025 04/15/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:29 14:29
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF001 = VN086277.D		RRF005 = VN086278.D		RRF010 = VN086279.D			
		RRF020 = VN086280.D		RRF050 = VN086281.D		RRF100 = VN086282.D			
COMPOUND	RRF001	RRF005	RRF010	RRF020	RRF050	RRF100	RRF	% RSD	
Chlorobenzene	1.235	1.175	1.058	1.117	1.050	1.059	1.116	6.8	
Ethyl Benzene	2.210	2.112	1.873	2.028	1.918	1.941	2.014	6.4	
m/p-Xylenes	0.809	0.795	0.705	0.757	0.727	0.734	0.754	5.4	
o-Xylene	0.791	0.792	0.709	0.753	0.713	0.718	0.746	5.2	
Styrene	1.306	1.288	1.156	1.245	1.218	1.246	1.243	4.3	
Bromoform	0.306	0.294	0.258	0.279	0.261	0.264	0.277	7	
1,1,2,2-Tetrachloroethane	1.347	1.294	1.181	1.212	1.035	0.985	1.176	12.1	
1,2-Dichloroethane-d4		0.746	0.745	0.707	0.719	0.708	0.725	2.6	
Dibromofluoromethane		0.267	0.255	0.230	0.215	0.194	0.232	12.8	
Toluene-d8		1.247	1.281	1.185	1.251	1.237	1.240	2.8	
4-Bromofluorobenzene		0.459	0.456	0.421	0.459	0.467	0.452	4	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.