

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01  
 Lab Code: CHEM Case No.: Q1940 SAS No.: Q1940 SDG NO.: Q1940  
 Lab File ID: VN086451.D Date Analyzed: 05/02/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 09:21  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT # | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|------|---------------|--------|
| 12 HOUR STD    | 187608        | 8.22  | 337415        | 9.10 | 308780        | 11.87  |
| UPPER LIMIT    | 375216        | 8.724 | 674830        | 9.6  | 617560        | 12.365 |
| LOWER LIMIT    | 93804         | 7.724 | 168708        | 8.6  | 154390        | 11.365 |
| EPA SAMPLE NO. |               |       |               |      |               |        |
| MW-1           | 131476        | 8.22  | 259020        | 9.10 | 245875        | 11.86  |
| MW-2           | 151630        | 8.22  | 288383        | 9.10 | 274301        | 11.87  |
| MW-3           | 141230        | 8.22  | 272575        | 9.10 | 263663        | 11.87  |
| VN0502WBL01    | 154685        | 8.22  | 298860        | 9.10 | 280972        | 11.87  |
| VN0502WBS01    | 205101        | 8.22  | 372276        | 9.10 | 334682        | 11.86  |
| VN0502WBSD01   | 145586        | 8.22  | 268054        | 9.10 | 242177        | 11.86  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.