

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1940 SAS No.: Q1940 SDG NO.: Q1940
 Lab File ID: VN086451.D Date Analyzed: 05/02/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:21
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	150946	13.788				
UPPER LIMIT	301892	14.288				
LOWER LIMIT	75473	13.288				
EPA SAMPLE NO.						
MW-1	108714	13.79				
MW-2	114473	13.79				
MW-3	115290	13.79				
VN0502WBL01	119397	13.79				
VN0502WBS01	155592	13.79				
VN0502WBSD01	108773	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.