

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1941

Project: Pre Treatment Plant 2025

RunNo.: LB135631

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.488	0.50	98	90-110	05/01/2025
Sample ID: CCV1 Orthophosphate as P	mg/L	0.501	0.5	100	90-110	05/01/2025
Sample ID: CCV2 Orthophosphate as P	mg/L	0.493	0.5	99	90-110	05/01/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1941

Project: Pre Treatment Plant 2025

RunNo.: LB135665

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.95	1	95	90-110	05/05/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.96	1	96	90-110	05/05/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.98	1	98	90-110	05/05/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.99	1	99	90-110	05/05/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.98	1	98	90-110	05/05/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1941

Project: Pre Treatment Plant 2025

RunNo.: LB135695

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.509	0.50	102	90-110	05/07/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.520	0.50	104	90-110	05/07/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.519	0.50	104	90-110	05/07/2025