

Preparation Blank Summary

Client: Holland Manufacturing Co.

SDG No.: Q1941

Project: Pre Treatment Plant 2025

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB135631BL Orthophosphate as P	mg/L	0.006	0.0250	J	0.004	0.05	05/01/2025
Sample ID: LB135645BL BOD5	mg/L	< 0.2000	0.2000	U	0.20	2.0	05/02/2025
Sample ID: LB135657BL Oil and Grease	mg/L	< 2.5000	2.5000	U	0.29	5.0	05/05/2025
Sample ID: LB135658BL TSS	mg/L	1	2.0000	J	1	4	05/05/2025
Sample ID: PB167831BL Phosphorus, Total	mg/L	< 0.0250	0.0250	U	0.005	0.05	05/07/2025
Sample ID: PB167844BL Ammonia as N	mg/L	0.038	0.0500	J	0.03	0.1	05/05/2025