

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q1941
Project: Pre Treatment Plant 2025	Sample ID: Q1924-02
Client ID: COMPDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	411		411		1	0.12		05/02/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q1941
Project: Pre Treatment Plant 2025	Sample ID: Q1930-01
Client ID: WATER-TREATMENT-DISCHARGEDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	2.30	OR	2.20	OR	1	4		05/05/2025
Ammonia as N	mg/L	+/-20	2.40	D	2.50	D	2	4		05/05/2025

Duplicate Sample Summary

Client:	Holland Manufacturing Co.	SDG No.:	Q1941
Project:	Pre Treatment Plant 2025	Sample ID:	Q1930-01
Client ID:	WATER-TREATMENT-DISCHARGEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	2.90	OR	3.20	OR	1	10		05/05/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q1941
Project: Pre Treatment Plant 2025	Sample ID: Q1941-01
Client ID: EFFLUENTDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Orthophosphate as P	mg/L	+/-20	0.047	J	0.046	J	1	2.15		05/01/2025
Phosphorus, Total	mg/L	+/-20	0.071		0.072		1	1.4		05/07/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q1941
Project: Pre Treatment Plant 2025	Sample ID: Q1941-01
Client ID: EFFLUENTMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Orthophosphate as P	mg/L	+/-20	0.51		0.51		1	1.37		05/01/2025
Phosphorus, Total	mg/L	+/-20	0.51		0.51		1	1.37		05/07/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q1941
Project: Pre Treatment Plant 2025	Sample ID: Q1941-01
Client ID: EFFLUENTMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	27.9		28.0		1	0.36		05/05/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q1941
Project: Pre Treatment Plant 2025	Sample ID: Q1949-04
Client ID: 002-35TH-AVE(MAY)DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	102		103		1	0.68		05/05/2025