

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1941

Project: Pre Treatment Plant 2025

Run No.: LB135631

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135631BS							
Orthophosphate as P	mg/L	0.5	0.51		102	1	90-110	05/01/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1941

Project: Pre Treatment Plant 2025

Run No.: LB135645

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135645BS							
BOD5	mg/L	198	185		93	1	84.6-115.4	05/02/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1941

Project: Pre Treatment Plant 2025

Run No.: LB135657

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135657BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	05/05/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1941

Project: Pre Treatment Plant 2025

Run No.: LB135658

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135658BS							
TSS	mg/L	550	533		97	1	90-110	05/05/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1941

Project: Pre Treatment Plant 2025

Run No.: LB135695

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167831BS							
Phosphorus, Total	mg/L	0.50	0.50		101	1	90-110	05/07/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1941

Project: Pre Treatment Plant 2025

Run No.: LB135665

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167844BS							
Ammonia as N	mg/L	1	0.96		96	1	90-110	05/05/2025