

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1942 SAS No.: Q1942 SDG NO.: Q1942

Continuing Calib Date: 05/06/2025 Initial Calibration Date(s): 04/09/2025 04/09/2025

Continuing Calib Time: 10:20 Initial Calibration Time(s): 09:41 19:09

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.48	4.48	4.41	4.55	0.00
Aroclor-1016-2	(2)	4.50	4.50	4.43	4.57	0.00
Aroclor-1016-3	(3)	4.56	4.56	4.49	4.63	0.00
Aroclor-1016-4	(4)	4.65	4.65	4.58	4.72	0.00
Aroclor-1016-5	(5)	4.94	4.94	4.87	5.01	0.01
Aroclor-1260-1	(1)	6.03	6.03	5.96	6.10	0.00
Aroclor-1260-2	(2)	6.28	6.28	6.21	6.35	0.00
Aroclor-1260-3	(3)	6.63	6.63	6.56	6.70	0.00
Aroclor-1260-4	(4)	6.85	6.85	6.78	6.92	0.01
Aroclor-1260-5	(5)	7.15	7.15	7.08	7.22	0.00
TCX		3.41	3.40	3.35	3.45	0.00
DCB		8.50	8.50	8.40	8.60	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1942 SAS No.: Q1942 SDG NO.: Q1942

Continuing Calib Date: 05/06/2025 Initial Calibration Date(s): 04/09/2025 04/09/2025

Continuing Calib Time: 10:20 Initial Calibration Time(s): 09:41 19:09

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	3.78	3.79	3.72	3.86	0.01
Aroclor-1016-2	(2)	3.80	3.80	3.73	3.87	0.00
Aroclor-1016-3	(3)	3.96	3.96	3.89	4.03	0.00
Aroclor-1016-4	(4)	4.01	4.01	3.94	4.08	0.00
Aroclor-1016-5	(5)	4.20	4.21	4.14	4.28	0.01
Aroclor-1260-1	(1)	5.20	5.20	5.13	5.27	0.00
Aroclor-1260-2	(2)	5.39	5.39	5.32	5.46	0.00
Aroclor-1260-3	(3)	5.53	5.53	5.46	5.60	0.00
Aroclor-1260-4	(4)	5.99	5.99	5.92	6.06	0.00
Aroclor-1260-5	(5)	6.23	6.23	6.16	6.30	0.00
TCX		2.78	2.78	2.73	2.83	0.00
DCB		7.54	7.55	7.45	7.65	0.01