

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1942

Client: Chemtech Consulting Group

Analytical Method: SW8260-Low

Datafile : VN086482.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0505WBSD01	Dichlorodifluoromethane	20	19.4	ug/L	97	1		69	116	20
	Chloromethane	20	18.3	ug/L	92	3		65	116	20
	Vinyl chloride	20	18.8	ug/L	94	1		65	117	20
	Ethyl Acetate	20	19.9	ug/L	100	7		75	115	20
	Bromomethane	20	24.9	ug/L	125	3		58	125	20
	Chloroethane	20	18.9	ug/L	95	1		56	128	20
	Trichlorofluoromethane	20	19.8	ug/L	99	2		73	115	20
	1,1,2-Trichlorotrifluoroethane	20	20.2	ug/L	101	1		80	112	20
	Tert butyl alcohol	100	98.1	ug/L	98	12		73	124	20
	1,1-Dichloroethene	20	18.9	ug/L	95	1		74	110	20
	Acrolein	100	90.4	ug/L	90	3		51	143	20
	Acrylonitrile	100	99.8	ug/L	100	8		73	113	20
	Acetone	100	100	ug/L	100	7		60	125	20
	Carbon disulfide	20	17.7	ug/L	89	0		64	112	20
	Methyl tert-butyl Ether	20	20.4	ug/L	102	8		78	114	20
	Methyl Acetate	20	18.1	ug/L	91	12		67	125	20
	Methylene Chloride	20	20.8	ug/L	104	4		72	114	20
	trans-1,2-Dichloroethene	20	20.0	ug/L	100	2		75	108	20
	Vinyl Acetate	100	98.4	ug/L	98	2		76	115	20
	1,1-Dichloroethane	20	19.7	ug/L	99	3		78	112	20
	Cyclohexane	20	17.7	ug/L	89	1		75	110	20
	2-Butanone	100	96.5	ug/L	97	10		65	122	20
	Carbon Tetrachloride	20	21.0	ug/L	105	2		77	113	20
	2,2-Dichloropropane	20	19.0	ug/L	95	3		75	116	20
	cis-1,2-Dichloroethene	20	20.3	ug/L	102	2		77	110	20
	Bromochloromethane	20	19.0	ug/L	95	3		70	124	20
	Chloroform	20	20.8	ug/L	104	3		79	113	20
	1,1,1-Trichloroethane	20	20.2	ug/L	101	3		80	108	20
	Methylcyclohexane	20	18.7	ug/L	94	2		72	115	20
	1,1-Dichloropropene	20	19.5	ug/L	98	2		79	110	20
	Benzene	20	19.9	ug/L	100	2		82	109	20
	1,2-Dichloroethane	20	21.7	ug/L	109	6		80	115	20
	Trichloroethene	20	21.0	ug/L	105	3		77	113	20
	1,2-Dichloropropane	20	20.1	ug/L	101	5		83	111	20
	Dibromomethane	20	22.9	ug/L	115	9	*	82	110	20
	Bromodichloromethane	20	21.2	ug/L	106	3		83	110	20
	4-Methyl-2-Pentanone	100	99.6	ug/L	100	7		74	118	20
	Toluene	20	20.6	ug/L	103	1		82	110	20
	t-1,3-Dichloropropene	20	20.5	ug/L	103	4		79	110	20
	cis-1,3-Dichloropropene	20	19.7	ug/L	99	4		82	110	20
	1,1,2-Trichloroethane	20	21.9	ug/L	110	6		83	112	20
	1,3-Dichloropropane	20	21.7	ug/L	109	7		83	111	20
	2-Chloroethyl vinyl ether	100	83.1	ug/L	83	1		75	124	20
	2-Hexanone	100	99.1	ug/L	99	5		73	117	20
	Dibromochloromethane	20	22.0	ug/L	110	4		82	110	20
	1,2-Dibromoethane	20	22.1	ug/L	111	7	*	81	110	20
	Tetrachloroethene	20	21.1	ug/L	106	6		67	123	20
	Chlorobenzene	20	20.9	ug/L	104	1		82	109	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1942

Client: Chemtech Consulting Group

Analytical Method: SW8260-Low

Datafile : VN086482.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0505WBSD01	1,1,1,2-Tetrachloroethane	20	21.2	ug/L	106	4		84	111	20
	Hexachloroethane	20	19.1	ug/L	96	6		80	113	20
	Ethyl Benzene	20	19.9	ug/L	100	1		83	109	20
	m/p-Xylenes	40	40.7	ug/L	102	1		82	110	20
	o-Xylene	20	20.1	ug/L	101	1		83	109	20
	Styrene	20	20.7	ug/L	104	4		80	111	20
	Bromoform	20	21.4	ug/L	107	6		79	109	20
	Isopropylbenzene	20	19.7	ug/L	99	2		83	112	20
	1,1,2,2-Tetrachloroethane	20	21.3	ug/L	106	0		76	118	20
	1,2,3-Trichloropropane	20	20.1	ug/L	101	1		75	112	20
	Bromobenzene	20	21.6	ug/L	108	3		77	116	20
	N-propylbenzene	20	19.6	ug/L	98	4		83	112	20
	2-Chlorotoluene	20	19.7	ug/L	99	3		83	110	20
	1,3,5-Trimethylbenzene	20	19.9	ug/L	100	2		85	112	20
	4-Chlorotoluene	20	20.0	ug/L	100	0		82	110	20
	tert-Butylbenzene	20	19.3	ug/L	97	5		83	112	20
	1,2,4-Trimethylbenzene	20	20.1	ug/L	101	0		85	111	20
	Sec-butylbenzene	20	19.8	ug/L	99	4		81	114	20
	p-Isopropyltoluene	20	19.9	ug/L	100	5		78	116	20
	1,3-Dichlorobenzene	20	21.3	ug/L	106	3		82	108	20
	1,4-Dichlorobenzene	20	20.9	ug/L	104	2		82	107	20
	n-Butylbenzene	20	19.1	ug/L	96	3		75	115	20
	1,2-Dichlorobenzene	20	20.7	ug/L	104	2		82	109	20
	1,2-Dibromo-3-Chloropropane	20	18.9	ug/L	95	2		68	112	20
	1,2,4-Trichlorobenzene	20	19.5	ug/L	98	4		75	113	20
	Hexachlorobutadiene	20	19.9	ug/L	100	4		81	121	20
	Naphthalene	20	18.9	ug/L	95	10		78	119	20
	1,2,3-Trichlorobenzene	20	19.6	ug/L	98	2		76	114	20