

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

SW-846

**SDG No.:** Q1943

**Client:** Tetra Tech NUS, Inc.

**Analytical Method:** 8270-Modified **DataFile:** BN036967.D

| Lab Sample ID | Parameter   | Spike | Result | Unit | Rec | RPD | Qual | RPD  | Low | Limits | RPD |
|---------------|-------------|-------|--------|------|-----|-----|------|------|-----|--------|-----|
|               |             |       |        |      |     |     |      | Qual |     | High   |     |
| PB167888BS    | 1,4-Dioxane | 0.4   | 0.29   | ug/L | 73  |     |      |      | 70  | 130    |     |