

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1944

Project: Tank Farm - Acid Analysis

RunNo.: LB135647

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.01	7	100	90-110	05/02/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	05/02/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	05/02/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1944

Project: Tank Farm - Acid Analysis

RunNo.: LB135648

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	05/02/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	05/02/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	05/02/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1944

Project: Tank Farm - Acid Analysis

RunNo.: LB135665

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.95	1	95	90-110	05/05/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.96	1	96	90-110	05/05/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.98	1	98	90-110	05/05/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.99	1	99	90-110	05/05/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.98	1	98	90-110	05/05/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1944

Project: Tank Farm - Acid Analysis

RunNo.: LB135671

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TKN	mg/L	4.8	5	96	90-110	05/05/2025
Sample ID: CCV1 TKN	mg/L	4.9	5	98	90-110	05/05/2025
Sample ID: CCV2 TKN	mg/L	5	5	100	90-110	05/05/2025
Sample ID: CCV3 TKN	mg/L	5	5	100	90-110	05/05/2025
Sample ID: CCV4 TKN	mg/L	4.8	5	96	90-110	05/05/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1944

Project: Tank Farm - Acid Analysis

RunNo.: LB135673

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
Phenolics		mg/L	0.94	1	94	90-110	05/05/2025
Sample ID:	CCV1						
Phenolics		mg/L	1.1	1	110	90-110	05/05/2025
Sample ID:	CCV2						
Phenolics		mg/L	1.2	1	120	90-110	05/05/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1944

Project: Tank Farm - Acid Analysis

RunNo.: LB135681

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
Phenolics		mg/L	0.93	1	93	90-110	05/06/2025
Sample ID:	CCV1						
Phenolics		mg/L	0.93	1	93	90-110	05/06/2025
Sample ID:	CCV2						
Phenolics		mg/L	0.95	1	95	90-110	05/06/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1944

Project: Tank Farm - Acid Analysis

RunNo.: LB135681

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------