

Duplicate Sample Summary

Client: METEM A GE POWER Business	SDG No.: Q1944
Project: Tank Farm - Acid Analysis	Sample ID: Q1930-01
Client ID: WATER-TREATMENT-DISCHARGEDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	2.30	OR	2.20	OR	1	4		05/05/2025
Ammonia as N	mg/L	+/-20	2.40	D	2.50	D	2	4		05/05/2025

Duplicate Sample Summary

Client: METEM A GE POWER Business	SDG No.: Q1944
Project: Tank Farm - Acid Analysis	Sample ID: Q1930-01
Client ID: WATER-TREATMENT-DISCHARGEMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	2.90	OR	3.20	OR	1	10		05/05/2025

Duplicate Sample Summary

Client: METEM A GE POWER Business	SDG No.: Q1944
Project: Tank Farm - Acid Analysis	Sample ID: Q1944-02
Client ID: CITY-WATERDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
pH	pH	+/-20	7.65		7.66		1	0.13		05/02/2025
Alkalinity	mg/L	+/-20	178		179		1	1		05/05/2025
Acidity	mg/L	+/-10	4.00	U	4.00	U	1	0		05/05/2025
TKN	mg/L	+/-20	0.11	J	0.11	J	1	0		05/05/2025

Duplicate Sample Summary

Client:	METEM A GE POWER Business	SDG No.:	Q1944
Project:	Tank Farm - Acid Analysis	Sample ID:	Q1944-02
Client ID:	CITY-WATERMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TKN	mg/L	+/-20	4.50		4.70		1	4		05/05/2025

Duplicate Sample Summary

Client: METEM A GE POWER Business	SDG No.: Q1944
Project: Tank Farm - Acid Analysis	Sample ID: Q1944-03
Client ID: CHILLER-WATERDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Alkalinity	mg/L	+/-20	5650		5680		1	1		05/05/2025
Phenolics	mg/L	+/-20	0.015	U	0.015	U	1	0		05/05/2025
Phenolics	mg/L	+/-20	0.015	U	0.015	U	1	0		05/06/2025

Duplicate Sample Summary

Client: METEM A GE POWER Business	SDG No.: Q1944
Project: Tank Farm - Acid Analysis	Sample ID: Q1944-03
Client ID: CHILLER-WATERMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Phenolics	mg/L	+/-20	1.40		1.30		1	7		05/05/2025
Phenolics	mg/L	+/-20	1.10		1.10		1	0		05/06/2025