

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135662

Review By	mohan	Review On	5/6/2025 10:06:50 AM
Supervise By	jaswal	Supervise On	5/6/2025 4:50:43 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85530,MP85531,MP85532,MP85533,MP85534,MP85535		
ICV Standard	MP85536		
CCV Standard	MP85538		
ICSA Standard			
CRI Standard	MP85540		
LCS Standard			
Chk Standard	MP85537,MP85539,MP85541,MP85543		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	05/05/25 10:59		mohan	OK
2	S0.2	S0.2	CAL2	05/05/25 11:01		mohan	OK
3	S2.5	S2.5	CAL3	05/05/25 11:03		mohan	OK
4	S5	S5	CAL4	05/05/25 11:06		mohan	OK
5	S7.5	S7.5	CAL5	05/05/25 11:08		mohan	OK
6	S10	S10	CAL6	05/05/25 11:13		mohan	OK
7	ICV114	ICV114	ICV	05/05/25 11:16		mohan	OK
8	ICB114	ICB114	ICB	05/05/25 11:18		mohan	OK
9	CCV71	CCV71	CCV	05/05/25 11:25		mohan	OK
10	CCB71	CCB71	CCB	05/05/25 11:30		mohan	OK
11	CRA	CRA	CRDL	05/05/25 11:33		mohan	OK
12	HighStd	HighStd	HIGH STD	05/05/25 11:35		mohan	OK
13	ChkStd	ChkStd	SAM	05/05/25 11:37		mohan	OK
14	PB167839BL	PB167839BL	MB	05/05/25 11:42		mohan	OK
15	PB167839BS	PB167839BS	LCS	05/05/25 11:44		mohan	OK
16	Q1916-04	WC-12	SAM	05/05/25 11:47		mohan	OK
17	Q1917-04	MH-JJ	SAM	05/05/25 11:49		mohan	OK
18	Q1922-04	MH-R	SAM	05/05/25 11:51		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135662

Review By	mohan	Review On	5/6/2025 10:06:50 AM
Supervise By	jaswal	Supervise On	5/6/2025 4:50:43 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85530,MP85531,MP85532,MP85533,MP85534,MP85535		
ICV Standard	MP85536		
CCV Standard	MP85538		
ICSA Standard			
CRI Standard	MP85540		
LCS Standard			
Chk Standard	MP85537,MP85539,MP85541,MP85543		

19	Q1922-08	MH-S	SAM	05/05/25 11:53		mohan	OK
20	Q1925-01	AUD-25-0068	SAM	05/05/25 11:56		mohan	OK
21	CCV72	CCV72	CCV	05/05/25 11:58		mohan	OK
22	CCB72	CCB72	CCB	05/05/25 12:03		mohan	OK
23	Q1925-02	AUD-25-0069	SAM	05/05/25 12:05		mohan	OK
24	Q1925-03	AUD-25-0070	SAM	05/05/25 12:07		mohan	OK
25	Q1928-04	MH-Q	SAM	05/05/25 12:10		mohan	OK
26	Q1929-03	WC-A4-02-C	SAM	05/05/25 12:12		mohan	OK
27	Q1929-07	WC-A1-03-C	SAM	05/05/25 12:14		mohan	OK
28	Q1929-11	WC-A1-04-C	SAM	05/05/25 12:16		mohan	OK
29	Q1932-02	COMP-4	SAM	05/05/25 12:19		mohan	OK
30	Q1932-04	COMP-5	SAM	05/05/25 12:21		mohan	OK
31	Q1932-06	COMP-6	SAM	05/05/25 12:23		mohan	OK
32	Q1932-08	COMP-7	SAM	05/05/25 12:26		mohan	OK
33	CCV73	CCV73	CCV	05/05/25 12:28		mohan	OK
34	CCB73	CCB73	CCB	05/05/25 12:38		mohan	OK
35	Q1935-04	MH-NN	SAM	05/05/25 12:40		mohan	OK
36	Q1935-08	MH-MM	SAM	05/05/25 12:43		mohan	OK
37	Q1935-08DUP	MH-MMDUP	DUP	05/05/25 12:45		mohan	OK
38	Q1935-08MS	MH-MMMS	MS	05/05/25 12:47		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135662

Review By	mohan	Review On	5/6/2025 10:06:50 AM
Supervise By	jaswal	Supervise On	5/6/2025 4:50:43 PM

STD. NAME	STD REF.#
ICAL Standard	MP85530,MP85531,MP85532,MP85533,MP85534,MP85535
ICV Standard	MP85536
CCV Standard	MP85538
ICSA Standard	
CRI Standard	MP85540
LCS Standard	
Chk Standard	MP85537,MP85539,MP85541,MP85543

39	Q1935-08MSD	MH-MMMSD	MSD	05/05/25 12:50		mohan	OK
40	PB167866BL	PB167866BL	MB	05/05/25 12:52		mohan	OK
41	PB167866BS	PB167866BS	LCS	05/05/25 12:54		mohan	OK
42	Q1944-02	CITY-WATER	SAM	05/05/25 12:56		mohan	OK
43	Q1944-03	CHILLER-WATER	SAM	05/05/25 12:59		mohan	OK
44	Q1945-01	GB1W	SAM	05/05/25 13:01		mohan	OK
45	CCV74	CCV74	CCV	05/05/25 13:03		mohan	OK
46	CCB74	CCB74	CCB	05/05/25 13:06		mohan	OK
47	Q1945-01DUP	GB1WDUP	DUP	05/05/25 13:17		mohan	OK
48	Q1945-01MS	GB1WMS	MS	05/05/25 13:19		mohan	OK
49	Q1945-01MSD	GB1WMSD	MSD	05/05/25 13:21		mohan	OK
50	Q1956-07	FB-05022025	SAM	05/05/25 13:24		mohan	OK
51	PB167815TB	PB167815TB	MB	05/05/25 13:26		mohan	OK
52	Q1935-08L	MH-MML	SD	05/05/25 13:28		mohan	OK
53	Q1935-08A	MH-MMA	PS	05/05/25 13:31		mohan	OK
54	Q1945-01L	GB1WL	SD	05/05/25 13:36		mohan	OK
55	Q1945-01A	GB1WA	PS	05/05/25 13:39		mohan	OK
56	CCV75	CCV75	CCV	05/05/25 13:41		mohan	OK
57	CCB75	CCB75	CCB	05/05/25 13:43		mohan	OK
58	Q1913-01	WC-12-S-202504	SAM	05/05/25 13:45		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135662

Review By	mohan	Review On	5/6/2025 10:06:50 AM
Supervise By	jaswal	Supervise On	5/6/2025 4:50:43 PM

STD. NAME	STD REF.#
ICAL Standard	MP85530,MP85531,MP85532,MP85533,MP85534,MP85535
ICV Standard	MP85536
CCV Standard	MP85538
ICSA Standard	
CRI Standard	MP85540
LCS Standard	
Chk Standard	MP85537,MP85539,MP85541,MP85543

59	Q1913-03	WC-13-S-202504	SAM	05/05/25 13:48		mohan	OK
60	PB167774TB	PB167774TB	MB	05/05/25 13:50		mohan	OK
61	CCV76	CCV76	CCV	05/05/25 13:52		mohan	OK
62	CCB76	CCB76	CCB	05/05/25 13:55		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135670

Review By	mohan	Review On	5/6/2025 10:11:57 AM
Supervise By	JANVI	Supervise On	5/6/2025 2:58:22 PM

STD. NAME	STD REF.#
ICAL Standard	MP85530,MP85531,MP85532,MP85533,MP85534,MP85535
ICV Standard	MP85536
CCV Standard	MP85538
ICSA Standard	
CRI Standard	MP85540
LCS Standard	
Chk Standard	MP85537,MP85539,MP85541,MP85543

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	05/05/25 14:51		mohan	OK
2	S0.2	S0.2	CAL2	05/05/25 14:54		mohan	OK
3	S2.5	S2.5	CAL3	05/05/25 14:56		mohan	OK
4	S5	S5	CAL4	05/05/25 14:58		mohan	OK
5	S7.5	S7.5	CAL5	05/05/25 15:01		mohan	OK
6	S10	S10	CAL6	05/05/25 15:03		mohan	OK
7	ICV116	ICV116	ICV	05/05/25 15:08		mohan	OK
8	ICB116	ICB116	ICB	05/05/25 15:11		mohan	OK
9	CCV79	CCV79	CCV	05/05/25 15:13		mohan	OK
10	CCB79	CCB79	CCB	05/05/25 15:15		mohan	OK
11	CRA	CRA	CRDL	05/05/25 15:18		mohan	OK
12	HighStd	HighStd	HIGH STD	05/05/25 15:20		mohan	OK
13	ChkStd	ChkStd	SAM	05/05/25 15:22		mohan	OK
14	PB167870BL	PB167870BL	MB	05/05/25 15:24		mohan	OK
15	PB167870BS	PB167870BS	LCS	05/05/25 15:30		mohan	OK
16	Q1944-01	MELTED-RUBBER-S	SAM	05/05/25 15:32		mohan	OK
17	Q1956-01	SB1-3-4	SAM	05/05/25 15:34		mohan	OK
18	Q1956-02	SB2-4-5	SAM	05/05/25 15:37		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QCBatch ID # LB135670

Review By	mohan	Review On	5/6/2025 10:11:57 AM
Supervise By	JANVI	Supervise On	5/6/2025 2:58:22 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85530,MP85531,MP85532,MP85533,MP85534,MP85535		
ICV Standard	MP85536		
CCV Standard	MP85538		
ICSA Standard			
CRI Standard	MP85540		
LCS Standard			
Chk Standard	MP85537,MP85539,MP85541,MP85543		

19	Q1956-02DUP	SB2-4-5DUP	DUP	05/05/25 15:39		mohan	OK
20	Q1956-03	SB2-4-5MS	MS	05/05/25 15:44		mohan	OK
21	CCV80	CCV80	CCV	05/05/25 15:46		mohan	OK
22	CCB80	CCB80	CCB	05/05/25 15:48		mohan	OK
23	Q1956-04	SB2-4-5MSD	MSD	05/05/25 15:51		mohan	OK
24	Q1956-06	SB91-3-4	SAM	05/05/25 15:53		mohan	OK
25	Q1956-02L	SB2-4-5L	SD	05/05/25 15:55		mohan	OK
26	Q1956-02A	SB2-4-5A	PS	05/05/25 15:58		mohan	OK
27	CCV81	CCV81	CCV	05/05/25 16:00		mohan	OK
28	CCB81	CCB81	CCB	05/05/25 16:02		mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135674

Review By	JANVI	Review On	5/6/2025 4:57:41 PM
Supervise By	jaswal	Supervise On	5/7/2025 1:30:24 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard			
LCS Standard			
Chk Standard	MP85557,MP85558		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	05/05/25 13:02		Kareem	OK
2	S1	S1	CAL2	05/05/25 13:07		Kareem	OK
3	S2	S2	CAL3	05/05/25 13:11		Kareem	OK
4	S3	S3	CAL4	05/05/25 13:15		Kareem	OK
5	S4	S4	CAL5	05/05/25 13:19		Kareem	OK
6	S5	S5	CAL6	05/05/25 13:24		Kareem	OK
7	ICV01	ICV01	ICV	05/05/25 14:18		Kareem	OK
8	LLICV01	LLICV01	LLICV	05/05/25 14:36		Kareem	OK
9	ICB01	ICB01	ICB	05/05/25 14:41		Kareem	OK
10	CRI01	CRI01	CRDL	05/05/25 14:47		Kareem	OK
11	ICSA01	ICSA01	ICSA	05/05/25 14:52		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	05/05/25 15:07		Kareem	OK
13	ICSADL	ICSADL	ICSA	05/05/25 15:11		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	05/05/25 15:16		Kareem	OK
15	CCV01	CCV01	CCV	05/05/25 15:20		Kareem	OK
16	CCB01	CCB01	CCB	05/05/25 15:26		Kareem	OK
17	Q1916-01	WC-12	SAM	05/05/25 15:30		Kareem	OK
18	Q1917-01	MH-JJ	SAM	05/05/25 15:35		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135674

Review By	JANVI	Review On	5/6/2025 4:57:41 PM
Supervise By	jaswal	Supervise On	5/7/2025 1:30:24 AM

STD. NAME	STD REF.#
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546
ICV Standard	MP85553
CCV Standard	MP85556
ICSA Standard	MP85554,MP85555
CRI Standard	
LCS Standard	
Chk Standard	MP85557,MP85558

19	Q1923-01	TR-05-04302025	SAM	05/05/25 15:39		Kareem	OK
20	Q1923-01DUP	TR-05-04302025DUP	DUP	05/05/25 15:43		Kareem	OK
21	Q1923-01L	TR-05-04302025L	SD	05/05/25 15:47		Kareem	OK
22	Q1923-01MS	TR-05-04302025MS	MS	05/05/25 15:51		Kareem	OK
23	Q1923-01MSD	TR-05-04302025MSD	MSD	05/05/25 15:55		Kareem	OK
24	Q1923-01A	TR-05-04302025A	PS	05/05/25 15:59		Kareem	OK
25	Q1944-01	MELTED-RUBBER-S	SAM	05/05/25 16:03		Kareem	OK
26	CCV02	CCV02	CCV	05/05/25 16:07		Kareem	OK
27	CCB02	CCB02	CCB	05/05/25 16:12		Kareem	OK
28	Q1865-08	SOM-25-00051	SAM	05/05/25 16:16		Kareem	OK
29	Q1916-04	WC-12	SAM	05/05/25 16:20		Kareem	OK
30	Q1906-04	WC-4	SAM	05/05/25 16:25		Kareem	OK
31	Q1944-02	CITY-WATER	SAM	05/05/25 16:29		Kareem	OK
32	Q1944-03DL	CHILLER-WATERDL	SAM	05/05/25 16:33	Straight 5x for all elements	Kareem	OK
33	Q1913-01	WC-12-S-202504	SAM	05/05/25 16:38		Kareem	OK
34	Q1913-03	WC-13-S-202504	SAM	05/05/25 16:42	MS/MSD fail for many parameters	Kareem	Not Ok
35	Q1913-03DUP	WC-13-S-202504DUP	DUP	05/05/25 16:47	MS/MSD fail for many parameters	Kareem	Not Ok
36	Q1913-03L	WC-13-S-202504L	SD	05/05/25 16:51	MS/MSD fail for many parameters	Kareem	Not Ok
37	CCV03	CCV03	CCV	05/05/25 16:56		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135674

Review By	JANVI	Review On	5/6/2025 4:57:41 PM
Supervise By	jaswal	Supervise On	5/7/2025 1:30:24 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard			
LCS Standard			
Chk Standard	MP85557,MP85558		

38	CCB03	CCB03	CCB	05/05/25 17:00		Kareem	OK
39	Q1913-03MS	WC-13-S-202504MS	MS	05/05/25 17:04	MS/MSD fail for many parameters	Kareem	Not Ok
40	Q1913-03MSD	WC-13-S-202504MSD	MSD	05/05/25 17:08	MS/MSD fail for many parameters	Kareem	Not Ok
41	Q1913-03A	WC-13-S-202504A	PS	05/05/25 17:13	MS/MSD fail for many parameters	Kareem	Not Ok
42	PB167837BL	PB167837BL	MB	05/05/25 17:17		Kareem	OK
43	PB167815TB	PB167815TB	MB	05/05/25 17:31		Kareem	OK
44	PB167837BS	PB167837BS	LCS	05/05/25 17:36		Kareem	OK
45	PB167817BL	PB167817BL	MB	05/05/25 17:41		Kareem	OK
46	PB167817BS	PB167817BS	LCS	05/05/25 17:46		Kareem	OK
47	PB167832BL	PB167832BL	MB	05/05/25 17:51		Kareem	OK
48	CCV04	CCV04	CCV	05/05/25 17:55		Kareem	OK
49	CCB04	CCB04	CCB	05/05/25 17:59		Kareem	OK
50	PB167832BS	PB167832BS	LCS	05/05/25 18:04		Kareem	OK
51	PB167868BL	PB167868BL	MB	05/05/25 18:10		Kareem	OK
52	PB167868BS	PB167868BS	LCS	05/05/25 18:14		Kareem	OK
53	PB167774TB	PB167774TB	MB	05/05/25 18:18		Kareem	OK
54	PB167858BL	PB167858BL	MB	05/05/25 18:22		Kareem	OK
55	PB167858BS	PB167858BS	LCS	05/05/25 18:27		Kareem	OK
56	Q1922-01	MH-R	SAM	05/05/25 18:31		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135674

Review By	JANVI	Review On	5/6/2025 4:57:41 PM
Supervise By	jaswal	Supervise On	5/7/2025 1:30:24 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard			
LCS Standard			
Chk Standard	MP85557,MP85558		

57	Q1922-05	MH-S	SAM	05/05/25 18:35		Kareem	OK
58	Q1928-01	MH-Q	SAM	05/05/25 18:39		Kareem	OK
59	CCV05	CCV05	CCV	05/05/25 18:43		Kareem	OK
60	CCB05	CCB05	CCB	05/05/25 18:47		Kareem	OK
61	Q1932-01	COMP-4	SAM	05/05/25 18:52	CCV fail for Be,Mg	Kareem	Not Ok
62	Q1932-03	COMP-5	SAM	05/05/25 18:56	CCV fail for Be,Mg	Kareem	Not Ok
63	Q1932-05	COMP-6	SAM	05/05/25 19:00	CCV fail for Be,Mg	Kareem	Not Ok
64	Q1932-07	COMP-7	SAM	05/05/25 19:04	CCV fail for Be,Mg	Kareem	Not Ok
65	Q1933-01	OK-01-050125	SAM	05/05/25 19:08	CCV fail for Be,Mg	Kareem	Not Ok
66	Q1933-01DUP	OK-01-050125DUP	DUP	05/05/25 19:12	CCV fail for Be,Mg	Kareem	Not Ok
67	Q1933-01L	OK-01-050125L	SD	05/05/25 19:16	CCV fail for Be,Mg	Kareem	Not Ok
68	Q1933-01MS	OK-01-050125MS	MS	05/05/25 19:20	CCV fail for Be,Mg	Kareem	Not Ok
69	Q1933-01MSD	OK-01-050125MSD	MSD	05/05/25 19:24	CCV fail for Be,Mg	Kareem	Not Ok
70	Q1933-01A	OK-01-050125A	PS	05/05/25 19:28	CCV fail for Be,Mg	Kareem	Not Ok
71	CCV06	CCV06	CCV	05/05/25 19:32	CCV fail for Be,Mg	Kareem	OK
72	CCB06	CCB06	CCB	05/05/25 19:37		Kareem	OK
73	Q1939-05	GB3	SAM	05/05/25 19:41	CCV fail for Be,Mg	Kareem	Not Ok
74	Q1935-01	MH-NN	SAM	05/05/25 19:45	CCV fail for Be,Mg	Kareem	Not Ok
75	Q1935-05	MH-MM	SAM	05/05/25 19:49	CCV fail for Be,Mg	Kareem	Not Ok
76	Q1939-01	GB1	SAM	05/05/25 19:53	CCV fail for Be,Mg	Kareem	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135674

Review By	JANVI	Review On	5/6/2025 4:57:41 PM
Supervise By	jaswal	Supervise On	5/7/2025 1:30:24 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard			
LCS Standard			
Chk Standard	MP85557,MP85558		

77	Q1939-03	GB2	SAM	05/05/25 19:57	CCV fail for Be,Mg	Kareem	Not Ok
78	Q1917-04	MH-JJ	SAM	05/05/25 20:01		Kareem	OK
79	Q1922-04	MH-R	SAM	05/05/25 20:06		Kareem	OK
80	Q1922-08	MH-S	SAM	05/05/25 20:10		Kareem	OK
81	Q1925-01	AUD-25-0068	SAM	05/05/25 20:15		Kareem	OK
82	CCV07	CCV07	CCV	05/05/25 20:19		Kareem	OK
83	CCB07	CCB07	CCB	05/05/25 20:23		Kareem	OK
84	Q1925-02	AUD-25-0069	SAM	05/05/25 20:28		Kareem	OK
85	Q1925-03	AUD-25-0070	SAM	05/05/25 20:32		Kareem	OK
86	Q1928-04	MH-Q	SAM	05/05/25 20:36		Kareem	OK
87	Q1929-03	WC-A4-02-C	SAM	05/05/25 20:41		Kareem	OK
88	Q1929-07	WC-A1-03-C	SAM	05/05/25 20:45		Kareem	OK
89	Q1929-11	WC-A1-04-C	SAM	05/05/25 20:50		Kareem	OK
90	Q1932-02	COMP-4	SAM	05/05/25 20:54		Kareem	OK
91	Q1932-04	COMP-5	SAM	05/05/25 20:59		Kareem	OK
92	Q1932-06	COMP-6	SAM	05/05/25 21:03		Kareem	OK
93	CCV08	CCV08	CCV	05/05/25 21:07		Kareem	OK
94	CCB08	CCB08	CCB	05/05/25 21:12		Kareem	OK
95	Q1932-08	COMP-7	SAM	05/05/25 21:16		Kareem	OK
96	Q1935-04	MH-NN	SAM	05/05/25 21:20		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135674

Review By	JANVI	Review On	5/6/2025 4:57:41 PM
Supervise By	jaswal	Supervise On	5/7/2025 1:30:24 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard			
LCS Standard			
Chk Standard	MP85557,MP85558		

97	Q1935-08	MH-MM	SAM	05/05/25 21:25		Kareem	OK
98	Q1935-08DUP	MH-MMDUP	DUP	05/05/25 21:29		Kareem	OK
99	Q1935-08L	MH-MML	SD	05/05/25 21:34		Kareem	OK
100	Q1935-08MS	MH-MMMS	MS	05/05/25 21:38		Kareem	OK
101	Q1935-08MSD	MH-MMMSD	MSD	05/05/25 21:42		Kareem	OK
102	Q1935-08A	MH-MMA	PS	05/05/25 21:46		Kareem	OK
103	Q1945-01	GB1W	SAM	05/05/25 21:51	Zn high	Kareem	Dilution
104	Q1956-07	FB-05022025	SAM	05/05/25 21:55		Kareem	OK
105	CCV09	CCV09	CCV	05/05/25 21:59		Kareem	OK
106	CCB09	CCB09	CCB	05/05/25 22:03		Kareem	OK
107	Q1930-01	WATER-TREATMENT	SAM	05/05/25 22:08		Kareem	OK
108	Q1930-01DUP	WATER-TREATMENT	DUP	05/05/25 22:12		Kareem	OK
109	Q1930-01L	WATER-TREATMENT	SD	05/05/25 22:16		Kareem	OK
110	Q1930-01MS	WATER-TREATMENT	MS	05/05/25 22:21		Kareem	OK
111	Q1930-01MSD	WATER-TREATMENT	MSD	05/05/25 22:25		Kareem	OK
112	Q1930-01A	WATER-TREATMENT	PS	05/05/25 22:29		Kareem	OK
113	CCV10	CCV10	CCV	05/05/25 22:33		Kareem	OK
114	CCB10	CCB10	CCB	05/05/25 22:37		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135689

Review By	JANVI	Review On	5/7/2025 11:23:01 AM
Supervise By	jaswal	Supervise On	5/7/2025 11:23:43 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard	MP85552		
LCS Standard			
Chk Standard	MP85557,MP85558		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	05/06/25 13:43		Kareem	OK
2	S1	S1	CAL2	05/06/25 13:48		Kareem	OK
3	S2	S2	CAL3	05/06/25 13:52		Kareem	OK
4	S3	S3	CAL4	05/06/25 13:56		Kareem	OK
5	S4	S4	CAL5	05/06/25 14:01		Kareem	OK
6	S5	S5	CAL6	05/06/25 14:05		Kareem	OK
7	ICV01	ICV01	ICV	05/06/25 14:47		Kareem	OK
8	LLICV01	LLICV01	LLICV	05/06/25 15:22		Kareem	OK
9	ICB01	ICB01	ICB	05/06/25 15:27		Kareem	OK
10	CRI01	CRI01	CRDL	05/06/25 15:31		Kareem	OK
11	ICSA01	ICSA01	ICSA	05/06/25 15:50		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	05/06/25 15:55		Kareem	OK
13	ICSADL	ICSADL	ICSA	05/06/25 15:59		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	05/06/25 16:03		Kareem	OK
15	CCV01	CCV01	CCV	05/06/25 16:29		Kareem	OK
16	CCB01	CCB01	CCB	05/06/25 16:34		Kareem	OK
17	Q1948-01	001-WILLETS-PT-BL	SAM	05/06/25 16:38		Kareem	OK
18	Q1948-02	002-35TH-AVE(APR)	SAM	05/06/25 16:42		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135689

Review By	JANVI	Review On	5/7/2025 11:23:01 AM
Supervise By	jaswal	Supervise On	5/7/2025 11:23:43 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard	MP85552		
LCS Standard			
Chk Standard	MP85557,MP85558		

19	Q1932-01	COMP-4	SAM	05/06/25 16:47		Kareem	OK
20	Q1932-03	COMP-5	SAM	05/06/25 16:51		Kareem	OK
21	Q1932-05	COMP-6	SAM	05/06/25 16:55		Kareem	OK
22	Q1932-07	COMP-7	SAM	05/06/25 16:59		Kareem	OK
23	Q1933-01	OK-01-050125	SAM	05/06/25 17:03		Kareem	OK
24	Q1933-01DUP	OK-01-050125DUP	DUP	05/06/25 17:07		Kareem	OK
25	Q1933-01L	OK-01-050125L	SD	05/06/25 17:11		Kareem	OK
26	CCV02	CCV02	CCV	05/06/25 17:15		Kareem	OK
27	CCB02	CCB02	CCB	05/06/25 17:20		Kareem	OK
28	Q1933-01MS	OK-01-050125MS	MS	05/06/25 17:24		Kareem	OK
29	Q1933-01MSD	OK-01-050125MSD	MSD	05/06/25 17:28		Kareem	OK
30	Q1933-01A	OK-01-050125A	PS	05/06/25 17:32		Kareem	OK
31	Q1913-03	WC-13-S-202504	SAM	05/06/25 17:43		Kareem	OK
32	Q1913-03DUP	WC-13-S-202504DUP	DUP	05/06/25 17:47		Kareem	OK
33	Q1913-03L	WC-13-S-202504L	SD	05/06/25 17:51		Kareem	OK
34	Q1913-03MS	WC-13-S-202504MS	MS	05/06/25 17:56		Kareem	OK
35	Q1913-03MSD	WC-13-S-202504MSD	MSD	05/06/25 18:00		Kareem	OK
36	Q1913-03A	WC-13-S-202504A	PS	05/06/25 18:04		Kareem	OK
37	CCV03	CCV03	CCV	05/06/25 18:08		Kareem	OK
38	CCB03	CCB03	CCB	05/06/25 18:13		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135689

Review By	JANVI	Review On	5/7/2025 11:23:01 AM
Supervise By	jaswal	Supervise On	5/7/2025 11:23:43 AM

STD. NAME	STD REF.#
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546
ICV Standard	MP85553
CCV Standard	MP85556
ICSA Standard	MP85554,MP85555
CRI Standard	MP85552
LCS Standard	
Chk Standard	MP85557,MP85558

39	PB167860BL	PB167860BL	MB	05/06/25 18:17		Kareem	OK
40	PB167860BS	PB167860BS	LCS	05/06/25 18:21		Kareem	OK
41	PB167781BL	PB167781BL	MB	05/06/25 18:25		Kareem	OK
42	PB167781BS	PB167781BS	LCS	05/06/25 18:37		Kareem	OK
43	PB167757BL	PB167757BL	MB	05/06/25 18:41		Kareem	OK
44	PB167757BS	PB167757BS	LCS	05/06/25 19:01		Kareem	OK
45	Q1949-01	001-WILLETS-PT-BL	SAM	05/06/25 19:05		Kareem	OK
46	CCV04	CCV04	CCV	05/06/25 19:16		Kareem	OK
47	CCB04	CCB04	CCB	05/06/25 19:20		Kareem	OK
48	Q1949-04	002-35TH-AVE(MAY)	SAM	05/06/25 19:24		Kareem	OK
49	Q1949-04DUP	002-35TH-AVE(MAY)	DUP	05/06/25 19:29		Kareem	OK
50	Q1949-04L	002-35TH-AVE(MAY)	SD	05/06/25 19:33		Kareem	OK
51	Q1949-04MS	002-35TH-AVE(MAY)	MS	05/06/25 19:37		Kareem	OK
52	Q1949-04MSD	002-35TH-AVE(MAY)	MSD	05/06/25 19:41		Kareem	OK
53	Q1949-04A	002-35TH-AVE(MAY)	PS	05/06/25 19:45		Kareem	OK
54	Q1939-05	GB3	SAM	05/06/25 19:49	Zn high	Kareem	Dilution
55	Q1935-01	MH-NN	SAM	05/06/25 19:53		Kareem	OK
56	Q1935-05	MH-MM	SAM	05/06/25 19:57		Kareem	OK
57	CCV05	CCV05	CCV	05/06/25 20:01		Kareem	OK
58	CCB05	CCB05	CCB	05/06/25 20:06		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135689

Review By	JANVI	Review On	5/7/2025 11:23:01 AM
Supervise By	jaswal	Supervise On	5/7/2025 11:23:43 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard	MP85552		
LCS Standard			
Chk Standard	MP85557,MP85558		

59	Q1939-01	GB1	SAM	05/06/25 20:10		Kareem	OK
60	Q1939-03	GB2	SAM	05/06/25 20:14		Kareem	OK
61	Q1945-01DL	GB1WDL	SAM	05/06/25 20:18	10x for Zn	Kareem	Confirms
62	Q1946-01	BB-CONCRETE	SAM	05/06/25 20:22		Kareem	OK
63	Q1961-01	EO-01-050525	SAM	05/06/25 20:27		Kareem	OK
64	CCV06	CCV06	CCV	05/06/25 20:31		Kareem	OK
65	CCB06	CCB06	CCB	05/06/25 20:35		Kareem	OK