

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental **SDG No.:** Q1945
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1945 **SAS No.:** Q1945

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167858BS							
Aluminum	ug/L	1000	869		87	80 - 120	P
Antimony	ug/L	400	379		95	80 - 120	P
Arsenic	ug/L	400	370		92	80 - 120	P
Barium	ug/L	100	87.2		87	80 - 120	P
Beryllium	ug/L	100	85.5		86	80 - 120	P
Cadmium	ug/L	100	92.1		92	80 - 120	P
Calcium	ug/L	500	448	J	90	80 - 120	P
Chromium	ug/L	200	183		92	80 - 120	P
Cobalt	ug/L	100	89.4		89	80 - 120	P
Copper	ug/L	150	140		93	80 - 120	P
Iron	ug/L	1500	1400		93	80 - 120	P
Lead	ug/L	500	457		91	80 - 120	P
Magnesium	ug/L	1000	866	J	87	80 - 120	P
Manganese	ug/L	100	89.5		90	80 - 120	P
Nickel	ug/L	250	228		91	80 - 120	P
Potassium	ug/L	5000	4510		90	80 - 120	P
Selenium	ug/L	1000	937		94	80 - 120	P
Silver	ug/L	37.5	34.0		91	80 - 120	P
Sodium	ug/L	1500	1370		91	80 - 120	P
Thallium	ug/L	1000	916		92	80 - 120	P
Vanadium	ug/L	150	134		89	80 - 120	P
Zinc	ug/L	100	91.9		92	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q1945

Contract: GENV01

Lab Code: CHEM

Case No.: Q1945

SAS No.: Q1945

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167866BS Mercury	ug/L	4.0	3.72		93	80 - 120	CV