

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Amsterdam		Date Received:	
Client Sample ID:	PB167889BSD		SDG No.:	Q1945
Lab Sample ID:	PB167889BSD		Matrix:	Water
Analytical Method:	SW8270		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOCMS Group1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050143.D	1	05/07/25 08:53	05/08/25 12:39	PB167889

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	23.2		3.90	10.0	ug/L
62-53-3	Aniline	27.0		1.50	5.00	ug/L
108-95-2	Phenol	43.7		0.91	5.00	ug/L
111-44-4	bis(2-Chloroethyl)ether	41.7		0.81	5.00	ug/L
95-57-8	2-Chlorophenol	43.7		0.58	5.00	ug/L
95-48-7	2-Methylphenol	42.7		1.10	5.00	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	41.6		1.30	5.00	ug/L
98-86-2	Acetophenone	42.8		0.74	5.00	ug/L
65794-96-9	3+4-Methylphenols	42.0		1.10	10.0	ug/L
621-64-7	n-Nitroso-di-n-propylamine	40.2		1.40	2.50	ug/L
67-72-1	Hexachloroethane	41.7		0.65	5.00	ug/L
98-95-3	Nitrobenzene	43.5		0.76	5.00	ug/L
78-59-1	Isophorone	40.9		0.75	5.00	ug/L
88-75-5	2-Nitrophenol	45.0		1.80	5.00	ug/L
105-67-9	2,4-Dimethylphenol	44.0		1.90	5.00	ug/L
111-91-1	bis(2-Chloroethoxy)methane	42.0		0.68	5.00	ug/L
120-83-2	2,4-Dichlorophenol	44.5		0.52	5.00	ug/L
91-20-3	Naphthalene	42.4		0.50	5.00	ug/L
106-47-8	4-Chloroaniline	16.2		0.84	5.00	ug/L
87-68-3	Hexachlorobutadiene	43.7		0.54	5.00	ug/L
105-60-2	Caprolactam	37.8		1.10	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	41.1		0.59	5.00	ug/L
91-57-6	2-Methylnaphthalene	41.1		0.56	5.00	ug/L
77-47-4	Hexachlorocyclopentadiene	98.8	E	3.60	10.0	ug/L
88-06-2	2,4,6-Trichlorophenol	45.6		0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	45.8		0.62	5.00	ug/L
92-52-4	1,1-Biphenyl	44.4		0.53	5.00	ug/L
91-58-7	2-Chloronaphthalene	44.6		0.61	5.00	ug/L
88-74-4	2-Nitroaniline	45.7		1.30	5.00	ug/L

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Amsterdam		Date Received:	
Client Sample ID:	PB167889BSD		SDG No.:	Q1945
Lab Sample ID:	PB167889BSD		Matrix:	Water
Analytical Method:	SW8270		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOCMS Group1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050143.D	1	05/07/25 08:53	05/08/25 12:39	PB167889

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
131-11-3	Dimethylphthalate	42.1		0.61	5.00	ug/L
208-96-8	Acenaphthylene	43.5		0.75	5.00	ug/L
606-20-2	2,6-Dinitrotoluene	43.9		0.92	5.00	ug/L
99-09-2	3-Nitroaniline	23.6		1.10	5.00	ug/L
83-32-9	Acenaphthene	43.4		0.55	5.00	ug/L
51-28-5	2,4-Dinitrophenol	81.0	E	6.00	10.0	ug/L
100-02-7	4-Nitrophenol	75.5		2.40	10.0	ug/L
132-64-9	Dibenzofuran	42.6		0.61	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	43.4		1.20	5.00	ug/L
84-66-2	Diethylphthalate	40.7		0.69	5.00	ug/L
7005-72-3	4-Chlorophenyl-phenylether	42.7		0.68	5.00	ug/L
86-73-7	Fluorene	42.4		0.63	5.00	ug/L
100-01-6	4-Nitroaniline	37.3		1.50	5.00	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	45.5		2.90	10.0	ug/L
86-30-6	n-Nitrosodiphenylamine	46.2		0.58	5.00	ug/L
101-55-3	4-Bromophenyl-phenylether	45.6		0.40	5.00	ug/L
118-74-1	Hexachlorobenzene	45.6		0.52	5.00	ug/L
1912-24-9	Atrazine	44.0		1.00	5.00	ug/L
87-86-5	Pentachlorophenol	97.1	E	1.60	10.0	ug/L
85-01-8	Phenanthrene	44.3		0.50	5.00	ug/L
120-12-7	Anthracene	44.6		0.61	5.00	ug/L
86-74-8	Carbazole	44.0		0.72	5.00	ug/L
84-74-2	Di-n-butylphthalate	43.1		1.20	5.00	ug/L
206-44-0	Fluoranthene	42.5		0.82	5.00	ug/L
129-00-0	Pyrene	46.5		0.50	5.00	ug/L
85-68-7	Butylbenzylphthalate	45.3		1.90	5.00	ug/L
91-94-1	3,3-Dichlorobenzidine	25.0		0.93	10.0	ug/L
56-55-3	Benzo(a)anthracene	45.0		0.45	5.00	ug/L
218-01-9	Chrysene	44.3		0.44	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	44.2		1.60	5.00	ug/L
117-84-0	Di-n-octyl phthalate	44.4		2.30	10.0	ug/L

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Amsterdam		Date Received:	
Client Sample ID:	PB167889BSD		SDG No.:	Q1945
Lab Sample ID:	PB167889BSD		Matrix:	Water
Analytical Method:	SW8270		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOCMS Group1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050143.D	1	05/07/25 08:53	05/08/25 12:39	PB167889

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
205-99-2	Benzo(b)fluoranthene	44.1		0.49	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	43.1		0.48	5.00	ug/L
50-32-8	Benzo(a)pyrene	44.6		0.55	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	50.2		0.59	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	50.4		0.67	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	50.4		0.69	5.00	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	45.8		0.52	5.00	ug/L
123-91-1	1,4-Dioxane	35.6		1.00	5.00	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	42.9		0.72	5.00	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	134		15 (10) - 110 (139)	89%	SPK: 150
13127-88-3	Phenol-d6	129		15 (10) - 110 (134)	86%	SPK: 150
4165-60-0	Nitrobenzene-d5	78.3		30 (49) - 130 (133)	78%	SPK: 100
321-60-8	2-Fluorobiphenyl	80.2		30 (52) - 130 (132)	80%	SPK: 100
118-79-6	2,4,6-Tribromophenol	131		15 (44) - 110 (137)	87%	SPK: 150
1718-51-0	Terphenyl-d14	85.8		30 (48) - 130 (125)	86%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	264000	7.74			
1146-65-2	Naphthalene-d8	911000	10.534			
15067-26-2	Acenaphthene-d10	549000	14.392			
1517-22-2	Phenanthrene-d10	983000	17.139			
1719-03-5	Chrysene-d12	868000	21.386			
1520-96-3	Perylene-d12	894000	24.38			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products