

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: GENV01

Lab Code: CHEM

SAS No.: Q1945

SDG NO.: Q1945

Lab File ID: BM050023.D

DFTPP Injection Date: 04/28/2025

Instrument ID: BNA_M

DFTPP Injection Time: 11:46

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	23.8
68	Less than 2.0% of mass 69	0.4 (1.3) 1
69	Mass 69 relative abundance	28.5
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	35.7
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	26.2
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	11.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.4 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BM050024.D	04/28/2025	12:30
SSTDICC005	SSTDICC005	BM050025.D	04/28/2025	13:09
SSTDICC010	SSTDICC010	BM050026.D	04/28/2025	13:48
SSTDICC020	SSTDICC020	BM050027.D	04/28/2025	14:27
SSTDICCC040	SSTDICCC040	BM050028.D	04/28/2025	15:06
SSTDICC050	SSTDICC050	BM050029.D	04/28/2025	15:45
SSTDICC060	SSTDICC060	BM050030.D	04/28/2025	16:24
SSTDICC080	SSTDICC080	BM050031.D	04/28/2025	17:04