

Surrogate Summary

SDG No.: **Q1945**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1945-01	GB1W	1,2-Dichloroethane-d4	50	52.0	104	70 (74)	130 (125)
		Dibromofluoromethane	50	50.1	100	70 (75)	130 (124)
		Toluene-d8	50	50.6	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.9	98	70 (77)	130 (121)
Q1945-02	GB3W	1,2-Dichloroethane-d4	50	52.3	105	70 (74)	130 (125)
		Dibromofluoromethane	50	50.3	101	70 (75)	130 (124)
		Toluene-d8	50	51.2	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.3	103	70 (77)	130 (121)
VX0506WBL01	VX0506WBL01	1,2-Dichloroethane-d4	50	53.4	107	70 (74)	130 (125)
		Dibromofluoromethane	50	50.3	101	70 (75)	130 (124)
		Toluene-d8	50	49.6	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.7	99	70 (77)	130 (121)
VX0506WBS01	VX0506WBS01	1,2-Dichloroethane-d4	50	45.4	91	70 (74)	130 (125)
		Dibromofluoromethane	50	45.2	90	70 (75)	130 (124)
		Toluene-d8	50	45.6	91	70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.8	89	70 (77)	130 (121)
VX0506WBSD0	VX0506WBSD01	1,2-Dichloroethane-d4	50	46.2	92	70 (74)	130 (125)
		Dibromofluoromethane	50	47.3	95	70 (75)	130 (124)
		Toluene-d8	50	46.6	93	70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.9	92	70 (77)	130 (121)

