

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1945 SAS No.: Q1945 SDG NO.: Q1945
 Lab File ID: VX046050.D Date Analyzed: 05/06/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:48
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	71104	12.018				
UPPER LIMIT	142208	12.518				
LOWER LIMIT	35552	11.518				
EPA SAMPLE NO.						
GB1W	48139	12.02				
GB3W	54778	12.02				
VX0506WBL01	54743	12.02				
VX0506WBS01	67454	12.02				
VX0506WBSD01	67703	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.