

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1948
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1948 **SAS No.:** Q1948
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	500	510	98	95 - 105	P	05/06/2025	14:47	LB135689
	Iron	10500	10000	105	95 - 105	P	05/06/2025	14:47	LB135689
	Lead	975	1000	98	95 - 105	P	05/06/2025	14:47	LB135689

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1948
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1948 **SAS No.:** Q1948
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.1	20.0	106	80 - 120	P	05/06/2025	15:22	LB135689
	Iron	98.2	100	98	80 - 120	P	05/06/2025	15:22	LB135689
	Lead	11.1	12.0	92	80 - 120	P	05/06/2025	15:22	LB135689

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc SDG No.: Q1948
Contract: TULL01 Lab Code: CHEM Case No.: Q1948 SAS No.: Q1948
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1210	1250	97	90 - 110	P	05/06/2025	16:29	LB135689
	Iron	5040	5000	101	90 - 110	P	05/06/2025	16:29	LB135689
	Lead	5030	5000	101	90 - 110	P	05/06/2025	16:29	LB135689
CCV02	Copper	1200	1250	96	90 - 110	P	05/06/2025	17:15	LB135689
	Iron	5000	5000	100	90 - 110	P	05/06/2025	17:15	LB135689
	Lead	4890	5000	98	90 - 110	P	05/06/2025	17:15	LB135689
CCV03	Copper	1170	1250	94	90 - 110	P	05/06/2025	18:08	LB135689
	Iron	4700	5000	94	90 - 110	P	05/06/2025	18:08	LB135689
	Lead	4690	5000	94	90 - 110	P	05/06/2025	18:08	LB135689
CCV04	Copper	1240	1250	99	90 - 110	P	05/06/2025	19:16	LB135689
	Iron	4820	5000	96	90 - 110	P	05/06/2025	19:16	LB135689
	Lead	4970	5000	99	90 - 110	P	05/06/2025	19:16	LB135689
CCV05	Copper	1230	1250	98	90 - 110	P	05/06/2025	20:01	LB135689
	Iron	4860	5000	97	90 - 110	P	05/06/2025	20:01	LB135689
	Lead	4900	5000	98	90 - 110	P	05/06/2025	20:01	LB135689
CCV06	Copper	1240	1250	99	90 - 110	P	05/06/2025	20:31	LB135689
	Iron	4920	5000	98	90 - 110	P	05/06/2025	20:31	LB135689
	Lead	4870	5000	97	90 - 110	P	05/06/2025	20:31	LB135689