



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc **SDG No.:** Q1948
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1948 **SAS No.:** Q1948

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	20.0	+/-20.0	U	20.0	P	05/06/2025	15:27	LB135689
	Iron	100	+/-100	U	100	P	05/06/2025	15:27	LB135689
	Lead	12.0	+/-12.0	U	12.0	P	05/06/2025	15:27	LB135689

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc **SDG No.:** Q1948
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1948 **SAS No.:** Q1948

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	20.0	+/-20.0	U	20.0	P	05/06/2025	16:34	LB135689
	Iron	100	+/-100	U	100	P	05/06/2025	16:34	LB135689
	Lead	12.0	+/-12.0	U	12.0	P	05/06/2025	16:34	LB135689
CCB02	Copper	20.0	+/-20.0	U	20.0	P	05/06/2025	17:20	LB135689
	Iron	100	+/-100	U	100	P	05/06/2025	17:20	LB135689
	Lead	12.0	+/-12.0	U	12.0	P	05/06/2025	17:20	LB135689
CCB03	Copper	20.0	+/-20.0	U	20.0	P	05/06/2025	18:13	LB135689
	Iron	23.9	+/-100	J	100	P	05/06/2025	18:13	LB135689
	Lead	12.0	+/-12.0	U	12.0	P	05/06/2025	18:13	LB135689
CCB04	Copper	20.0	+/-20.0	U	20.0	P	05/06/2025	19:20	LB135689
	Iron	100	+/-100	U	100	P	05/06/2025	19:20	LB135689
	Lead	12.0	+/-12.0	U	12.0	P	05/06/2025	19:20	LB135689
CCB05	Copper	20.0	+/-20.0	U	20.0	P	05/06/2025	20:06	LB135689
	Iron	100	+/-100	U	100	P	05/06/2025	20:06	LB135689
	Lead	12.0	+/-12.0	U	12.0	P	05/06/2025	20:06	LB135689
CCB06	Copper	20.0	+/-20.0	U	20.0	P	05/06/2025	20:35	LB135689
	Iron	100	+/-100	U	100	P	05/06/2025	20:35	LB135689
	Lead	12.0	+/-12.0	U	12.0	P	05/06/2025	20:35	LB135689