

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tully Environmental, Inc **level:** low **sdg no.:** Q1948
contract: TULL01 **lab code:** CHEM **case no.:** Q1948 **sas no.:** Q1948
matrix: Water **sample id:** Q1949-04 **client id:** 002-35TH-AVE(MAY)MS
Percent Solids for Sample: NA **Spiked ID:** Q1949-04MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	148		3.07	J	150	96		P
Iron	ug/L	75 - 125	13900		12100		1500	120		P
Lead	ug/L	75 - 125	481		13.9		500	93		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q1948</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>Q1948</u> sas no.: <u>Q1948</u>
matrix: <u>Water</u>	sample id: <u>Q1949-04</u>	client id: <u>002-35TH-AVE(MAY)MSD</u>
Percent Solids for Sample: NA	Spiked ID: Q1949-04MSD	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	151		3.07	J	150	99		P
Iron	ug/L	75 - 125	14300		12100		1500	148		P
Lead	ug/L	75 - 125	493		13.9		500	96		P