

## Initial and Continuing Calibration Verification

**Client:** Tully Environmental, Inc

**SDG No.:** Q1949

**Project:** Transfer Station-SPDES

**RunNo.:** LB135665

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: <b>ICV1</b> Ammonia as N	mg/L	0.95	1	95	90-110	05/05/2025
Sample ID: <b>CCV1</b> Ammonia as N	mg/L	0.96	1	96	90-110	05/05/2025
Sample ID: <b>CCV2</b> Ammonia as N	mg/L	0.98	1	98	90-110	05/05/2025
Sample ID: <b>CCV3</b> Ammonia as N	mg/L	0.99	1	99	90-110	05/05/2025
Sample ID: <b>CCV4</b> Ammonia as N	mg/L	0.98	1	98	90-110	05/05/2025