

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q1949
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1949 **SAS No.:** Q1949
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	-0.0080	2.0		-18	22	05/06/2025	15:50	LB135689
	Iron	91100	101000	90	85600	116500	05/06/2025	15:50	LB135689
	Lead	0.71			-12	12	05/06/2025	15:50	LB135689
ICSAB01	Copper	516	511	101	434	588	05/06/2025	15:55	LB135689
	Iron	94200	99300	95	84400	114500	05/06/2025	15:55	LB135689
	Lead	48.7	49.0	99	37	61	05/06/2025	15:55	LB135689
ICSA	Copper	9.82	2.0	491	-18	22	05/06/2025	15:59	LB135689
	Iron	94800	101000	94	85600	116500	05/06/2025	15:59	LB135689
	Lead	-29.7			-12	12	05/06/2025	15:59	LB135689
ICSAB	Copper	522	511	102	434	588	05/06/2025	16:03	LB135689
	Iron	91400	99300	92	84400	114500	05/06/2025	16:03	LB135689
	Lead	34.1	49.0	70	37	61	05/06/2025	16:03	LB135689