

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: Q1949 SAS No.: Q1949 SDG No.: Q1949

Instrument ID: MSVOA_N Calibration Date/Time: 05/07/2025 11:01

Lab File ID: VN086528.D Init. Calib. Date(s): 04/23/2025 04/23/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:22 13:17

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.682	1.671	0.5	-0.65	
Toluene	2.008	2.038	0.4	1.49	
Ethyl Benzene	2.260	2.276	0.1	0.71	
m/p-Xylenes	0.858	0.880	0.3	2.56	
o-Xylene	0.840	0.865	0.3	2.98	
1,2-Dichloroethane-d4	2.892	2.674	0.01	-7.54	
Toluene-d8	1.650	1.615	0.01	-2.12	
4-Bromofluorobenzene	0.590	0.568	0.2	-3.73	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.