

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
 Lab Code: CHEM Case No.: Q1949 SAS No.: Q1949 SDG NO.: Q1949
 Lab File ID: VN086528.D Date Analyzed: 05/07/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:01
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (MAY)	0	0.00				
001-WILLETS-PT-BLVD (MAY) MS	0	0.00				
001-WILLETS-PT-BLVD (MAY) MSD	0	0.00				
002-35TH-AVE (MAY)	0	0.00				
VN0507WBL01	0	0.00				
VN0507WBS01	0	0.00				

IS4 =

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.