

Initial and Continuing Calibration Blank Summary

Client: CDM Smith

SDG No.: Q1956

Project: Bergen Point Fueling System

RunNo.: LB135696

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	0.0016	0.0025	J	0.00096	0.005	05/07/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	0.00098	0.0025	J	0.00096	0.005	05/07/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	0.0011	0.0025	J	0.00096	0.005	05/07/2025