



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	CDM Smith				SDG No.:	Q1956				
Contract:	CAMP02	Lab Code:	CHEM		Case No.:	Q1956	SAS No.:	Q1956		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB114	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	11:18	LB135662	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q1956

Contract: CAMP02 **Lab Code:** CHEM

Case No.: Q1956

SAS No.: Q1956

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB71	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	11:30	LB135662
CCB72	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	12:03	LB135662
CCB73	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	12:38	LB135662
CCB74	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	13:06	LB135662
CCB75	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	13:43	LB135662
CCB76	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	13:55	LB135662

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>CDM Smith</u>				SDG No.:	<u>Q1956</u>				
Contract:	<u>CAMP02</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1956</u>	SAS No.:	<u>Q1956</u>		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB116	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	15:11	LB135670	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	CDM Smith				SDG No.:	Q1956			
Contract:	CAMP02	Lab Code:	CHEM		Case No.:	Q1956		SAS No.:	Q1956
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB79	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	15:15	LB135670
CCB80	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	15:48	LB135670
CCB81	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	16:02	LB135670

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1956							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1956	SAS No.: Q1956						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	05/05/2025	14:41	LB135674
	Antimony	50.0	+/-50.0	U	50.0	P	05/05/2025	14:41	LB135674
	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	14:41	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	14:41	LB135674
	Beryllium	6.00	+/-6.00	U	6.00	P	05/05/2025	14:41	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	14:41	LB135674
	Calcium	2000	+/-2000	U	2000	P	05/05/2025	14:41	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	14:41	LB135674
	Cobalt	30.0	+/-30.0	U	30.0	P	05/05/2025	14:41	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	14:41	LB135674
	Iron	100	+/-100	U	100	P	05/05/2025	14:41	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	14:41	LB135674
	Magnesium	2000	+/-2000	U	2000	P	05/05/2025	14:41	LB135674
	Manganese	20.0	+/-20.0	U	20.0	P	05/05/2025	14:41	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	14:41	LB135674
	Potassium	2000	+/-2000	U	2000	P	05/05/2025	14:41	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	14:41	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	14:41	LB135674
	Sodium	2000	+/-2000	U	2000	P	05/05/2025	14:41	LB135674
	Thallium	40.0	+/-40.0	U	40.0	P	05/05/2025	14:41	LB135674
	Vanadium	40.0	+/-40.0	U	40.0	P	05/05/2025	14:41	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	14:41	LB135674

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith SDG No.: Q1956
Contract: CAMP02 Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	05/05/2025	15:26	LB135674
	Antimony	50.0	+/-50.0	U	50.0	P	05/05/2025	15:26	LB135674
	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	15:26	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	15:26	LB135674
	Beryllium	6.00	+/-6.00	U	6.00	P	05/05/2025	15:26	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	15:26	LB135674
	Calcium	2000	+/-2000	U	2000	P	05/05/2025	15:26	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	15:26	LB135674
	Cobalt	30.0	+/-30.0	U	30.0	P	05/05/2025	15:26	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	15:26	LB135674
	Iron	100	+/-100	U	100	P	05/05/2025	15:26	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	15:26	LB135674
	Magnesium	2000	+/-2000	U	2000	P	05/05/2025	15:26	LB135674
	Manganese	20.0	+/-20.0	U	20.0	P	05/05/2025	15:26	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	15:26	LB135674
	Potassium	2000	+/-2000	U	2000	P	05/05/2025	15:26	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	15:26	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	15:26	LB135674
	Sodium	2000	+/-2000	U	2000	P	05/05/2025	15:26	LB135674
	Thallium	40.0	+/-40.0	U	40.0	P	05/05/2025	15:26	LB135674
	Vanadium	40.0	+/-40.0	U	40.0	P	05/05/2025	15:26	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	15:26	LB135674
CCB02	Aluminum	100	+/-100	U	100	P	05/05/2025	16:12	LB135674
	Antimony	50.0	+/-50.0	U	50.0	P	05/05/2025	16:12	LB135674
	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	16:12	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	16:12	LB135674
	Beryllium	6.00	+/-6.00	U	6.00	P	05/05/2025	16:12	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	16:12	LB135674
	Calcium	2000	+/-2000	U	2000	P	05/05/2025	16:12	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	16:12	LB135674
	Cobalt	30.0	+/-30.0	U	30.0	P	05/05/2025	16:12	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	16:12	LB135674
	Iron	100	+/-100	U	100	P	05/05/2025	16:12	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	16:12	LB135674
	Magnesium	2000	+/-2000	U	2000	P	05/05/2025	16:12	LB135674
	Manganese	20.0	+/-20.0	U	20.0	P	05/05/2025	16:12	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	16:12	LB135674
	Potassium	2000	+/-2000	U	2000	P	05/05/2025	16:12	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	16:12	LB135674

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1956							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1956	SAS No.: Q1956						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	16:12	LB135674
	Sodium	2000	+/-2000	U	2000	P	05/05/2025	16:12	LB135674
	Thallium	40.0	+/-40.0	U	40.0	P	05/05/2025	16:12	LB135674
	Vanadium	40.0	+/-40.0	U	40.0	P	05/05/2025	16:12	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	16:12	LB135674
CCB03	Aluminum	100	+/-100	U	100	P	05/05/2025	17:00	LB135674
	Antimony	50.0	+/-50.0	U	50.0	P	05/05/2025	17:00	LB135674
	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	17:00	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	17:00	LB135674
	Beryllium	6.00	+/-6.00	U	6.00	P	05/05/2025	17:00	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	17:00	LB135674
	Calcium	2000	+/-2000	U	2000	P	05/05/2025	17:00	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	17:00	LB135674
	Cobalt	30.0	+/-30.0	U	30.0	P	05/05/2025	17:00	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	17:00	LB135674
	Iron	100	+/-100	U	100	P	05/05/2025	17:00	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	17:00	LB135674
	Magnesium	2000	+/-2000	U	2000	P	05/05/2025	17:00	LB135674
	Manganese	20.0	+/-20.0	U	20.0	P	05/05/2025	17:00	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	17:00	LB135674
	Potassium	2000	+/-2000	U	2000	P	05/05/2025	17:00	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	17:00	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	17:00	LB135674
	Sodium	2000	+/-2000	U	2000	P	05/05/2025	17:00	LB135674
	Thallium	40.0	+/-40.0	U	40.0	P	05/05/2025	17:00	LB135674
	Vanadium	40.0	+/-40.0	U	40.0	P	05/05/2025	17:00	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	17:00	LB135674
CCB04	Aluminum	100	+/-100	U	100	P	05/05/2025	17:59	LB135674
	Antimony	50.0	+/-50.0	U	50.0	P	05/05/2025	17:59	LB135674
	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	17:59	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	17:59	LB135674
	Beryllium	6.00	+/-6.00	U	6.00	P	05/05/2025	17:59	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	17:59	LB135674
	Calcium	2000	+/-2000	U	2000	P	05/05/2025	17:59	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	17:59	LB135674
	Cobalt	30.0	+/-30.0	U	30.0	P	05/05/2025	17:59	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	17:59	LB135674
	Iron	100	+/-100	U	100	P	05/05/2025	17:59	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	17:59	LB135674

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1956							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1956	SAS No.: Q1956						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	05/05/2025	17:59	LB135674
	Manganese	20.0	+/-20.0	U	20.0	P	05/05/2025	17:59	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	17:59	LB135674
	Potassium	2000	+/-2000	U	2000	P	05/05/2025	17:59	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	17:59	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	17:59	LB135674
	Sodium	2000	+/-2000	U	2000	P	05/05/2025	17:59	LB135674
	Thallium	40.0	+/-40.0	U	40.0	P	05/05/2025	17:59	LB135674
	Vanadium	40.0	+/-40.0	U	40.0	P	05/05/2025	17:59	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	17:59	LB135674
CCB05	Aluminum	100	+/-100	U	100	P	05/05/2025	18:47	LB135674
	Antimony	50.0	+/-50.0	U	50.0	P	05/05/2025	18:47	LB135674
	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	18:47	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	18:47	LB135674
	Beryllium	6.00	+/-6.00	U	6.00	P	05/05/2025	18:47	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	18:47	LB135674
	Calcium	2000	+/-2000	U	2000	P	05/05/2025	18:47	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	18:47	LB135674
	Cobalt	30.0	+/-30.0	U	30.0	P	05/05/2025	18:47	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	18:47	LB135674
	Iron	100	+/-100	U	100	P	05/05/2025	18:47	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	18:47	LB135674
	Magnesium	2000	+/-2000	U	2000	P	05/05/2025	18:47	LB135674
	Manganese	20.0	+/-20.0	U	20.0	P	05/05/2025	18:47	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	18:47	LB135674
	Potassium	2000	+/-2000	U	2000	P	05/05/2025	18:47	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	18:47	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	18:47	LB135674
	Sodium	2000	+/-2000	U	2000	P	05/05/2025	18:47	LB135674
	Thallium	40.0	+/-40.0	U	40.0	P	05/05/2025	18:47	LB135674
	Vanadium	40.0	+/-40.0	U	40.0	P	05/05/2025	18:47	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	18:47	LB135674
CCB06	Aluminum	100	+/-100	U	100	P	05/05/2025	19:37	LB135674
	Antimony	50.0	+/-50.0	U	50.0	P	05/05/2025	19:37	LB135674
	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	19:37	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	19:37	LB135674
	Beryllium	6.00	+/-6.00	U	6.00	P	05/05/2025	19:37	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	19:37	LB135674
	Calcium	2000	+/-2000	U	2000	P	05/05/2025	19:37	LB135674

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1956							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1956	SAS No.: Q1956						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	19:37	LB135674
	Cobalt	30.0	+/-30.0	U	30.0	P	05/05/2025	19:37	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	19:37	LB135674
	Iron	100	+/-100	U	100	P	05/05/2025	19:37	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	19:37	LB135674
	Magnesium	2000	+/-2000	U	2000	P	05/05/2025	19:37	LB135674
	Manganese	20.0	+/-20.0	U	20.0	P	05/05/2025	19:37	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	19:37	LB135674
	Potassium	2000	+/-2000	U	2000	P	05/05/2025	19:37	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	19:37	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	19:37	LB135674
	Sodium	2000	+/-2000	U	2000	P	05/05/2025	19:37	LB135674
	Thallium	40.0	+/-40.0	U	40.0	P	05/05/2025	19:37	LB135674
	Vanadium	40.0	+/-40.0	U	40.0	P	05/05/2025	19:37	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	19:37	LB135674
CCB07	Aluminum	100	+/-100	U	100	P	05/05/2025	20:23	LB135674
	Antimony	50.0	+/-50.0	U	50.0	P	05/05/2025	20:23	LB135674
	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	20:23	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	20:23	LB135674
	Beryllium	6.00	+/-6.00	U	6.00	P	05/05/2025	20:23	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	20:23	LB135674
	Calcium	2000	+/-2000	U	2000	P	05/05/2025	20:23	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	20:23	LB135674
	Cobalt	30.0	+/-30.0	U	30.0	P	05/05/2025	20:23	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	20:23	LB135674
	Iron	100	+/-100	U	100	P	05/05/2025	20:23	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	20:23	LB135674
	Magnesium	2000	+/-2000	U	2000	P	05/05/2025	20:23	LB135674
	Manganese	20.0	+/-20.0	U	20.0	P	05/05/2025	20:23	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	20:23	LB135674
	Potassium	2000	+/-2000	U	2000	P	05/05/2025	20:23	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	20:23	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	20:23	LB135674
	Sodium	2000	+/-2000	U	2000	P	05/05/2025	20:23	LB135674
	Thallium	40.0	+/-40.0	U	40.0	P	05/05/2025	20:23	LB135674
	Vanadium	40.0	+/-40.0	U	40.0	P	05/05/2025	20:23	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	20:23	LB135674
CCB08	Aluminum	100	+/-100	U	100	P	05/05/2025	21:12	LB135674
	Antimony	50.0	+/-50.0	U	50.0	P	05/05/2025	21:12	LB135674

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1956							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1956	SAS No.: Q1956						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	21:12	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	21:12	LB135674
	Beryllium	6.00	+/-6.00	U	6.00	P	05/05/2025	21:12	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	21:12	LB135674
	Calcium	2000	+/-2000	U	2000	P	05/05/2025	21:12	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	21:12	LB135674
	Cobalt	30.0	+/-30.0	U	30.0	P	05/05/2025	21:12	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	21:12	LB135674
	Iron	100	+/-100	U	100	P	05/05/2025	21:12	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	21:12	LB135674
	Magnesium	2000	+/-2000	U	2000	P	05/05/2025	21:12	LB135674
	Manganese	20.0	+/-20.0	U	20.0	P	05/05/2025	21:12	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	21:12	LB135674
	Potassium	2000	+/-2000	U	2000	P	05/05/2025	21:12	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	21:12	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	21:12	LB135674
	Sodium	2000	+/-2000	U	2000	P	05/05/2025	21:12	LB135674
	Thallium	40.0	+/-40.0	U	40.0	P	05/05/2025	21:12	LB135674
	Vanadium	40.0	+/-40.0	U	40.0	P	05/05/2025	21:12	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	21:12	LB135674
CCB09	Aluminum	100	+/-100	U	100	P	05/05/2025	22:03	LB135674
	Antimony	50.0	+/-50.0	U	50.0	P	05/05/2025	22:03	LB135674
	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	22:03	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	22:03	LB135674
	Beryllium	6.00	+/-6.00	U	6.00	P	05/05/2025	22:03	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	22:03	LB135674
	Calcium	2000	+/-2000	U	2000	P	05/05/2025	22:03	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	22:03	LB135674
	Cobalt	30.0	+/-30.0	U	30.0	P	05/05/2025	22:03	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	22:03	LB135674
	Iron	100	+/-100	U	100	P	05/05/2025	22:03	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	22:03	LB135674
	Magnesium	2000	+/-2000	U	2000	P	05/05/2025	22:03	LB135674
	Manganese	20.0	+/-20.0	U	20.0	P	05/05/2025	22:03	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	22:03	LB135674
	Potassium	2000	+/-2000	U	2000	P	05/05/2025	22:03	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	22:03	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	22:03	LB135674
	Sodium	2000	+/-2000	U	2000	P	05/05/2025	22:03	LB135674

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1956							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1956	SAS No.: Q1956						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	05/05/2025	22:03	LB135674
	Vanadium	40.0	+/-40.0	U	40.0	P	05/05/2025	22:03	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	22:03	LB135674
CCB10	Aluminum	100	+/-100	U	100	P	05/05/2025	22:37	LB135674
	Antimony	50.0	+/-50.0	U	50.0	P	05/05/2025	22:37	LB135674
	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	22:37	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	22:37	LB135674
	Beryllium	6.00	+/-6.00	U	6.00	P	05/05/2025	22:37	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	22:37	LB135674
	Calcium	2000	+/-2000	U	2000	P	05/05/2025	22:37	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	22:37	LB135674
	Cobalt	30.0	+/-30.0	U	30.0	P	05/05/2025	22:37	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	22:37	LB135674
	Iron	100	+/-100	U	100	P	05/05/2025	22:37	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	22:37	LB135674
	Magnesium	2000	+/-2000	U	2000	P	05/05/2025	22:37	LB135674
	Manganese	20.0	+/-20.0	U	20.0	P	05/05/2025	22:37	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	22:37	LB135674
	Potassium	2000	+/-2000	U	2000	P	05/05/2025	22:37	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	22:37	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	22:37	LB135674
	Sodium	2000	+/-2000	U	2000	P	05/05/2025	22:37	LB135674
	Thallium	40.0	+/-40.0	U	40.0	P	05/05/2025	22:37	LB135674
	Vanadium	40.0	+/-40.0	U	40.0	P	05/05/2025	22:37	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	22:37	LB135674

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1956							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1956	SAS No.: Q1956						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	05/07/2025	13:09	LB135699
	Antimony	50.0	+/-50.0	U	50.0	P	05/07/2025	13:09	LB135699
	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	13:09	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	13:09	LB135699
	Beryllium	6.00	+/-6.00	U	6.00	P	05/07/2025	13:09	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	13:09	LB135699
	Calcium	2000	+/-2000	U	2000	P	05/07/2025	13:09	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	13:09	LB135699
	Cobalt	30.0	+/-30.0	U	30.0	P	05/07/2025	13:09	LB135699
	Copper	20.0	+/-20.0	U	20.0	P	05/07/2025	13:09	LB135699
	Iron	100	+/-100	U	100	P	05/07/2025	13:09	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	13:09	LB135699
	Magnesium	2000	+/-2000	U	2000	P	05/07/2025	13:09	LB135699
	Manganese	20.0	+/-20.0	U	20.0	P	05/07/2025	13:09	LB135699
	Nickel	40.0	+/-40.0	U	40.0	P	05/07/2025	13:09	LB135699
	Potassium	2000	+/-2000	U	2000	P	05/07/2025	13:09	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	13:09	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	13:09	LB135699
	Sodium	2000	+/-2000	U	2000	P	05/07/2025	13:09	LB135699
	Thallium	40.0	+/-40.0	U	40.0	P	05/07/2025	13:09	LB135699
	Vanadium	40.0	+/-40.0	U	40.0	P	05/07/2025	13:09	LB135699
	Zinc	40.0	+/-40.0	U	40.0	P	05/07/2025	13:09	LB135699

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1956							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1956	SAS No.: Q1956						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	05/07/2025	13:43	LB135699
	Antimony	50.0	+/-50.0	U	50.0	P	05/07/2025	13:43	LB135699
	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	13:43	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	13:43	LB135699
	Beryllium	6.00	+/-6.00	U	6.00	P	05/07/2025	13:43	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	13:43	LB135699
	Calcium	2000	+/-2000	U	2000	P	05/07/2025	13:43	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	13:43	LB135699
	Cobalt	30.0	+/-30.0	U	30.0	P	05/07/2025	13:43	LB135699
	Copper	20.0	+/-20.0	U	20.0	P	05/07/2025	13:43	LB135699
	Iron	100	+/-100	U	100	P	05/07/2025	13:43	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	13:43	LB135699
	Magnesium	2000	+/-2000	U	2000	P	05/07/2025	13:43	LB135699
	Manganese	20.0	+/-20.0	U	20.0	P	05/07/2025	13:43	LB135699
	Nickel	40.0	+/-40.0	U	40.0	P	05/07/2025	13:43	LB135699
	Potassium	2000	+/-2000	U	2000	P	05/07/2025	13:43	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	13:43	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	13:43	LB135699
	Sodium	2000	+/-2000	U	2000	P	05/07/2025	13:43	LB135699
	Thallium	40.0	+/-40.0	U	40.0	P	05/07/2025	13:43	LB135699
	Vanadium	40.0	+/-40.0	U	40.0	P	05/07/2025	13:43	LB135699
	Zinc	40.0	+/-40.0	U	40.0	P	05/07/2025	13:43	LB135699
CCB02	Aluminum	100	+/-100	U	100	P	05/07/2025	14:39	LB135699
	Antimony	50.0	+/-50.0	U	50.0	P	05/07/2025	14:39	LB135699
	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	14:39	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	14:39	LB135699
	Beryllium	6.00	+/-6.00	U	6.00	P	05/07/2025	14:39	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	14:39	LB135699
	Calcium	2000	+/-2000	U	2000	P	05/07/2025	14:39	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	14:39	LB135699
	Cobalt	30.0	+/-30.0	U	30.0	P	05/07/2025	14:39	LB135699
	Copper	20.0	+/-20.0	U	20.0	P	05/07/2025	14:39	LB135699
	Iron	100	+/-100	U	100	P	05/07/2025	14:39	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	14:39	LB135699
	Magnesium	2000	+/-2000	U	2000	P	05/07/2025	14:39	LB135699
	Manganese	20.0	+/-20.0	U	20.0	P	05/07/2025	14:39	LB135699
	Nickel	40.0	+/-40.0	U	40.0	P	05/07/2025	14:39	LB135699
	Potassium	2000	+/-2000	U	2000	P	05/07/2025	14:39	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	14:39	LB135699

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1956							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1956	SAS No.: Q1956						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	14:39	LB135699
	Sodium	2000	+/-2000	U	2000	P	05/07/2025	14:39	LB135699
	Thallium	40.0	+/-40.0	U	40.0	P	05/07/2025	14:39	LB135699
	Vanadium	40.0	+/-40.0	U	40.0	P	05/07/2025	14:39	LB135699
	Zinc	40.0	+/-40.0	U	40.0	P	05/07/2025	14:39	LB135699
CCB03	Aluminum	100	+/-100	U	100	P	05/07/2025	15:27	LB135699
	Antimony	50.0	+/-50.0	U	50.0	P	05/07/2025	15:27	LB135699
	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	15:27	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	15:27	LB135699
	Beryllium	6.00	+/-6.00	U	6.00	P	05/07/2025	15:27	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	15:27	LB135699
	Calcium	2000	+/-2000	U	2000	P	05/07/2025	15:27	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	15:27	LB135699
	Cobalt	30.0	+/-30.0	U	30.0	P	05/07/2025	15:27	LB135699
	Copper	20.0	+/-20.0	U	20.0	P	05/07/2025	15:27	LB135699
	Iron	100	+/-100	U	100	P	05/07/2025	15:27	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	15:27	LB135699
	Magnesium	2000	+/-2000	U	2000	P	05/07/2025	15:27	LB135699
	Manganese	20.0	+/-20.0	U	20.0	P	05/07/2025	15:27	LB135699
	Nickel	40.0	+/-40.0	U	40.0	P	05/07/2025	15:27	LB135699
	Potassium	2000	+/-2000	U	2000	P	05/07/2025	15:27	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	15:27	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	15:27	LB135699
	Sodium	2000	+/-2000	U	2000	P	05/07/2025	15:27	LB135699
	Thallium	40.0	+/-40.0	U	40.0	P	05/07/2025	15:27	LB135699
	Vanadium	40.0	+/-40.0	U	40.0	P	05/07/2025	15:27	LB135699
	Zinc	40.0	+/-40.0	U	40.0	P	05/07/2025	15:27	LB135699
CCB04	Aluminum	100	+/-100	U	100	P	05/07/2025	16:09	LB135699
	Antimony	50.0	+/-50.0	U	50.0	P	05/07/2025	16:09	LB135699
	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	16:09	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	16:09	LB135699
	Beryllium	6.00	+/-6.00	U	6.00	P	05/07/2025	16:09	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	16:09	LB135699
	Calcium	2000	+/-2000	U	2000	P	05/07/2025	16:09	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	16:09	LB135699
	Cobalt	30.0	+/-30.0	U	30.0	P	05/07/2025	16:09	LB135699
	Copper	20.0	+/-20.0	U	20.0	P	05/07/2025	16:09	LB135699
	Iron	100	+/-100	U	100	P	05/07/2025	16:09	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	16:09	LB135699

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1956							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1956	SAS No.: Q1956						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	05/07/2025	16:09	LB135699
	Manganese	20.0	+/-20.0	U	20.0	P	05/07/2025	16:09	LB135699
	Nickel	40.0	+/-40.0	U	40.0	P	05/07/2025	16:09	LB135699
	Potassium	2000	+/-2000	U	2000	P	05/07/2025	16:09	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	16:09	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	16:09	LB135699
	Sodium	2000	+/-2000	U	2000	P	05/07/2025	16:09	LB135699
	Thallium	40.0	+/-40.0	U	40.0	P	05/07/2025	16:09	LB135699
	Vanadium	40.0	+/-40.0	U	40.0	P	05/07/2025	16:09	LB135699
	Zinc	40.0	+/-40.0	U	40.0	P	05/07/2025	16:09	LB135699
CCB05	Aluminum	100	+/-100	U	100	P	05/07/2025	16:52	LB135699
	Antimony	50.0	+/-50.0	U	50.0	P	05/07/2025	16:52	LB135699
	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	16:52	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	16:52	LB135699
	Beryllium	6.00	+/-6.00	U	6.00	P	05/07/2025	16:52	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	16:52	LB135699
	Calcium	2000	+/-2000	U	2000	P	05/07/2025	16:52	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	16:52	LB135699
	Cobalt	30.0	+/-30.0	U	30.0	P	05/07/2025	16:52	LB135699
	Copper	20.0	+/-20.0	U	20.0	P	05/07/2025	16:52	LB135699
	Iron	100	+/-100	U	100	P	05/07/2025	16:52	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	16:52	LB135699
	Magnesium	2000	+/-2000	U	2000	P	05/07/2025	16:52	LB135699
	Manganese	20.0	+/-20.0	U	20.0	P	05/07/2025	16:52	LB135699
	Nickel	40.0	+/-40.0	U	40.0	P	05/07/2025	16:52	LB135699
	Potassium	2000	+/-2000	U	2000	P	05/07/2025	16:52	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	16:52	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	16:52	LB135699
	Sodium	2000	+/-2000	U	2000	P	05/07/2025	16:52	LB135699
	Thallium	40.0	+/-40.0	U	40.0	P	05/07/2025	16:52	LB135699
	Vanadium	40.0	+/-40.0	U	40.0	P	05/07/2025	16:52	LB135699
	Zinc	40.0	+/-40.0	U	40.0	P	05/07/2025	16:52	LB135699

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	CDM Smith				SDG No.:	Q1956			
Contract:	CAMP02	Lab Code:	CHEM		Case No.:	Q1956	SAS No.:	Q1956	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number