

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: CDM Smith **SDG No.:** Q1956
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1956 **SAS No.:** Q1956
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	240000	255000	94	216000	294000	05/05/2025	14:52	LB135674
	Antimony	-3.42			-50	50	05/05/2025	14:52	LB135674
	Arsenic	4.47			-20	20	05/05/2025	14:52	LB135674
	Barium	0.40	6.0	7	-94	106	05/05/2025	14:52	LB135674
	Beryllium	1.29			-6	6	05/05/2025	14:52	LB135674
	Cadmium	-1.99	1.0	199	-5	7	05/05/2025	14:52	LB135674
	Calcium	226000	245000	92	208000	282000	05/05/2025	14:52	LB135674
	Chromium	53.8	52.0	104	42	62	05/05/2025	14:52	LB135674
	Cobalt	1.10			-30	30	05/05/2025	14:52	LB135674
	Copper	12.1	2.0	605	-18	22	05/05/2025	14:52	LB135674
	Iron	91500	101000	91	85600	116500	05/05/2025	14:52	LB135674
	Lead	-1.42			-12	12	05/05/2025	14:52	LB135674
	Magnesium	242000	255000	95	216000	294000	05/05/2025	14:52	LB135674
	Manganese	1.97	7.0	28	-13	27	05/05/2025	14:52	LB135674
	Nickel	1.16	2.0	58	-38	42	05/05/2025	14:52	LB135674
	Potassium	-11.6			0	0	05/05/2025	14:52	LB135674
	Selenium	-15.2			-20	20	05/05/2025	14:52	LB135674
	Silver	0.79			-10	10	05/05/2025	14:52	LB135674
	Sodium	7.57			0	0	05/05/2025	14:52	LB135674
	Thallium	5.95			-40	40	05/05/2025	14:52	LB135674
	Vanadium	1.94			-40	40	05/05/2025	14:52	LB135674
	Zinc	3.23			-40	40	05/05/2025	14:52	LB135674
ICSAB01	Aluminum	233000	247000	94	209000	285000	05/05/2025	15:07	LB135674
	Antimony	651	618	105	525	711	05/05/2025	15:07	LB135674
	Arsenic	112	104	108	88.4	120	05/05/2025	15:07	LB135674
	Barium	496	537	92	437	637	05/05/2025	15:07	LB135674
	Beryllium	504	495	102	420	570	05/05/2025	15:07	LB135674
	Cadmium	1050	972	108	826	1120	05/05/2025	15:07	LB135674
	Calcium	219000	235000	93	199000	271000	05/05/2025	15:07	LB135674
	Chromium	572	542	106	460	624	05/05/2025	15:07	LB135674
	Cobalt	517	476	109	404	548	05/05/2025	15:07	LB135674
	Copper	520	511	102	434	588	05/05/2025	15:07	LB135674
	Iron	93700	99300	94	84400	114500	05/05/2025	15:07	LB135674
	Lead	48.4	49.0	99	37	61	05/05/2025	15:07	LB135674
	Magnesium	234000	248000	94	210000	286000	05/05/2025	15:07	LB135674
	Manganese	491	507	97	430	584	05/05/2025	15:07	LB135674
	Nickel	1020	954	107	810	1100	05/05/2025	15:07	LB135674
	Potassium	-13.2			0	0	05/05/2025	15:07	LB135674
	Selenium	35.7	46.0	78	26	66	05/05/2025	15:07	LB135674
	Silver	195	201	97	170	232	05/05/2025	15:07	LB135674
	Sodium	-28.0			0	0	05/05/2025	15:07	LB135674
	Thallium	104	108	96	68	148	05/05/2025	15:07	LB135674

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>CDM Smith</u>	SDG No.: <u>Q1956</u>
Contract: <u>CAMP02</u>	Lab Code: <u>CHEM</u>
ICS Source: <u>EPA</u>	Case No.: <u>Q1956</u>
	SAS No.: <u>Q1956</u>
	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	487	491	99	417	565	05/05/2025	15:07	LB135674
	Zinc	1070	952	112	809	1095	05/05/2025	15:07	LB135674
ICSA01	Aluminum	233000	255000	91	216000	294000	05/07/2025	13:21	LB135699
	Antimony	-1.02			-50	50	05/07/2025	13:21	LB135699
	Arsenic	-1.60			-20	20	05/07/2025	13:21	LB135699
	Barium	-1.29	6.0	22	-94	106	05/07/2025	13:21	LB135699
	Beryllium	1.26			-6	6	05/07/2025	13:21	LB135699
	Cadmium	-2.63	1.0	263	-5	7	05/07/2025	13:21	LB135699
	Calcium	225000	245000	92	208000	282000	05/07/2025	13:21	LB135699
	Chromium	54.1	52.0	104	42	62	05/07/2025	13:21	LB135699
	Cobalt	1.12			-30	30	05/07/2025	13:21	LB135699
	Copper	-2.32	2.0	116	-18	22	05/07/2025	13:21	LB135699
	Iron	94700	101000	94	85600	116500	05/07/2025	13:21	LB135699
	Lead	-2.88			-12	12	05/07/2025	13:21	LB135699
	Magnesium	239000	255000	94	216000	294000	05/07/2025	13:21	LB135699
	Manganese	7.87	7.0	112	-13	27	05/07/2025	13:21	LB135699
	Nickel	2.04	2.0	102	-38	42	05/07/2025	13:21	LB135699
	Potassium	0.38			0	0	05/07/2025	13:21	LB135699
	Selenium	-9.82			-20	20	05/07/2025	13:21	LB135699
	Silver	1.08			-10	10	05/07/2025	13:21	LB135699
	Sodium	35.9			0	0	05/07/2025	13:21	LB135699
	Thallium	7.75			-40	40	05/07/2025	13:21	LB135699
	Vanadium	2.40			-40	40	05/07/2025	13:21	LB135699
	Zinc	3.10			-40	40	05/07/2025	13:21	LB135699
ICSAB01	Aluminum	231000	247000	94	209000	285000	05/07/2025	13:25	LB135699
	Antimony	642	618	104	525	711	05/07/2025	13:25	LB135699
	Arsenic	117	104	112	88.4	120	05/07/2025	13:25	LB135699
	Barium	502	537	94	437	637	05/07/2025	13:25	LB135699
	Beryllium	528	495	107	420	570	05/07/2025	13:25	LB135699
	Cadmium	1070	972	110	826	1120	05/07/2025	13:25	LB135699
	Calcium	221000	235000	94	199000	271000	05/07/2025	13:25	LB135699
	Chromium	586	542	108	460	624	05/07/2025	13:25	LB135699
	Cobalt	523	476	110	404	548	05/07/2025	13:25	LB135699
	Copper	509	511	100	434	588	05/07/2025	13:25	LB135699
	Iron	90800	99300	91	84400	114500	05/07/2025	13:25	LB135699
	Lead	55.5	49.0	113	37	61	05/07/2025	13:25	LB135699
	Magnesium	235000	248000	95	210000	286000	05/07/2025	13:25	LB135699
	Manganese	512	507	101	430	584	05/07/2025	13:25	LB135699
	Nickel	1040	954	109	810	1100	05/07/2025	13:25	LB135699
	Potassium	-13.9			0	0	05/07/2025	13:25	LB135699
	Selenium	41.2	46.0	90	26	66	05/07/2025	13:25	LB135699
	Silver	197	201	98	170	232	05/07/2025	13:25	LB135699

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: CDM Smith **SDG No.:** Q1956
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1956 **SAS No.:** Q1956
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Sodium	-3.21			0	0	05/07/2025	13:25	LB135699
	Thallium	115	108	106	68	148	05/07/2025	13:25	LB135699
	Vanadium	501	491	102	417	565	05/07/2025	13:25	LB135699
	Zinc	1080	952	113	809	1095	05/07/2025	13:25	LB135699